

Röntgenteraapia aparatuur. Koordinaadid, mehhanismid ja astmikud

Radiotherapy equipment - Coordinates, movements and scales

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 61217:2012 sisaldab Euroopa standardi EN 61217:2012 ingliskeelset teksti.	This Estonian standard EVS-EN 61217:2012 consists of the English text of the European standard EN 61217:2012.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 13.04.2012.	Date of Availability of the European standard is 13.04.2012.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.50, 13.280

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:
Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

English version

**Radiotherapy equipment -
Coordinates, movements and scales
(IEC 61217:2011)**

Appareils utilisés en radiothérapie -
Coordonnées, mouvements et échelles
(CEI 61217:2011)

Strahlentherapie-Einrichtungen -
Koordinaten, Bewegungen und Skalen
(IEC 61217:2011)

This European Standard was approved by CENELEC on 2012-01-11. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 62C/530/FDIS, future edition 2 of IEC 61217, prepared by SC 62C, "Equipment for radiotherapy, nuclear medicine and radiation dosimetry", of IEC/TC 62, "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61217:2012.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2012-10-11
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2015-01-11

This document supersedes EN 61217:1996 + A1:2001 + A2:2008.

EN 61217:2012 constitutes a technical revision to include imager and focus coordinate systems in 3.12. Beyond this subclause, changes were only introduced where needed to include the above coordinate systems.

In this standard, the following print types are used:

– Requirements and definitions: roman type.

– *Test specifications: italic type.*

– Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

– TERMS USED THROUGHOUT THIS STANDARD THAT HAVE BEEN LISTED IN THE INDEX OF DEFINED TERMS: SMALL CAPITALS.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

– “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;

– “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

– “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For the relationship with EU Directive(s) see informative Annex ZZ, which is an integral part of this document.

Endorsement notice

The text of the International Standard IEC 61217:2011 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60976:2007 NOTE Harmonized as EN 60976:2007 (not modified).

IEC 61168:1993 NOTE Harmonized as EN 61168:1994 (not modified).

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment -	EN 60601-1	2006
+ corr. December	2006	Part 1: General requirements for basic safety	+ corr. March	2010
+ corr. December	2007	and essential performance	+ A11	2011
IEC 60601-1-3	2008	Medical electrical equipment -	EN 60601-1-3	2008
		Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	+ corr. March	2010
IEC 60601-2-1	2009	Medical electrical equipment - Part 2-1: Particular requirements for the basic safety and essential performance of electron accelerators in the range 1 MeV to 50 MeV	EN 60601-2-1	201X ¹⁾
IEC 60601-2-11	1997	Medical electrical equipment - Part 2-11: Particular requirements for the safety of gamma beam therapy equipment	EN 60601-2-11	1997
IEC 60601-2-29	2008	Medical electrical equipment - Part 2-29: Particular requirements for the basic safety and essential performance of radiotherapy simulators	EN 60601-2-29 + A11	2008 2011
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 62083	2009	Medical electrical equipment - Requirements for the safety of radiotherapy treatment planning systems	EN 62083	2009

¹⁾ To be published.

Annex ZZ
(informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers the relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC for ERs 9.1 and 11.2.1 last sentence only.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope and object.....	10
2 Normative references	10
3 Coordinate systems.....	10
3.1 General.....	10
3.2 General rules	11
3.3 Fixed reference system ("f") (Figure 1a)	12
3.4 GANTRY coordinate system ("g") (Figure 4)	12
3.5 BEAM LIMITING DEVICE or DELINEATOR coordinate system ("b") (Figure 5)	13
3.6 WEDGE FILTER coordinate system ("w") (Figure 7)	13
3.7 X-RAY IMAGE RECEPTOR coordinate system ("r") (Figures 6 and 8)	14
3.8 PATIENT SUPPORT coordinate system ("s") (Figure 9).....	14
3.9 Table top eccentric rotation coordinate system ("e") (Figures 10 and 11).....	15
3.10 Table top coordinate system ("t") (Figures 10, 11, 18 and 19).....	15
3.11 PATIENT coordinate system ("p") (Figures 17a and 17b).....	16
3.12 Imager coordinate system ("i") and focus coordinate system ("o")	17
3.12.1 General	17
3.12.2 The imager coordinate system ("i")	17
3.12.3 Focus coordinate system ("o")	18
4 Identification of scales and digital DISPLAYS.....	18
5 Designation of ME EQUIPMENT movements	19
6 ME EQUIPMENT zero positions.....	19
7 List of scales, graduations, directions and DISPLAYS.....	20
7.1 General.....	20
7.2 Rotation of the GANTRY (Figures 14a and 14b).....	20
7.3 Rotation of the BEAM LIMITING DEVICE or DELINEATOR (Figures 15a and 15b)	20
7.4 Rotation of the WEDGE FILTER (Figures 7 and 14a).....	20
7.5 RADIATION FIELD or DELINEATED RADIATION FIELD	21
7.5.1 General	21
7.5.2 Edges of RADIATION FIELD or DELINEATED RADIATION FIELD (Figure 16a).....	21
7.5.3 DISPLAY of RADIATION FIELD or DELINEATED RADIATION FIELD (Figures 16a to 16k)	22
7.6 PATIENT SUPPORT isocentric rotation	23
7.7 Table top eccentric rotation	23
7.8 Table top linear and angular movements	24
7.8.1 Vertical displacement of the table top	24
7.8.2 Longitudinal displacement of the table top	24
7.8.3 Lateral displacement of the table top	24
7.8.4 Pitch of the table top	24
7.8.5 Roll of the table top	24
7.9 X-RAY IMAGE RECEPTOR movements.....	24
7.9.1 X-RAY IMAGE RECEPTOR rotation.....	24
7.9.2 X-RAY IMAGE RECEPTOR radial displacement from RADIATION SOURCE (SID)	25
7.9.3 X-RAY IMAGE RECEPTOR radial displacement from ISOCENTRE	25

7.9.4 X-RAY IMAGE RECEPTOR longitudinal displacement	25
7.9.5 X-RAY IMAGE RECEPTOR lateral displacement	25
7.10 Other scales	25
Annex A (informative) Examples of coordinate transformations between individual coordinate systems	57
Annex B (informative) Coordinate transformations between IEC and DICOM PATIENT coordinates	64
Bibliography	65
Index of defined terms	66
Figure 1a – Coordinate systems for an isocentric RADIOTHERAPY EQUIPMENT (see 3.1) with all angular positions set to zero	27
Figure 1b – Translation of origin I_d along X_m , Y_m , Z_m and rotation around axis Z_d parallel to Z_m (see 3.2d))	28
Figure 1c – Translation of origin I_d along X_m , Y_m , Z_m and rotation around axis Y_d parallel to Y_m (see 3.2d))	28
Figure 2 – X Y Z right-hand coordinate mother system (isometric drawing) showing ψ , ϕ , θ directions of positive rotation for daughter system (see 3.2a))	29
Figure 3 – Hierarchical structure among coordinate systems (see 3.2c) and 3.2e))	30
Figure 4 – Rotation ($\phi_g = 15^\circ$) of GANTRY coordinate system X_g , Y_g , Z_g in fixed coordinate system X_f , Y_f , Z_f (see 3.4)	31
Figure 5 – Rotation ($\theta_b = 15^\circ$) of BEAM LIMITING DEVICE or DELINEATOR coordinate system X_b , Y_b , Z_b in GANTRY coordinate system X_g , Y_g , Z_g , and resultant rotation of RADIATION FIELD or DELINEATED RADIATION FIELD of dimensions FX and FY (see 3.5)	32
Figure 6 – Displacement of image intensifier type X-RAY IMAGE RECEPTOR coordinate system origin, I_r , in GANTRY coordinate system, by $R_x = -8$, $R_y = +10$, $R_z = -40$ (see 3.7)	33
Figure 7 – Rotation ($\theta_w = 270^\circ$) and translation of WEDGE FILTER coordinate system X_w , Y_w , Z_w in BEAM LIMITING DEVICE coordinate system X_b , Y_b , Z_b , the BEAM LIMITING DEVICE coordinate system having a rotation $\theta_b = 345^\circ$ (see 3.6)	34
Figure 8 – Rotation ($\theta_r = 90^\circ$) and displacement of X-RAY IMAGE RECEPTOR coordinate system X_r , Y_r , Z_r in GANTRY coordinate system X_g , Y_g , Z_g (see 3.7)	35
Figure 9 – Rotation ($\theta_s = 345^\circ$) of PATIENT SUPPORT coordinate system X_s , Y_s , Z_s in fixed coordinate system X_f , Y_f , Z_f (see 3.8)	36
Figure 10 – Table top eccentric coordinate system rotation θ_e in PATIENT SUPPORT coordinate system which has been rotated by θ_s in the fixed coordinate system with $\theta_e = 360^\circ - \theta_s$ (see 3.9 and 3.10)	37
Figure 11a – Table top displaced below ISOCENTRE by $T_z = -20$ cm (see 3.9 and 3.10)	37
Figure 11b – Table top coordinate system displacement $T_x = +5$, $T_y = L_e + 10$ in PATIENT SUPPORT coordinate system X_s , Y_s , Z_s rotation ($\theta_s = 330^\circ$) in fixed coordinate system X_f , Y_f , Z_f (see 3.9 and 3.10)	38
Figure 11c – Table top coordinate system rotation ($\theta_e = 30^\circ$) about table top eccentric system. PATIENT SUPPORT rotation ($\theta_s = 330^\circ$) in fixed coordinate system $T_x = 0$, $T_y = L_e$ (see 3.9 and 3.10)	38
Figure 12a – Example of BEAM LIMITING DEVICE scale, pointer on mother system (GANTRY), scale on daughter system (BEAM LIMITING DEVICE), viewed from ISOCENTRE (see 3.2f)2) and Clause 4)	39
Figure 12b – Example of BEAM LIMITING DEVICE scale, pointer on daughter system (BEAM LIMITING DEVICE), scale on mother system (GANTRY), viewed from ISOCENTRE (see 3.2f)2) and Clause 4)	40

Figure 12c – Examples of scales (see Clause 4)	40
Figure 13a – Rotary GANTRY (adapted from IEC 60601-2-1) with identification of axes 1 to 8, directions 9 to 13, and dimensions 14 and 15 (see Clause 5)	41
Figure 13b – ISOCENTRIC RADIOTHERAPY SIMULATOR or TELERADIOTHERAPY EQUIPMENT, with identification of axes 1; 4 to 6; 19, of directions 9 to 12; 16 to 18 and of dimensions 14; 15 (see Clause 5)	42
Figure 13c – View from radiation source of teleradiotherapy radiation field or radiotherapy simulator delineated radiation field (see Clause 5)	43
Figure 14a – Example of ISOCENTRIC TELERADIOTHERAPY EQUIPMENT (see 7.2 and 7.4)	44
Figure 14b – Example of ISOCENTRIC RADIOTHERAPY SIMULATOR equipment (see 7.2)	45
Figure 15a – Rotated ($\theta_b = 30^\circ$) symmetrical rectangular RADIATION FIELD ($F_X \times F_Y$) at NORMAL TREATMENT DISTANCE, viewed from ISOCENTRE looking toward RADIATION SOURCE (see 7.3)	46
Figure 15b – Same rotated ($\theta_b = 30^\circ$) symmetrical rectangular RADIATION FIELD ($F_X \times F_Y$) at NORMAL TREATMENT DISTANCE, viewed from RADIATION SOURCE (see 7.3)	46
Figure 16a – Rectangular and symmetrical RADIATION FIELD or DELINEATED RADIATION FIELD, viewed from RADIATION SOURCE (see 7.5)	47
Figure 16b – Rectangular and asymmetrical in Y_b RADIATION FIELD or DELINEATED RADIATION FIELD, viewed from RADIATION SOURCE (see 7.5)	47
Figure 16c – Rectangular and asymmetrical in X_b RADIATION FIELD or DELINEATED RADIATION FIELD, viewed from RADIATION SOURCE (see 7.5)	48
Figure 16d – Rectangular and asymmetrical in X_b and Y_b RADIATION FIELD or DELINEATED RADIATION FIELD, viewed from RADIATION SOURCE (see 7.5)	48
Figure 16e – Rectangular and symmetrical RADIATION FIELD, rotated by $\theta_b = 30^\circ$, viewed from RADIATION SOURCE (see 7.5)	49
Figure 16f – Rectangular and asymmetrical in Y_b RADIATION FIELD, rotated by $\theta_b = 30^\circ$, viewed from RADIATION SOURCE (see 7.5)	49
Figure 16g – Rectangular and asymmetrical in X_b RADIATION FIELD, rotated by $\theta_b = 30^\circ$, viewed from RADIATION SOURCE (see 7.5)	50
Figure 16h – Rectangular and asymmetrical in X_b and Y_b RADIATION FIELD, rotated by $\theta_b = 30^\circ$, viewed from RADIATION SOURCE (see 7.5)	51
Figure 16i – Irregular multi-element (multileaf) contiguous RADIATION FIELD, viewed from RADIATION SOURCE, with element motion in X_b direction (see 7.5)	52
Figure 16j – Irregular multi-element (multileaf) two-part RADIATION FIELD, viewed from RADIATION SOURCE, with element motion in X_b direction (see 7.5)	53
Figure 16k – Irregular multi-element (multileaf) contiguous RADIATION FIELD, viewed from RADIATION SOURCE, with element motion in Y_b direction (see 7.5)	54
Figure 17a – PATIENT coordinate system (PATIENT is supine)	55
Figure 17b – Rotation of PATIENT coordinate system	55
Figure 18 – Table top pitch rotation of table top coordinate system X_t , Y_t , Z_t (see 3.10 and 7.8.4)	56
Figure 19 – Table top roll rotation of table top coordinate system X_t , Y_t , Z_t (see 3.10 and 7.8.5)	56
Figure B.1 – Coordinate transformations between IEC and DICOM PATIENT coordinates	64

Table 1 – ME EQUIPMENT movements and designations 19

Table 2 – Individual coordinate systems..... 26

Table A.1 – Rotation matrices 58

This document is a preview generated by EVS

INTRODUCTION

RADIOTHERAPY is performed in medical centres where a variety of ME EQUIPMENT from different MANUFACTURERS is usually concentrated in the RADIOTHERAPY department. In order to plan and simulate the TREATMENT, set up the PATIENT and direct the RADIATION BEAM, such ME EQUIPMENT can be put in different angular and linear positions and, in the case of MOVING BEAM RADIOTHERAPY, can be rotated and translated during the IRRADIATION of the PATIENT. It is essential that the position of the PATIENT, and the dimensions, directions, and qualities of the RADIATION BEAM prescribed in the treatment plan, be set up or varied by programmes on the radiotherapy EQUIPMENT with accuracy and without misunderstanding. Standard identification and scaling of coordinates is required for ME used in RADIOTHERAPY, including RADIOTHERAPY SIMULATORS and ME EQUIPMENT used to take images during or in connection with RADIOTHERAPY, because differences in the marking and scaling of similar movements on the various types of ME EQUIPMENT used in the same department may increase the probability of error. In addition, data from ME EQUIPMENT used to evaluate the tumour region, such as ultrasound, X-ray, CT and MRI should be presented to the treatment planning system in a form which is consistent with the RADIOTHERAPY coordinate system. Coordinate systems for individual geometrical parameters are required in order to facilitate the mathematical transformation of points and vectors from one coordinate system to another.

A goal of this standard is to avoid ambiguity, confusion, and errors which could be caused when using different types of ME EQUIPMENT. Hence, its scope applies to all types of TELERADIOTHERAPY ME EQUIPMENT, RADIOTHERAPY SIMULATORS, information from diagnostic ME EQUIPMENT when used for RADIOTHERAPY, recording and verification equipment, and to data input for the TREATMENT PLANNING process.

Movement nomenclature is classified as defined terms according to IEC/TR 60788:2004 as well as terms defined in IEC 60601-2-1 and IEC 60601-2-29 (see index of defined terms).

This standard is issued as a publication separate from the IEC 60601 series of safety standards. It is not a safety code and does not contain performance requirements. Thus, the present requirements will not appear in future editions of the IEC 60601-2 series, which deals exclusively with safety requirements.

IEC 60601-2-1, IEC 60601-2-11, IEC 60601-2-29, IEC 60976, IEC 60977, IEC 61168 and IEC 61170 include ME EQUIPMENT movements and scale conventions. A number of changes and additions have been made in this standard.

A major value of a standard coordinate system is its contribution to safety in RADIOTHERAPY TREATMENT PLANNING. The scales that are demonstrated in this standard are consistent with the coordinate systems described herein. USERS may use other scale conventions. It is anticipated that MANUFACTURERS will normally employ the scale conventions of this standard for new ME EQUIPMENT.

It is anticipated that future amendments may address the following:

- three-dimensional RADIOTHERAPY SIMULATORS;
- CT type RADIOTHERAPY SIMULATORS.

Amendment 2, published in 2007, had extended the rotation of the PATIENT support devices around the Z-axis in the IEC fixed coordinate system to two additional rotations – rolling around the PATIENT'S longitudinal axis and pitching around the patient's transversal axis.

The use of the two new additional degrees of freedom (pitch and roll) generalizes the coordinate systems to include systematically 3 rotations and 3 translations, therefore supporting 6 degrees of freedom in a systematic way. Modern patient support devices with 6 degrees of freedom can use a combined translation and rotation to get the same result as the eccentric table top rotation. When changing table position data using the new IEC systems,

the definition of isocentric rotations is sufficient to transfer all treatment-related information. The eccentric table top coordinate system is however maintained for backward compatibility.

NOTE It is quite common in proton therapy to use a treatment chair, where the PATIENT can be rotated and tilted, while the beam line has a fixed direction.

This document is a preview generated by EVS

RADIOTHERAPY EQUIPMENT – COORDINATES, MOVEMENTS AND SCALES

1 Scope and object

This International Standard applies to equipment and data related to the process of TELERADIOtherapy, including PATIENT image data used in relation with RADIOTHERAPY TREATMENT PLANNING SYSTEMS, RADIOTHERAPY SIMULATORS, isocentric GAMMA BEAM THERAPY EQUIPMENT, isocentric medical ELECTRON ACCELERATORS, and non-isocentric equipment when relevant.

The object of this standard is to define a consistent set of coordinate systems for use throughout the process of TELERADIOtherapy, to define the marking of scales (where provided), to define the movements of ME EQUIPMENT used in this process, and to facilitate computer control when used.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment*

IEC 60601-2-1:2009, *Medical electrical equipment – Part 2-1: Particular requirements for the basic safety and essential performance of electron accelerators in the range 1 MeV to 50 MeV*

IEC 60601-2-11:1997, *Medical electrical equipment – Part 2: Particular requirements for the safety of gamma beam therapy equipment*

IEC 60601-2-29:2008, *Medical electrical equipment – Part 2-29: Particular requirements for the basic safety and essential performance of radiotherapy simulators*

IEC 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 62083:2009, *Medical electrical equipment – Requirements for the safety of radiotherapy treatment planning systems*

3 Coordinate systems

3.1 General

An individual coordinate system is assigned to each major part of the ME EQUIPMENT which can potentially be moved in relation to another part, as illustrated in Figure 1a and summarized in Table 1. Furthermore a fixed reference system is defined. Each major part (e.g. GANTRY, RADIATION HEAD) is always stationary with respect to its own coordinate system.