In vitro diagnostikasüsteemid. Nõuded diabeetikute enesekontrolli veresuhkru jälgimissüsteemidele

In vitro diagnostic test systems - Requirements for bloodglucose monitoring systems for self-testing in itus Ordinario de la constante managing diabetes mellitus (ISO 15197:2013)



EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

	This Estonian standard EVS-EN ISO 15197:2013
sisaldab Euroopa standardi EN ISO 15197:2013	consists of the English text of the European standard
ingliskeelset teksti.	EN ISO 15197:2013.
, , , , , , , , , , , , , , , , , , , ,	This standard has been endorsed with a notification
avaldamisega EVS Teatajas.	published in the official bulletin of the Estonian Centre
	for Standardisation.
_	
	Date of Availability of the European standard is
·	22.05.2013.
kättesaadavaks 22.05.2013.	
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for
	Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.100.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega: Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation: Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 15197

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2013

ICS 11.100.10

Supersedes EN ISO 15197:2003

English Version

In vitro diagnostic test systems - Requirements for bloodglucose monitoring systems for self-testing in managing diabetes mellitus (ISO 15197:2013)

Systèmes d'essais de diagnostic in vitro - Exigences relatives aux systèmes d'autosurveillance de la glycémie destinés à la prise en charge du diabète sucré (ISO 15197:2013)

Testsysteme für die In-vitro-Diagnostik - Anforderungen an Blutzuckermesssysteme zur Eigenanwendung beim Diabetes mellitus (ISO 15197:2013)

This European Standard was approved by CEN on 15 May 2013.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

This document (EN ISO 15197:2013) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with the Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2013, and conflicting national standards shall be withdrawn at the latest by November 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15197:2003.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 15197:2013 has been approved by CEN as EN ISO 15197:2013 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 98/79/EC in vitro diagnostic medical devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 98/79/EC *in vitro* diagnostic medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and Directive 98/79/EC in vitro diagnostic medical devices

Clauses of this international Standard	Essential requirements (ERs) of Directive 98/79/EC	Qualifying Comments/Notes
4.2, 6	A.3	
4.3, 4.4	A.4	Aspects related to mechanical stress are not covered
4.3, 4.4, 6.5, 7	A.5	Packaging aspects are not covered.
4.3, 5.2, 5.3, 5.4, 5.5, 5.6, 5.8, 5.10, 5.11, 5.12	B.3.3	
5.7	B.3.4	0
4.4	B.3.6	
6, 7	B.4.1	3
6	B.6.1	
5.2	B.6.3	
5.3, 5.6	B.6.4.1	"Life time resistance to mechanical stress" is not covered.
5.2	B.6.4.4	Clause 5.2 only partially covers this ER, as it deals only with electrical terminals and connectors.
4.4, 7, 8	B.7.1	7
7	B.8.1	The sixth paragraph of this ER (translation into other languages) is not covered.
7	B.8.2	Covered by reference to EN ISO 18113-1.

7	B.8.3	This aspect is covered by normative reference to EN ISO 18113-4 and EN ISO 18113-5. However, essential requirement B.8.3 of Directive 98/79/EC should be consulted for a comprehensive list of the information required
7	B.8.4	Not all the requirements of ER B.8.4 are covered, but only those required by EN ISO 18113-1, EN ISO 18113-4 and EN ISO 18113-5, related to IVD medical devices for self – testing
7	B.8.4(a)	If the manufacturer is not located in the EU, the manufacturer is required to designate an "EC Authorised Representative" established in the EU. In such cases and to comply with this ER, the name and address of the Authorised Representative are required
7	B.8.4(d)	Full compliance with this ER requires the use of the symbol [LOT] and the symbol [SN], as applicable.
7	B.8.6	Covered by normative reference to EN ISO 18113-4 and EN ISO 18113-5.
4.2, 7	B.8.7	Not all the requirements of this ER are covered, but only those required by the labelling standards EN ISO 18113-1, EN ISO 18113-4 and EN ISO 18113-5, related to IVD medical devices for self – testing

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

Con	tents	Page
Forew	ord	iv
Intro	uction	v
1	Scope	
_	0.	
2	Normative references	
3	Terms and definitions	2
4	Design and development	8
	4.1 General requirements	8
	4.2 Metrological traceability	
	4.3 Safety and risk management	
	4.4 Ergonomics and human factors	
	4.5 User verification requirements	10
5	Safety and reliability testing	
	5.1 General requirements	
	5.2 Protection against electric shock	
	5.3 Protection against mechanical hazards	
	5.4 Electromagnetic compatibility	
	5.5 Resistance to heat	
	5.6 Resistance to moisture and liquids	
	5.7 Protection against liberated gases, explosion and implosion.5.8 Meter components.	
	5.9 Performance test	
	5.10 Mechanical resistance to vibration and shock	
	5.11 Equipment temperature exposure limits for storage	
	5.12 Equipment humidity exposure limits for storage	13
6	Analytical performance evaluation	
O	6.1 General requirements	14 1./.
	6.2 Measurement precision	14 16
	6.3 System accuracy	19
	6.4 Influence quantities	25
	6.5 Stability of reagents and materials	30
7	Information supplied by the manufacturer	
,	7.1 General requirements	30 30
	7.2 Performance characteristics	31
	7.3 Options for supplying instructions for use	31
8	User performance evaluation	
0	8.1 General requirements	
	8.2 Acceptance criteria and evaluation of results	
	8.3 Selection and preparation of subjects	32
	8.4 Execution of study protocol	32
	8.5 Glucose reference values	33
	8.6 Human factors	33
	8.7 Data analysis and presentation of results	
	8.8 Evaluation of instructions for use	34
Annex	A (informative) Possible interfering substances	35
	B (informative) Traceability chain	
	C (informative) Rationale for the analytical performance requirements	
	graphy	

Introduction

Blood-glucose monitoring systems are *in vitro* diagnostic medical devices used predominantly by individuals affected by diabetes mellitus. Diabetes mellitus is caused by a deficiency in insulin secretion or by insulin resistance leading to abnormally high concentrations of glucose in the blood, which may result in acute and chronic health complications. When used properly, a glucose monitoring system allows the user to monitor and take action to control the concentration of glucose present in the blood.

This International Standard is intended for blood-glucose monitoring systems used by lay persons. The primary objectives are to establish requirements that result in acceptable performance and to specify procedures for demonstrating conformance to this International Standard.

Minimum performance criteria for blood-glucose monitoring systems were established from the analytical requirements (precision and trueness) for individual glucose measurement results. "System accuracy" is the term used in this International Standard to communicate the analytical capability of a blood-glucose monitoring system to the intended users (i.e. lay persons), who would not be familiar with metrological terms commonly used in laboratory medicine. System accuracy describes the ability of a glucose monitoring system to produce measurement results that agree with true glucose values when the system is used as intended. The concept of "system accuracy" includes measurement bias and measurement precision.

The requirements for system accuracy are based on three considerations:

- the effectiveness of current technology for monitoring patients with diabetes mellitus;
- recommendations of diabetes researchers as well as existing product standards and regulatory guidelines; and
- the state-of-the-art of blood-glucose monitoring technology.

In arriving at the performance requirements specified in the second edition of this International Standard, desirable goals had to be weighed against the capabilities of existing blood-glucose monitoring technology. The revised performance criteria in this edition are the result of improvements in technology since publication of the first edition. The considerations that formed the basis for the minimum acceptable analytical performance of a blood-glucose measuring device intended for self-monitoring are described in Annex C.

Requirements that are unique to self-monitoring devices for blood-glucose are addressed in this International Standard. Requirements that apply in general to all *in vitro* diagnostic medical devices are incorporated by reference to other standards where appropriate.

Although this International Standard does not apply to glucose monitoring systems that provide measured values on an ordinal scale (e.g. visual, semiquantitative measurement procedures) or medical devices that measure blood-glucose continuously for self-monitoring, it may be useful as a guide for developing procedures to evaluate the performance of such systems.

In vitro diagnostic test systems — Requirements for bloodglucose monitoring systems for self-testing in managing diabetes mellitus

1 Scope

This International Standard specifies requirements for *in vitro* glucose monitoring systems that measure glucose concentrations in capillary blood samples, for specific design verification procedures and for the validation of performance by the intended users. These systems are intended for self-measurement by lay persons for management of diabetes mellitus.

This International Standard is applicable to manufacturers of such systems and those other organizations (e.g. regulatory authorities and conformity assessment bodies) having the responsibility for assessing the performance of these systems.

This International Standard does not:

- provide a comprehensive evaluation of all possible factors that could affect the performance of these systems,
- pertain to glucose concentration measurement for the purpose of diagnosing diabetes mellitus,
- address the medical aspects of diabetes mellitus management,
- apply to measurement procedures with measured values on an ordinal scale (e.g. visual, semiquantitative measurement procedures), or to continuous glucose monitoring systems,
- apply to glucose meters intended for use in medical applications other than self-testing for the management of diabetes mellitus

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 13485, Medical devices — Quality management systems — Requirements for regulatory purposes

ISO 14971, Medical devices — Application of risk management to medical devices

ISO 17511, In vitro diagnostic medical devices — Measurement of quantities in biological samples — Metrological traceability of values assigned to calibrators and control materials

ISO 18113-1, In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 1: Terms, definitions and general requirements

ISO 18113-4, In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 4: In vitro diagnostic reagents for self-testing

ISO 18113-5, In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 5: In vitro diagnostic instruments for self-testing

ISO 23640, In vitro diagnostic medical devices — Evaluation of stability of in vitro diagnostic reagents

IEC 60068-2-64, Environmental testing — Part 2-64: Tests — Test Fh: Vibration, broadband random and guidance