

**IEC 60601-1**  
(Third edition - 2005)

**Medical electrical equipment –  
Part 1: General requirements for basic safety  
and essential performance**

**CORRIGENDUM 2**

**7.9.3 Technical description**

**7.9.3.1 General**

*Delete the first bullet in the list that immediately follows the first paragraph of this subclause, which reads:*

- the information required in 7.2;

**Table D.2 – Safety signs**

*For rows 5, 6 and 7, instead of:*

5		ISO 7010-P017	Pushing prohibited
6		ISO 7010-P018	Sitting prohibited
7		ISO 7010-P019	Stepping prohibited

read:

5		ISO 7010-P017	No pushing
6		ISO 7010-P018	No sitting
7		ISO 7010-P019	No stepping on surface

**CEI 60601-1**  
(Troisième édition - 2005)

**Appareils électromédicaux –  
Partie 1: Exigences générales pour la sécurité de base  
et les performances essentielles**

**CORRIGENDUM 2**

**7.9.3 Description technique**

**7.9.3.1 Généralités**

*Supprimer le premier tiret de la liste qui suit le premier alinéa:*

- les informations exigées en 7.2;

**Tableau D.2 – Signes de sécurité**

*Aux lignes 5, 6 et 7, au lieu de:*

5		ISO 7010-P017	Interdiction de pousser
6		ISO 7010-P018	Interdiction de s'asseoir
7		ISO 7010-P019	Interdiction de monter

lire:

5		ISO 7010-P017	Interdiction de pousser
6		ISO 7010-P018	Interdiction de s'asseoir
7		ISO 7010-P019	Interdiction de monter dessus