



**INTERNATIONAL STANDARD ISO 407:1991**  
**TECHNICAL CORRIGENDUM 1**

Published 1999-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

# **Small medical gas cylinders — Pin-index yoke-type valve connections**

## **TECHNICAL CORRIGENDUM 1**

*Petites bouteilles à gaz médicaux — Raccords de robinets du type à étrier avec ergots de sécurité*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to International Standard ISO 407:1991 was prepared by Technical Committee ISO/TC 58, *Gas cylinders*, Subcommittee SC 2, *Cylinder fittings*.

---

*Page 7*

**Figure 5**, dimension  $q$ , should read  $24^{+0,5}_0$  (instead of 24 max.).

*Page 8*

**Figure 6**, dimension  $q_1$ , should read  $24^{+0,5}_0$  (instead of 24 max.).