



INTERNATIONAL STANDARD ISO 11979-8:1999
TECHNICAL CORRIGENDUM 1

Published 2003-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Ophthalmic implants — Intraocular lenses —
Part 8:
Fundamental requirements

TECHNICAL CORRIGENDUM 1

Implants ophtalmiques — Lentilles intraoculaires —

Partie 8: Exigences fondamentales

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 11979-8:1999 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

Page 2, Clause 8

At the end of this clause add the following text:

IOLs sterilized by ethylene oxide shall have residual ethylene oxide levels and release rates in accordance with the limits given in ISO 10993-7.