



**INTERNATIONAL STANDARD ISO 10282:2002**  
**TECHNICAL CORRIGENDUM 1**

Published 2005-06-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

## **Single-use sterile rubber surgical gloves — Specification**

### **TECHNICAL CORRIGENDUM 1**

*Gants en caoutchouc à usage chirurgical, stériles, non réutilisables — Spécifications*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 10282:2002 was prepared by Technical Committee ISO/TC 45, *Rubber and rubber products*, Subcommittee SC 4, *Products (other than hoses)*.

---

#### *Page 1, Clause 2*

Replace the reference to ISO 4648:1991 by ISO 23529:2004, *Rubber — General procedures for preparing and conditioning test pieces for physical test methods*.

#### *Page 2, Subclause 3.4*

Add the following sentence at the end of Note 1:

Powder-free is also referred to as “powderless”, “no powder” or “non-powdered” or other words to that effect.

#### *Page 3*

In Subclause 5.1, line 2, replace “acceptable” by “acceptance”.

In Table 1, second column, third row, replace “I” by “G-I”.

In Subclause 6.1, fifth paragraph, and in the note at the end of the subclause, replace ISO 4648 by ISO 23529.

---

**ICS 11.140; 83.140.99**

**Ref. No. ISO 10282:2002/Cor.1:2005(E)**

© ISO 2005 – All rights reserved

Published in Switzerland

*Page 4, Subclause 6.3.1*

Replace “taking three test pieces” by “taking three type 2 dumb-bell test pieces”.