

English version
Version Française
Deutsche Fassung

Dentistry - Number coding system for rotary instruments - Part 1: General characteristics (ISO 6360-1:2004/Cor 1:2007)

Art dentaire - Système de codification numérique pour instruments rotatifs - Partie 1: Caractéristiques générales (ISO 6360-1:2004/Cor 1:2007)

Zahnheilkunde - Nummernsystem für rotierende Instrumente - Teil 1: Allgemeine Grundlagen (ISO 6360-1:2004/Cor 1:2007)

This corrigendum becomes effective on 5 December 2007 for incorporation in the three official language versions of the EN.

Ce corrigendum prendra effet le 5 décembre 2007 pour incorporation dans les trois versions linguistiques officielles de la EN.

Die Berichtigung tritt am 5. Dezember 2007 zur Einarbeitung in die drei offiziellen Sprachfassungen der EN in Kraft.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

English version

Endorsement Notice

The text of ISO 6360-1:2004/Cor.1:2007 has been approved by CEN as a European Corrigendum without any modifications.

Version française

Notice d'entérinement

Le texte de l'ISO 6360-1:2004/Cor.1:2007 a été approuvé par le CEN comme Corrigendum européen sans aucune modification.



INTERNATIONAL STANDARD ISO 6360-1:2004 TECHNICAL CORRIGENDUM 1

Published 2007-09-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Dentistry — Number coding system for rotary instruments — Part 1: General characteristics

TECHNICAL CORRIGENDUM 1

Art dentaire — Système de codification numérique pour instruments rotatifs —

Partie 1: Caractéristiques générales

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 6360-1:2004 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

Page 14

5.5 Overall length

Paragraph 4 shall read:

For hand-held root-canal instruments, the code number for the sixth digit shall be 0.

A new paragraph 5 is added:

For powered (electrical) root-canal instruments of Type 1 and Type 2 of ISO 1797-1:1992 and ISO 1792-2:1992, the codes of Table 5 are used:

2 = miniature 3 = short 4 = standard 5 = long 6 = extra long