



INTERNATIONAL STANDARD ISO 8655-7:2005
TECHNICAL CORRIGENDUM 1

Published 2008-12-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Piston-operated volumetric apparatus —

Part 7:

Non-gravimetric methods for the assessment of equipment performance

TECHNICAL CORRIGENDUM 1

Appareils volumétriques à piston —

Partie 7: Méthodes non gravimétriques pour l'estimation de la performance d'équipement

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 8655-7:2002 was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, Subcommittee SC 6, *Laboratory and volumetric ware*.

Page 5, 8.1

At the end of the first sentence, add the following text:

"(or 27 °C in the case where the piston-operated volumetric apparatus under test has been adjusted for this standard reference temperature)."