



**ISO/IEC GUIDE 98-3/Suppl.1:2008**  
TECHNICAL CORRIGENDUM 1

Published 2009-05-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION  
INTERNATIONAL ELECTROTECHNICAL COMMISSION • МЕЖДУНАРОДНАЯ ЭЛЕКТРОТЕХНИЧЕСКАЯ КОМИССИЯ • COMMISSION ÉLECTROTECHNIQUE INTERNATIONALE

## **Uncertainty of measurement —**

**Part 3:**

**Guide to the expression of uncertainty in measurement  
(GUM:1995)**

**Supplement 1:**

**Propagation of distributions using a Monte Carlo method**

TECHNICAL CORRIGENDUM 1

*Incertitude de mesure —*

*Partie 3: Guide pour l'expression de l'incertitude de mesure (GUM:1995)*

*Supplément 1: Propagation de distributions par une méthode de Monte Carlo*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO/IEC Guide 98-3/Suppl.1:2008 was prepared by Working Group 1 of the JCGM.

---

*Page 68, Table C.5*

Replace Table C.5 with the following. The symbols have been corrected.

---

**ICS 17.020**

**Ref. No. ISO/IEC GUIDE 98-3/  
Suppl.1:2008/Cor.1:2009(E)**

**Table C.5 — A  $t$ -distribution pseudo-random number generator (Clause C.6)**

Input parameter	
$\nu$	Degrees of freedom
Output parameter	
$t$	Draw from a $t$ -distribution with $\nu$ degrees of freedom
Computation	
a) Generate two draws $r_1$ and $r_2$ independently from the rectangular distribution $R(0, 1)$ b) If $r_1 < 1/2$ , form $t = 1/(4r_1 - 1)$ and $v = r_2/t^2$ ; otherwise form $t = 4r_1 - 3$ and $v = r_2$ c) If $v < 1 -  t /2$ or $v < (1 + t^2/\nu)^{-(\nu+1)/2}$ , accept $t$ as a draw from the $t$ -distribution; otherwise repeat from step a)	

NOTE  $\nu$  must be greater than two for the standard deviation of the  $t$ -distribution with  $\nu$  degrees of freedom to be finite.