

INTERNATIONAL
STANDARD

ISO/IEEE
11073-
10424

First edition
2016-06-15

Corrigendum 1
2018-01

**Health informatics — Personal health
device communication —**

Part 10424:

**Device specialization — Sleep apnoea
breathing therapy equipment (SABTE)**

TECHNICAL CORRIGENDUM 1

*Informatique de la santé — Communication entre dispositifs de santé
personnels —*

*Partie 10424: Spécialisation de dispositif — Équipement de thérapie
respiratoire de l'apnée du sommeil (SABTE)*

RECTIFICATIF TECHNIQUE 1



Reference number
ISO/IEEE 11073-10424:2016/Cor 1:2018(E)

© IEEE 2017



COPYRIGHT PROTECTED DOCUMENT

© IEEE 2017

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO or IEEE at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Published in Switzerland

Institute of Electrical and Electronics Engineers, Inc
3 Park Avenue, New York
NY 10016-5997, USA

stds.ipr@ieee.org
www.ieee.org

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is called to the possibility that implementation of this standard may require the use of subject matter covered by patent rights. By publication of this standard, no position is taken with respect to the existence or validity of any patent rights in connection therewith. ISO/IEEE is not responsible for identifying essential patents or patent claims for which a license may be required, for conducting inquiries into the legal validity or scope of patents or patent claims or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance or a Patent Statement and Licensing Declaration Form, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from ISO or the IEEE Standards Association.

ISO/IEEE 11073-10424 was prepared by the IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society (as IEEE Std 11073-10427-2014/Cor 1-2017). It was adopted by Technical Committee ISO/TC 215, *Health informatics*, in parallel with its approval by the ISO member bodies, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE. IEEE is responsible for the maintenance of this document with participation and input from ISO member bodies.

Health informatics—Personal health device communication

Part 10424: Device Specialization— Sleep Apnoea Breathing Therapy Equipment (SABTE)

Corrigendum 1

Sponsor

IEEE 11073™ Standards Committee
of the
IEEE Engineering in Medicine and Biology Society

Approved 23 March 2017

IEEE-SA Standards Board

Abstract: Within the context of the ISO/IEEE 11073 family of standards for device communication, a normative definition of the communication between sleep apnoea breathing therapy equipment (SABTE) devices (agents) and managers (e.g., cell phones, personal computers, personal health appliances, set-top boxes), in a manner that enables plug-and-play interoperability, is established in IEEE Std 11073-10424-2014. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. IEEE Std 11073-10424-2014 defines a common core of communication functionality for SABTE. In this context, SABTE is defined as a device that is intended to alleviate the symptoms of a patient who suffers from sleep apnoea by delivering a therapeutic breathing pressure to the patient. SABTE is primarily used in the home health-care environment by a lay operator without direct professional supervision. This corrigendum corrects errors that have been identified in IEEE Std 11073-10424-2014 to make it easier to implement the standard in an interoperable fashion.

Keywords: IEEE 11073-10424™, medical device communication, personal health devices, SABTE, sleep apnoea breathing therapy equipment

The Institute of Electrical and Electronics Engineers, Inc.
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2017 by The Institute of Electrical and Electronics Engineers, Inc.
All rights reserved. Published 23 May 2017. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: ISBN 978-1-5044-4053-0 STD22607

IEEE prohibits discrimination, harassment and bullying. For more information, visit <http://www.ieee.org/web/aboutus/whatis/policies/p9-26.html>. No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page, appear in all standards and may be found under the heading “Important Notices and Disclaimers Concerning IEEE Standards Documents.” They can also be obtained on request from IEEE or viewed at <http://standards.ieee.org/IPR/disclaimers.html>.

Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents (standards, recommended practices, and guides), both full-use and trial-use, are developed within IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (“IEEE-SA”) Standards Board. IEEE (“the Institute”) develops its standards through a consensus development process, approved by the American National Standards Institute (“ANSI”), which brings together volunteers representing varied viewpoints and interests to achieve the final product. IEEE Standards are documents developed through scientific, academic, and industry-based technical working groups. Volunteers in IEEE working groups are not necessarily members of the Institute and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE Standards do not guarantee or ensure safety, security, health, or environmental protection, or ensure against interference with or from other devices or networks. Implementers and users of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.

IEEE does not warrant or represent the accuracy or content of the material contained in its standards, and expressly disclaims all warranties (express, implied and statutory) not included in this or any other document relating to the standard, including, but not limited to, the warranties of: merchantability; fitness for a particular purpose; non-infringement; and quality, accuracy, effectiveness, currency, or completeness of material. In addition, IEEE disclaims any and all conditions relating to: results; and workmanlike effort. IEEE standards documents are supplied “AS IS” and “WITH ALL FAULTS.”

Use of an IEEE standard is wholly voluntary. The existence of an IEEE standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE should be considered the approved IEEE standard.

Official statements

A statement, written or oral, that is not processed in accordance with the IEEE-SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, or be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that his or her views should be considered the personal views of that individual rather than the formal position of IEEE.

Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE. However, IEEE does not provide consulting information or advice pertaining to IEEE Standards documents. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its societies and Standards Coordinating Committees are not able to provide an instant response to comments or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in revisions to an IEEE standard is welcome to join the relevant IEEE working group.

Comments on standards should be submitted to the following address:

Secretary, IEEE-SA Standards Board
445 Hoes Lane
Piscataway, NJ 08854 USA

Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not imply compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so.

Copyrights

IEEE draft and approved standards are copyrighted by IEEE under U.S. and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, IEEE does not waive any rights in copyright to the documents.

Photocopies

Subject to payment of the appropriate fee, IEEE will grant users a limited, non-exclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, non-commercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

Every IEEE standard is subjected to review at least every ten years. When a document is more than ten years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit the IEEE Xplore at <http://ieeexplore.ieee.org/> or contact IEEE at the address listed previously. For more information about the IEEE-SA or IEEE's standards development process, visit the IEEE-SA Website at <http://standards.ieee.org>.

Errata

Errata, if any, for all IEEE standards can be accessed on the IEEE-SA Website at the following URL: <http://standards.ieee.org/findstds/errata/index.html>. Users are encouraged to check this URL for errata periodically.

Patents

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE-SA Website at <http://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patents Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

Participants

At the time this standard was submitted to the IEEE-SA Standards Board for approval, the 11073 Personal Health Device Working Group had the following membership:

Daidi Zhong, *Co-Chair*
Michael J. Kirwan, *Co-Chair*
Christoph Fischer, *Vice Chair*

Karsten Aalders
Charles R. Abbruscato
Nabil Abujbara
Maher Abuzaid
James Agnew
Haidar Ahmad
Manfred Aigner
Jorge Alberola
Murtaza Ali
Rolf Ambuehl
David Aparisi
Paolo Ariano
Lawrence Arne
Diego B. Arquillo
Serafin Arroyo
Muhammad Asim
Merat Bagha
Doug Baird
David Baker
Anindya Bakshi
Ananth Balasubramanian
Sunlee Bang
M. Jonathan Barkley
Gilberto Barrón
David Bean
John Bell
Rudy Belliard
Kathryn M. Bennett
Daniel Bernstein
George A. Bertos
Chris Biernacki
Ola Björnsne
Thomas Blackadar
Marc Blanchet
Thomas Bluethner
Douglas P. Bogia
Xavier Boniface
Shannon Boucousis
Julius Broma
Lyle G. Bullock, Jr.
Bernard Burg
Chris Burns
Anthony Butt
Jeremy Byford-Rew
Satya Calloji
Xiaoying Cao
Carole C. Carey
Craig Carlson
Santiago Carot-Nemesio
Randy W. Carroll
Simon Carter

Seungchul Chae
Rahul Chauhan
James Cheng
Peggy Chien
David Chiu
Jinyong Choi
Chia-Chin Chong
Saeed A. Choudhary
Jinhan Chung
Malcolm Clarke
John A. Cogan
John T. Collins
Cory Condek
Todd H. Cooper
David Cornejo
Douglas Coup
Nigel Cox
Hans Crommenacker
Tomio Crosley
Allen Curtis
Ndifor Cyril Fru
Jesús Daniel Trigo
Eyal Dassau
David Davenport
Russell Davis
Sushil K. Deka
Ciro de la Vega
Pedro de-las-Heras-Quiros
Jim Dello Stritto
Matthew d'Entremont
Kent Dicks
Hyoungdo Do
Alistair Donaldson
Xiaolian Duan
Brian Dubreuil
Sourav Dutta
Jakob Ehrensvar
Fredrik Einberg
Michihiro Enokida
Javier Escayola Calvo
Mark Estes
Leonardo Estevez
Roger Feeley
Hailing Feng
Bosco T. Fernandes
Christoph Fischer
Morten Flintrup
Joseph W. Forler
Russell Foster
Eric Freudenthal
Matthias Frohner

Ken Fuchs
Jing Gao
Marcus Garbe
John Garguilo
Rick Geimer
Igor Gejdos
Ferenc Gerbovics
Nicolae Goga
Julian Goldman
Raul Gonzalez Gomez
Chris Gough
Channa Gowda
Charles M. Gropper
Amit Gupta
Jeff Guttmacher
Rasmus Haahr
Christian Habermann
Michael Hagerty
Jerry Hahn
Robert Hall
Shu Han
Nathaniel Hamming
Rickey L. Hampton
Sten Hanke
Aki Harma
Jordan Hartmann
Kai Hassing
Wolfgang Heck
Nathaniel Heintzman
Charles Henderson
Jun-Ho Her
Helen B. Hernandez
Takashi Hibino
Timothy L. Hirou
Allen Hobbs
Alex Holland
Arto Holopainen
Kris Holtzclaw
Robert Hoy
Frank Hsu
Anne Huang
Sen-Der Huang
Ron Huby
David Hughes
Robert D. Hughes
Jiyoung Huh
Hugh Hunter
Hitoshi Ikeda
Yutaka Ikeda
Philip O. Isaacson
Atsushi Ito

Michael Jaffe	Alexander Mense	Stefan Sauermann
Praduman Jain	Behnaz Minaei	John Sawyer
Danny Jochelson	Jinsei Miyazaki	Guillaume Schatz
Phaneeth Junga	Erik Moll	Alois Schloegl
Akiyoshi Kabe	Darr Moore	Paul S. Schluter
Steve Kahle	Carsten Mueglitz	Lars Schmitt
Tomio Kamioka	Piotr Murawski	Mark G. Schnell
Kei Kariya	Soundharya Nagasubramanian	Richard A. Schrenker
Andy Kaschl	Jae-Wook Nah	Antonio Scorpiniti
Junzo Kashiwara	Alex Neefus	Kwang Seok Seo
Kohichi Kashiwagi	Trong-Nghia Nguyen-Dobinsky	Riccardo Serafin
Ralph Kent	Michael E. Nidd	Sid Shaw
Laurie M. Kermes	Tetsu Nishimura	Frank Shen
Ikuo Keshi	Jim Niswander	Min Shih
Junhyung Kim	Hongliang Niu	Mazen Shihabi
Minho Kim	Hiroaki Niwamoto	Redmond Shouldice
Min-Joon Kim	Thomas Norgall	Sternly K. Simon
Taekon Kim	Anand Noubade	Marjorie Skubic
Tetsuya Kimura	Yoshiteru Nozoe	Robert Smith
Michael J. Kirwan	Abraham Ofek	Ivan Soh
Alfred Kloos	Brett Olive	Motoki Sone
Jeongmee Koh	Begonya Otal	Emily Sopensky
Jean-Marc Koller	Marco Paleari	Rajagopalan Srinivasan
John Koon	Charles Palmer	Andreas Staubert
Patty Krantz	Bud Panjwani	Nicholas Steblay
Raymond Krasinski	Carl Pantiskas	Lars Steubesand
Alexander Kraus	Harry P. Pappas	John (Ivo) Stivorice
Ramesh Krishna	Hanna Park	Raymond A. Strickland
Geoffrey Kruse	Jong-Tae Park	Chandrasekaran Subramaniam
Falko Kuester	Myungeun Park	Hermann Suominen
Rafael Lajara	Soojun Park	Lee Surprenant
Pierre Landau	Phillip E. Pash	Ravi Swami
Jaechul Lee	TongBi Pei	Ray Sweidan
JongMuk Lee	Lucian Pestritu	Jin Tan
Kyong Ho Lee	Soren Petersen	Yi Tang
Rami Lee	James Petisce	Haruyuyki Tatsumi
Sungkee Lee	Peter Piction	John W. Thomas
Woojae Lee	Michael Pliskin	Jonas Tirén
Yonghee Lee	Varshney Prabodh	Alexandra Todiruta
Joe Lenart	Jeff Price	Janet Traub
Kathryn A. Lesh	Harald Prinzhorn	Gary Tschautscher
Catherine Li	Harry Qiu	Masato Tsuchid
Qiong Li	Arif Rahman	Ken Tubman
Patrick Lichter	Tanzilur Rahman	Yoshihiro Uchida
Jisoon Lim	Steve Ray	Akib Uddin
Joon-Ho Lim	Phillip Raymond	Sunil Unadkat
John Lin	Terrie Reed	Fabio Urbani
Xiaoming Liu	Tim Reilly	Philipp Urbauer
Wei-Jung Lo	Barry Reinhold	Laura Vanzago
Charles Lowe	Brian Reinhold	Alpo Värri
Don Ludolph	Melvin I. Reynolds	Andrei Vasilateanu
Christian Luszick	John G. Rhoads	Dalimar Velez
Bob MacWilliams	Jeffrey S. Robbins	Martha Veleziz
Srikanth Madhurboothswaran	Chris Roberts	Rudi Voon
Miriam L. Makhoul	Moskowitz Robert	Barry Vornbrock
Romain Marmot	Timothy Robertson	Isobel Walker
Sandra Martinez	David Rosales	David Wang
Miguel Martínez	Fatemeh Saki	Jerry P. Wang
de Espronceda Cámara	Bill Saltzstein	Yao Wang
Peter Mayhew	Benedikt Salzbrunn	Yi Wang
Jim McCain	Giovanna Sannino	Steve Warren
László Meleg	Jose A. Santos-Cadenas	Fujio Watanabe

Toru Watsuji
Mike Weng
Kathleen Wible
Paul Williamson
Jan Wittenber
Jia-Rong Wu
Will Wykeham

Ariton Xhafa
Qifeng Yan
Ricky Yang
Melanie S. Yeung
Qiang Yin
Done-Sik Yoo

Jianchao Zeng
Jason Zhang
Zhiqiang Zhang
Thomas Zhao
Daidi Zhong
Miha Zoubek
Szymon Zyskoter

The following members of the individual balloting committee voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Keith Chow
Malcolm Clarke
Christoph Fischer
Randall Groves
Werner Hoelzl

Noriyuki Ikeuchi
Atsushi Ito
Piotr Karocki
H. Moll
Melvin Reynolds
Bartien Sayogo

Walter Struppler
Chandrasekaran Subramaniam
Jan Wittenber
Oren Yuen
Daidi Zhong

When the IEEE-SA Standards Board approved this standard on 23 March 2017, it had the following membership:

Jean-Philippe Faure, *Chair*
Gary Hoffman, *Vice Chair*
John D. Kulick, *Past Chair*
Konstantinos Karachalios, *Secretary*

Chuck Adams
Masayuki Ariyoshi
Ted Burse
Stephen Dukes
Doug Edwards
J. Travis Griffith
Michael Janezic

Thomas Koshy
Joseph L. Koepfinger*
Kevin Lu
Daleep Mohla
Damir Novosel
Ronald C. Petersen
Annette D. Reilly

Robby Robson
Dorothy Stanley
Adrian Stephens
Mehmet Ulema
Phil Wennblom
Howard Wolfman
Yu Yuan

*Member Emeritus

Introduction

This introduction is not part of IEEE Std 11073-10424-2014/Cor 1-2017, IEEE Standard for Part 10424: Device Specialization—Sleep Apnoea Breathing Therapy Equipment (SABTE)—Corrigendum 1.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. Within the context of the ISO/IEEE 11073 family of standards for device communication, IEEE Std 11073-10424-2014 establishes a normative definition of the communication between sleep apnoea breathing therapy equipment (SABTE) devices (agents) and managers (e.g., cell phones, personal computers, personal health appliances, set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability. IEEE Std 11073-10424-2014 defines a common core of communication functionality for SABTE. In this context, SABTE is defined as a device that is intended to alleviate the symptoms of a patient who suffers from sleep apnoea by delivering a therapeutic breathing pressure to the patient. SABTE is primarily used in the home health-care environment by a lay operator without direct professional supervision.

This corrigendum corrects errors that have been identified in the IEEE Std 11073-10424-2014 to make it easier to implement the standard in an interoperable fashion.

Contents

6. Sleep apnoea breathing therapy equipment domain information model.....	12
6.7 Numeric objects.....	12
6.8 Real-time sample array objects.....	19
Annex C (normative) Allocation of identifiers.....	20
C.1 Definitions of terms and codes.....	20
C.2 Systematic derivations of terms and codes	22
Annex E (informative) Protocol data unit examples	27
E.4 GET MDS attributes service	27

Health informatics—Personal health device communication

Part 10424: Device Specialization— Sleep Apnoea Breathing Therapy Equipment (SABTE)

Corrigendum 1

NOTE—The editing instructions contained in this **corrigendum** define how to merge the material contained therein into the existing base standard and its amendments to form the comprehensive standard.

The editing instructions are shown in ***bold italic***. Four editing instructions are used: change, delete, insert, and replace. ***Change*** is used to make corrections in existing text or tables. The editing instruction specifies the location of the change and describes what is being changed by using ~~strike through~~ (to remove old material) and underscore (to add new material). ***Delete*** removes existing material. ***Insert*** adds new material without disturbing the existing material. Insertions may require renumbering. If so, renumbering instructions are given in the editing instruction. ***Replace*** is used to make changes in figures or equations by removing the existing figure or equation and replacing it with a new one. Editing instructions, change markings, and this NOTE will not be carried over into future editions because the changes will be incorporated into the base standard.

6. Sleep apnoea breathing therapy equipment domain information model

6.7 Numeric objects

6.7.3 Apnoea hypopnoea index (AHI)

Change the last word in the following sentence as shown:

The Type attribute is used to distinguish the modality of particular AHI between total AHI (i.e., MDC_SABTE_AHI_TOTAL), uAHI (i.e., MDC_SABTE_AHI_UNCLASS), oAHI (i.e., MDC_SABTE_AHI_OBSTRUC), or cAHI (i.e., MDC_SABTE_AHI_CENTRAL).

Change Table 7 as shown:

Table 7—AHI numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_AHI_TOTAL} or {MDC_PART_PHD_DM, MDC_SABTE_AHI_UNCLASS} or {MDC_PART_PHD_DM, MDC_SABTE_AHI_OBSTRUC} or {MDC_PART_PHD_DM, MDC_SABTE_AHI_CENTRAL}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-aperiodic, mss-acc-manager-initiated, mss-acc-agent-initiated, mss-cat-calculation.	M
Unit-Code	MDC_DIM_EVT_PER_HR	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

6.7.4 Therapy Pressure

Change Table 9 and the text immediately following it as shown:

Table 9—Therapy pressure numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_PRESS_INSTANT} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_MAX} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_MEAN} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_P50} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_P90} or {MDC_PART_PHD_DM, MDC_SABTE_PRESS_P95}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-aperiodic, mss-msmt-aperiodic, mss-acc-manager-initiated, mss-acc-agent-initiated.	M
Unit-Code	MDC_DIM_HECTO_PASCAL	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

NOTE—See IEEE Std 11073-20601a-2010 for information on whether an attribute is static or dynamic.

The Type attribute is used to distinguish the modality of particular therapy pressure between instantaneous value (i.e., MDC_SABTE_PRESS_INSTANT), ~~minimum of a usage session (i.e., MDC_SABTE_PRESS_MIN),~~ maximum of a usage session (i.e., MDC_SABTE_PRESS_MAX), ~~minimum of a usage session (i.e., MDC_SABTE_PRESS_MIN),~~ arithmetic mean of a usage session (i.e., MDC_SABTE_PRESS_MEAN), 50th percentile of a usage session (i.e., MDC_SABTE_PRESS_P50), 90th percentile of a usage session (i.e., MDC_SABTE_PRESS_P90), or 95th percentile of a usage session (i.e., MDC_SABTE_PRESS_P95).

6.7.5 Leakage

Change Table 10 and the text immediately following it as shown:

Table 10—Leakage numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_INSTANT} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_MAX} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_MEAN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_P50} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_P90} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_LEAK_P95}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-aperiodic, mss-msmt-aperiodic, mss-acc-manager-initiated, mss-acc-agent-initiated.	M
Unit-Code	MDC_DIM_L_PER_MIN	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

NOTE—See IEEE Std 11073-20601a-2010 for information on whether an attribute is static or dynamic.

The Type attribute is used to distinguish the modality of particular leakage between instantaneous value (i.e., MDC_SABTE_VOL_LEAK_INSTANT), ~~minimum of a usage session (i.e., MDC_SABTE_VOL_LEAK_MIN)~~, maximum of a usage session (i.e., MDC_SABTE_VOL_LEAK_MAX), ~~minimum of a usage session (i.e., MDC_SABTE_VOL_LEAK_MIN)~~, arithmetic mean of a usage session (i.e., MDC_SABTE_VOL_LEAK_MEAN), 50th percentile of a usage session (i.e., MDC_SABTE_VOL_LEAK_P50), 90th percentile of a usage session (i.e., MDC_SABTE_VOL_LEAK_P90), or 95th percentile of a usage session (i.e., MDC_SABTE_VOL_LEAK_P95).

6.7.6 Respiratory rate

Change Table 11 and the text immediately following it as shown:

Table 11—Respiratory rate numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_INSTANT} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_MAX} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_MEAN} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_P50} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_P90} or {MDC_PART_PHD_DM, MDC_SABTE_RESP_RATE_P95}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-aperiodic, mss-msmt-aperiodic, mss-acc-manager-initiated, mss-acc-agent-initiated.	M
Unit-Code	MDC_DIM_RESP_PER_MIN	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

NOTE—See IEEE Std 11073-20601a-2010 for information on whether an attribute is static or dynamic.

The Type attribute is used to distinguish the modality of particular respiratory rate between instantaneous value (i.e., MDC_SABTE_RESP_RATE_INSTANT), ~~minimum of a usage session (i.e., MDC_SABTE_RESP_RATE_MIN)~~, maximum of a usage session (i.e., MDC_SABTE_RESP_RATE_MAX), ~~minimum of a usage session (i.e., MDC_SABTE_RESP_RATE_MIN)~~, arithmetic mean of a usage session (i.e., MDC_SABTE_RESP_RATE_MEAN), 50th percentile of a usage session (i.e., MDC_SABTE_RESP_RATE_P50), 90th percentile of a usage session (i.e., MDC_SABTE_RESP_RATE_P90), or 95th percentile of a usage session (i.e., MDC_SABTE_RESP_RATE_P95).

6.7.7 Tidal volume

Change the Table 12 and the text immediately following it as shown:

Table 12—Tidal volume numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_INSTANT} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_MAX} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_MEAN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_P50} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_P90} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_TIDAL_P95}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-a-periodic, mss-msmt-a-periodic, mss-acc-manager-initiated, mss-acc-agent-initiated.	M
Unit-Code	MDC_DIM_MILLI_L	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

NOTE—See IEEE Std 11073-20601a-2010 for information on whether an attribute is static or dynamic.

The Type attribute is used to distinguish the modality of particular tidal volume between instantaneous value (i.e., MDC_SABTE_VOL_TIDAL_INSTANT), ~~minimum of a usage session (i.e., MDC_SABTE_VOL_TIDAL_MIN)~~, maximum of a usage session (i.e., MDC_SABTE_VOL_TIDAL_MAX), ~~minimum of a usage session (i.e., MDC_SABTE_VOL_TIDAL_MIN)~~, arithmetic mean of a usage session (i.e., MDC_SABTE_VOL_TIDAL_MEAN), 50th percentile of a usage session (i.e., MDC_SABTE_VOL_TIDAL_P50), 90th percentile of a usage session (i.e., MDC_SABTE_VOL_TIDAL_P90), or 95th percentile of a usage session (i.e., MDC_SABTE_VOL_TIDAL_P95).

6.7.8 Respiratory minute volume

Change Table 13 and the text immediately following it as shown:

Table 13—Respiratory minute volume numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_INSTANT} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_MAX} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_MEAN} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_P50} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_P90} or {MDC_PART_PHD_DM, MDC_SABTE_VOL_MINUTE_P95}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-aperiodic, mss-msmt-aperiodic, mss-acc-manager-initiated, mss-acc-agent-initiated.	M
Unit-Code	MDC_DIM_L_PER_MIN	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

NOTE—See IEEE Std 11073-20601a-2010 for information on whether an attribute is static or dynamic.

The Type attribute is used to distinguish the modality of particular respiratory minute volume between instantaneous value (i.e., MDC_SABTE_VOL_MINUTE_INSTANT), ~~minimum of a usage session (i.e., MDC_SABTE_VOL_MINUTE_MIN)~~, maximum of a usage session (i.e., MDC_SABTE_VOL_MINUTE_MAX), ~~minimum of a usage session (i.e., MDC_SABTE_VOL_MINUTE_MIN)~~, arithmetic mean of a usage session (i.e., MDC_SABTE_VOL_MINUTE_MEAN), 50th percentile of a usage session (i.e., MDC_SABTE_VOL_MINUTE_P50), 90th percentile of a usage session (i.e., MDC_SABTE_VOL_MINUTE_P90), or 95th percentile of a usage session (i.e., MDC_SABTE_VOL_MINUTE_P95).

6.7.9 I:E ratio

Change Table 14 and the text immediately following it as shown:

Table 14—I:E ratio duration numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601a-2010.	M
Type	{MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_INSTANT} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_MAX} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_MIN} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_MEAN} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_P50} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_P90} or {MDC_PART_PHD_DM, MDC_SABTE_RATIO_IE_P95}	M
Metric-Spec-Small	mss-avail-intermittent, mss-avail-stored-data, mss-upd-aperiodic, mss-acc-manager-initiated, mss-acc-agent-initiated, mss-cat-calculation.	M
Unit-Code	MDC_DIM_PERCENT	M
Attribute-Value-Map	See IEEE Std 11073-20601a-2010.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601a-2010.	R

NOTE—See IEEE Std 11073-20601a-2010 for information on whether an attribute is static or dynamic.

The Type attribute is used to distinguish the modality of particular I:E ratio between instantaneous value (i.e., MDC_SABTE_RATIO_IE_INSTANT), ~~minimum of a usage session (i.e., MDC_SABTE_RATIO_IE_MIN)~~, maximum of a usage session (i.e., MDC_SABTE_RATIO_IE_MAX), ~~minimum of a usage session (i.e., MDC_SABTE_RATIO_IE_MIN)~~, arithmetic mean of a usage session (i.e., MDC_SABTE_RATIO_IE_MEAN), 50th percentile of a usage session (i.e., MDC_SABTE_RATIO_IE_P50), 90th percentile of a usage session (i.e., MDC_SABTE_RATIO_IE_P90), or 95th percentile of a usage session (i.e., MDC_SABTE_RATIO_IE_P95).

6.8 Real-time sample array objects

6.8.1 General

Change the first sentence in 6.8.1 as shown:

The SABTE DIM for metric objects (see Figure 5) contains two three RT-SA objects for therapy pressure, leakage, and airflow waveform data. — -

Annex C

(normative)

Allocation of identifiers

C.1 Definitions of terms and codes

Change the code value of MDC_DEV_SPEC_PROFILE_SABTE as shown:

```
#define MDC_DEV_SPEC_PROFILE_SABTE          41240    /* */
```

Change the Reference ID of MDC_SABTE_AHI_CENT as shown:

```
#define MDC_SABTE_AHI_CENTRAL                22196    /* */
```

Change the definitions of multiple nomenclature codes, starting from MDC_SABTE_PRESS through MDC_SABTE_VOL_TIDAL_INSTANT, as shown:

```
#define MDC_SABTE_PRESS                      22336    /* */
#define MDC_SABTE_PRESS_INSTANT              22336    /* */
#define MDC_SABTE_PRESS_MAX                  22337    /* */
#define MDC_SABTE_PRESS_MIN                  22338    /* */
#define MDC_SABTE_PRESS_MEAN                 22339    /* */
#define MDC_SABTE_PRESS_P50                  22343    /* */
#define MDC_SABTE_PRESS_P90                  22345    /* */
#define MDC_SABTE_PRESS_P95                  22346    /* */
#define MDC_SABTE_PRESS_TARGET               22352    /* */
#define MDC_SABTE_PRESS_CPAP_SET             22356    /* */
#define MDC_SABTE_PRESS_CPAP_AUTO_MAX_SET   22360    /* */
#define MDC_SABTE_PRESS_CPAP_AUTO_MIN_SET   22364    /* */
#define MDC_SABTE_PRESS_IPAP_SET             22368    /* */
#define MDC_SABTE_PRESS_EPAP_SET             22372    /* */
#define MDC_SABTE_PRESS_RAMP_START_SET       22376    /* */
#define MDC_SABTE_RESP_RATE_INSTANT          22384    /* */
#define MDC_SABTE_RESP_RATE_MAX              22385    /* */
#define MDC_SABTE_RESP_RATE_MIN              22386    /* */
#define MDC_SABTE_RESP_RATE_MEAN             22387    /* */
#define MDC_SABTE_RESP_RATE_P50              22391    /* */
#define MDC_SABTE_RESP_RATE_P90              22393    /* */
#define MDC_SABTE_RESP_RATE_P95              22394    /* */
#define MDC_SABTE_RESP_RATE_SET              22480    /* */
#define MDC_SABTE_RATIO_IE_INSTANT           22400    /* */
#define MDC_SABTE_RATIO_IE_MAX               22401    /* */
#define MDC_SABTE_RATIO_IE_MIN               22402    /* */
#define MDC_SABTE_RATIO_IE_MEAN              22403    /* */
```


#define MDC_SABTE_RATIO_IE_P50	22407	/*	*/
#define MDC_SABTE_RATIO_IE_P90	22409	/*	*/
#define MDC_SABTE_RATIO_IE_P95	22410	/*	*/
#define MDC_SABTE_RATIO_IE_SET	22484	/*	*/
#define MDC_SABTE_VOL_LEAK	22432	/*	*/
#define MDC_SABTE_VOL_LEAK_INSTANT	22432	/*	*/
#define MDC_SABTE_VOL_LEAK_MAX	22433	/*	*/
#define MDC_SABTE_VOL_LEAK_MIN	22434	/*	*/
#define MDC_SABTE_VOL_LEAK_MEAN	22435	/*	*/
#define MDC_SABTE_VOL_LEAK_P50	22439	/*	*/
#define MDC_SABTE_VOL_LEAK_P90	22441	/*	*/
#define MDC_SABTE_VOL_LEAK_P95	22442	/*	*/
#define MDC_SABTE_VOL_MINUTE_INSTANT	22448	/*	*/
#define MDC_SABTE_VOL_MINUTE_MAX	22449	/*	*/
#define MDC_SABTE_VOL_MINUTE_MIN	22450	/*	*/
#define MDC_SABTE_VOL_MINUTE_MEAN	22451	/*	*/
#define MDC_SABTE_VOL_MINUTE_P50	22455	/*	*/
#define MDC_SABTE_VOL_MINUTE_P90	22457	/*	*/
#define MDC_SABTE_VOL_MINUTE_P95	22458	/*	*/
#define MDC_SABTE_VOL_TIDAL_INSTANT	22464	/*	*/
#define MDC_SABTE_VOL_TIDAL_MAX	22465	/*	*/
#define MDC_SABTE_VOL_TIDAL_MIN	22466	/*	*/
#define MDC_SABTE_VOL_TIDAL_MEAN	22467	/*	*/
#define MDC_SABTE_VOL_TIDAL_P50	22471	/*	*/
#define MDC_SABTE_VOL_TIDAL_P90	22473	/*	*/
#define MDC_SABTE_VOL_TIDAL_P95	22474	/*	*/

Change the format of units to superscript, for the nomenclature codes MDC_DIM_L_PER_MIN and MDC_DIM_EVT_PER_HR, as shown:

```

/*****
* From Dimensions (MDC_PART_DIM) (4)
*****/
#define MDC_DIM_L_PER_MIN          3072    /* 1 min-1 */
#define MDC_DIM_EVT_PER_HR        4732    /* event h-1 */

```

C.2 Systematic derivations of terms and codes

Change the code value of MDC_DEV_SPEC_PROFILE_SABTE in Table C.1 as shown:

Table C.1—Infrastructure nomenclature and codes (MDC_PART_INFRA)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Profile Device SABTE	Sleep apnoea breathing therapy equipment	SABTE	Profile of SABTE device specialization.	MDC_DEV_SPEC_PROFILE_SABTE	41240

Change the Reference ID of MDC_SABTE_AHI_CENT in Table C.2 as shown:

Table C.2—Personal Health Device Disease Management nomenclature and codes (MDC_PART_PHD_DM)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Index Ratio (Number, Duration) Apnoea, Hypopnoea, Central SABTE	Central apnoea hypopnoea index	cAHI	Total number of all central apnoea and central hypopnoea events occurring during a usage session divided by the hours of sleep. See 5.3.1 and 6.7.3.	MDC_SABTE_AHI_CENTRAL	22196

Replace the definitions of multiple nomenclature codes in Table C.2, starting from MDC_SABTE_PRESS through the end of the table, as shown:

Table C.2—Personal Health Device Disease Management nomenclature and codes (MDC_PART_PHD_DM)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Pressure Gas SABTE	Therapy pressure waveform		Sequence of therapy pressure samples. See 5.3.3 and 6.8.2.	MDC_SABTE_PRESS	22336
Pressure Instantaneous Gas SABTE	Instantaneous therapy pressure	P	Instantaneous value of delivered therapy pressure. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_INSTANT	22336
Pressure Maximum Gas SABTE	Maximum therapy pressure	P max	Maximum delivered therapy pressure during a usage session. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_MAX	22337
Pressure Minimum Gas SABTE	Minimum therapy pressure	P min	Minimum delivered therapy pressure during a usage session. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_MIN	22338
Pressure Mean Gas SABTE	Mean therapy pressure	P mean	Mean delivered therapy pressure during a usage session. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_MEAN	22339

Table C.2—Personal Health Device Disease Management nomenclature and codes (MDC_PART_PHD_DM) (continued)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Pressure P50 Gas SABTE	50 th percentile of therapy pressure	P50	50 th percentile of delivered therapy pressure during a usage session. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_P50	22343
Pressure P90 Gas SABTE	90 th percentile of therapy pressure	P90	90 th percentile of delivered therapy pressure during a usage session. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_P90	22345
Pressure P95 Gas SABTE	95 th percentile of therapy pressure	P95	95 th percentile of delivered therapy pressure during a usage session. See 5.3.2 and 6.7.5.	MDC_SABTE_PRESS_P95	22346
Pressure Target Gas SABTE	Target therapy pressure waveform		Sequence of target therapy pressure samples. See 5.3.3 and 6.8.2.	MDC_SABTE_PRESS_TARGET	22352
Pressure CPAP, Setting SABTE	CPAP pressure set	P CPAP set	Setting of target therapy pressure in CPAP mode during a therapy session. See 5.6.7.1 and 6.7.18.	MDC_SABTE_PRESS_CPAP_SET	22356
Pressure CPAP, Auto, Maximum, Setting SABTE	Auto-CPAP maximum pressure set	Pmax APAP set	Setting of maximum target therapy pressure in Auto-CPAP mode during a therapy session. See 5.6.8.2 and 6.7.20.	MDC_SABTE_PRESS_CPAP_AUTO_MAX_SET	22360
Pressure CPAP, Auto, Minimum, Setting SABTE	Auto-CPAP minimum pressure set	Pmin APAP set	Setting of minimum target therapy pressure in Auto-CPAP mode during a therapy session. See 5.6.8.1 and 6.7.19.	MDC_SABTE_PRESS_CPAP_AUTO_MIN_SET	22364
Pressure IPAP, Setting SABTE	IPAP pressure set	P IPAP set	Setting of target inspiration therapy pressure in BiLevel PAP mode during a breath cycle. See 5.6.9.1 and 6.7.21.	MDC_SABTE_PRESS_IPAP_SET	22368
Pressure EPAP, Setting SABTE	EPAP pressure set	P EPAP set	Setting of target expiration therapy pressure in BiLevel PAP mode during a breath cycle. See 5.6.9.2 and 6.7.22.	MDC_SABTE_PRESS_EPAP_SET	22372
Pressure Start, Setting Ramp SABTE	Ramp start pressure set		Setting of length of the sleep ramp. See 5.6.4.1 and 6.7.15.	MDC_SABTE_PRESS_RAMP_START_SET	22376
Rate Instantaneous Breath SABTE	Instantaneous respiration rate	RR	Instantaneous value of respiration rate. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_INSTANT	22384
Rate Maximum Breath SABTE	Maximum respiration rate	RR max	Maximum respiration rate during a usage session. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_MAX	22385
Rate Minimum Breath SABTE	Minimum respiration rate	RR min	Minimum respiration rate during a usage session. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_MIN	22386
Rate Mean Breath SABTE	Mean respiration rate	RR mean	Mean respiration rate during a usage session. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_MEAN	22387

Table C.2—Personal Health Device Disease Management nomenclature and codes (MDC_PART_PHD_DM) (continued)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Rate P50 Breath SABTE	50 th percentile of respiration rate		50 th percentile of respiration rate during a usage session. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_P50	22391
Rate P90 Breath SABTE	90 th percentile of respiration rate		90 th percentile of respiration rate during a usage session. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_P90	22393
Rate P95 Breath SABTE	95 th percentile of respiration rate		95 th percentile of respiration rate during a usage session. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_P95	22394
Rate Setting Breath SABTE	Respiratory rate set	RR set	Instantaneous value of respiration rate. See 5.3.7 and 6.7.7.	MDC_SABTE_RESP_RATE_SET	22480
Ratio Duration(Inspiration Phase), Duration(Expiration Phase), Instantaneous Gas SABTE	Instantaneous I:E ratio	TI/TE	Setting of target breathing frequency in BiLevel PAP mode during a therapy session. See 5.6.9.3 and 6.7.23.	MDC_SABTE_RATIO_IE_INSTANT	22400
Ratio Duration(Inspiration Phase), Duration(Expiration Phase), Maximum Gas SABTE	Maximum I:E ratio	TI/TE max	Instantaneous value of I:E ratio. See 5.3.10 and 6.7.10.	MDC_SABTE_RATIO_IE_MAX	22401
Ratio Duration(Inspiration Phase), Duration(Expiration Phase), Minimum Gas SABTE	Minimum I:E ratio	TI/TE min	Minimum I:E ratio during a usage session. See 5.3.10 and 6.7.10.	MDC_SABTE_RATIO_IE_MIN	22402
Ratio Duration(Inspiration Phase), Duration(Expiration Phase), Mean Gas SABTE	Mean I:E ratio	TI/TE mean	Maximum I:E ratio during a usage session. See 5.3.10 and 6.7.10.	MDC_SABTE_RATIO_IE_MEAN	22403
Ratio Duration(Inspiration Phase), Duration(Expiration Phase), P50 Gas SABTE	50 th percentile of I:E ratio		Minimum I:E ratio during a usage session. See 5.3.10 and 6.7.10.	MDC_SABTE_RATIO_IE_P50	22407
Ratio Duration(Inspiration Phase), Duration(Expiration Phase), P90 Gas SABTE	90 th percentile of I:E ratio		Mean I:E ratio during a usage session. See 5.3.10 and 6.7.10.	MDC_SABTE_RATIO_IE_P90	22409

Table C.2—Personal Health Device Disease Management nomenclature and codes (MDC_PART_PHD_DM) (continued)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Ratio Duration(InspirationPhase), Duration(ExpirationPhase), P95 Gas SABTE	95 th percentile of I:E ratio		95 th percentile of I:E ratio during a usage session. See 5.3.10 and 6.7.10.	MDC_SABTE_RATIO_IE_P95	22410
Ratio Duration(InspirationPhase), Duration(ExpirationPhase), Setting Gas SABTE	I:E ratio set	TI/TE set	Setting of target ratio between duration of the inspiration to the duration of the expiration in BiLevel PAP mode during a breath cycle. See 5.6.9.4 and 6.7.24.	MDC_SABTE_RATIO_IE_SET	22484
Volume Leakage SABTE	Leakage waveform		Sequence of leakage samples. See 5.3.5 and 6.8.3.	MDC_SABTE_VOL_LEAK	22432
Volume Instantaneous Leakage SABTE	Instantaneous leakage		Instantaneous value of leakage. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_INSTANT	22432
Volume Maximum Leakage SABTE	Maximum leakage		Maximum leakage during a usage session. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_MAX	22433
Volume Minimum Leakage SABTE	Minimum leakage		Minimum leakage during a usage session. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_MIN	22434
Volume Mean Leakage SABTE	Mean leakage		Mean leakage during a usage session. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_MEAN	22435
Volume P50 Leakage SABTE	50 th percentile of leakage		50 th percentile of leakage during a usage session. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_P50	22439
Volume P90 Leakage SABTE	90 th percentile of leakage		90 th percentile of leakage during a usage session. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_P90	22441
Volume P95 Leakage SABTE	95 th percentile of leakage		95 th percentile of leakage during a usage session. See 5.3.4 and 6.7.6.	MDC_SABTE_VOL_LEAK_P95	22442
Volume OneMinute, Instantaneous Gas SABTE	Instantaneous respir. minute volume	RMV	Instantaneous value of respiratory minute volume. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_INSTANT	22448
Volume OneMinute, Maximum Gas SABTE	Maximum respir. minute volume	RMV max	Maximum respiratory minute volume during a usage session. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_MAX	22449
Volume OneMinute, Minimum Gas SABTE	Minimum respir. minute volume	RMV min	Minimum respiratory minute volume during a usage session. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_MIN	22450
Volume OneMinute, Mean Gas SABTE	Mean respir. minute volume	RMV mean	Mean respiratory minute volume during a usage session. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_MEAN	22451

Table C.2—Personal Health Device Disease Management nomenclature and codes (MDC_PART_PHD_DM) (continued)

Systematic name	Common term	Acronym	Description/definition	Reference ID	Code
Volume OneMinute, P50 Gas SABTE	50 th percentile of respir. minute volume		50 th percentile of respiratory minute volume during a usage session. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_P50	22455
Volume OneMinute, P90 Gas SABTE	90 th percentile of respir. minute volume		90 th percentile of respiratory minute volume during a usage session. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_P90	22457
Volume OneMinute, P95 Gas SABTE	95 th percentile of respir. minute volume		95 th percentile of respiratory minute volume during a usage session. See 5.3.9 and 6.7.9.	MDC_SABTE_VOL_MINUTE_P95	22458
Volume Instantaneous Lung, Tidal SABTE	Instantaneous respir. tidal volume	VT	Instantaneous value of respiratory tidal volume. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_INSTANT	22464
Volume Maximum Lung, Tidal SABTE	Maximum respir. tidal volume	VT max	Maximum respiratory tidal volume during a usage session. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_MAX	22465
Volume Minimum Lung, Tidal SABTE	Minimum respir. tidal volume	VT min	Minimum respiratory tidal volume during a usage session. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_MIN	22466
Volume Mean Lung, Tidal SABTE	Mean respir. tidal volume	VT mean	Mean respiratory tidal volume during a usage session. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_MEAN	22467
Volume P50 Lung, Tidal SABTE	50 th percentile of respir. tidal volume		50 th percentile of respiratory tidal volume during a usage session. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_P50	22471
Volume P90 Lung, Tidal SABTE	90 th percentile of respir. tidal volume		90 th percentile of respiratory tidal volume during a usage session. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_P90	22473
Volume P95 Lung, Tidal SABTE	95 th percentile of respir. tidal volume		95 th percentile of respiratory tidal volume during a usage session. See 5.3.8 and 6.7.8.	MDC_SABTE_VOL_TIDAL_P95	22474

Annex E

(informative)

Protocol data unit examples

E.4 GET MDS attributes service

Change the subclause number from E.4.1.1 to E.4.1 as shown:

~~E.4.1.1~~ E.4.1 General

Change the subclause number from E.4.1.2 to E.4.2 as shown:

E.4.1.2 E.4.2 Get all medical device system attributes request

Change the subclause number from E.4.1.3 to E.4.3 as shown:

E.4.1.3 E.4.3 Get response with all MDS attributes

Change the line starting with “0x10 0x19” as shown:

```
0x10 0x198 type = MDC_DEV_SPEC_PROFILE_SABTE
```

