

INTERNATIONAL  
STANDARD

**ISO**  
**9022-19**

First edition  
1994-07-15

---

---

**Optics and optical instruments —  
Environmental test methods —**

**Part 19:**

Temperature cycles combined with sinusoidal  
or random vibration

*Optique et instruments d'optique — Méthodes d'essais  
d'environnement —*

*Partie 19: Essai combiné cycles de températures-vibrations sinusoïdales  
ou aléatoires*



Reference number  
ISO 9022-19:1994(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 9022-19 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 1, *Fundamental standards*.

ISO 9022 consists of the following parts, under the general title *Optics and optical instruments — Environmental test methods*:

- Part 1: *Definitions, extent of testing*
- Part 2: *Cold, heat, humidity*
- Part 3: *Mechanical stress*
- Part 4: *Salt mist*
- Part 5: *Combined cold, low air pressure*
- Part 6: *Dust*
- Part 7: *Drip, rain*
- Part 8: *High pressure, low pressure, immersion*
- Part 9: *Solar radiation*
- Part 10: *Combined sinusoidal vibration, dry heat or cold*

© ISO 1994

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

- Part 11: Mould growth
- Part 12: Contamination
- Part 13: Combined shock, bump or free fall, dry heat or cold
- Part 14: Dew, hoarfrost, ice
- Part 15: Combined random vibration wide band: reproducibility medium, in dry heat or cold
- Part 16: Combined bounce or steady-state acceleration, in dry heat or cold
- Part 17: Combined contamination, solar radiation
- Part 18: Combined damp heat and low internal pressure
- Part 19: Temperature cycles combined with sinusoidal or random vibration
- Part 20: Humid atmosphere containing sulfur dioxide or hydrogen sulfide

Annex A of this part of ISO 9022 is for information only.

This document is a preview generated by EVS

## Introduction

Optical instruments are affected during their use by a number of different environmental parameters which they are required to resist without significant reduction in performance.

The type and severity of these parameters depend on the conditions of use of the instrument (for example, in the laboratory or workshop) and on its geographical location. The environmental effects on optical instrument performance in the tropics and subtropics are totally different from those found when they are used in the arctic regions. Individual parameters cause a variety of different and overlapping effects on instrument performance.

The manufacturer attempts to ensure, and the user naturally expects, that instruments will resist the likely rigours of their environment throughout their life. This expectation can be assessed by exposure of the instrument to a range of simulated environmental parameters under controlled laboratory conditions. The severity of these conditions is often increased to obtain meaningful results in a relatively short period of time.

In order to allow assessment and comparison of the response of optical instruments to appropriate environmental conditions, ISO 9022 contains details of a number of laboratory tests which reliably simulate a variety of different environments. The tests are based largely on IEC standards, modified where necessary to take into account features special to optical instruments.

It should be noted that, as a result of continuous progress in all fields, optical instruments are no longer only precision-engineered optical products, but, depending on their range of application, also contain additional assemblies from other fields. For this reason, the principal function of the instrument must be assessed to determine which International Standard should be used for testing. If the optical function is of primary importance, then ISO 9022 is applicable, but if other functions take precedence then the appropriate International Standard in the field concerned should be applied. Cases may arise where application of both ISO 9022 and other appropriate International Standards will be necessary.

# Optics and optical instruments — Environmental test methods

## Part 19:

### Temperature cycles combined with sinusoidal or random vibration

#### 1 Scope

This part of ISO 9022 specifies the method of testing of optical instruments and instruments containing optical components under equivalent conditions, to assess their resistance to temperature cycles combined with sinusoidal or random vibration.

The purpose of testing is to investigate on a number of specimens to be arranged to what extent the optical, thermal, mechanical, chemical or electrical performance characteristics of the specimens are affected by the combined effect of temperature cycles and mechanical vibrations using specific variations of the operating condition.

#### 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 9022. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 9022 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 9022-1:1994<sup>1)</sup>, *Optics and optical instruments — Environmental test methods — Part 1: Definitions, extent of testing.*

1) To be published.

IEC 68-2-6:1982, *Environmental testing — Part 2: Tests — Test Fc and guidance: Vibration (sinusoidal).*

IEC 68-2-36:1973, *Environmental testing — Part 2: Tests — Test Fdb: Random vibration wide band — Reproducibility: medium.*

IEC 68-2-37:1973, *Environmental testing — Part 2: Tests — Test Fde: Random vibration wide band — Reproducibility: low.*

IEC 68-2-47:1982, *Environmental testing — Part 2: Tests — Mounting of components, equipment and other articles for tests including shock (Ea), bump (Eb), vibration (Fe and Fd) and steady-state acceleration (Ga) and guidance.*

#### 3 General information and test conditions

The size of the test chamber and the arrangement of the specimens shall be selected in such a way that a uniform temperature of all specimens contained therein is guaranteed. If condensation occurs, the specimens shall be protected from falling drops.

The test fixture for the specimens shall conform to the requirements of IEC 68-2-47.

For the purposes of this International Standard, the acceleration of free fall shall be taken as  $g = 9,81 \text{ m}\cdot\text{s}^{-2}$ .