

Dentistry - Polymer-based restorative materials (ISO 4049:2019)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 4049:2019 sisaldab Euroopa standardi EN ISO 4049:2019 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 4049:2019 consists of the English text of the European standard EN ISO 4049:2019.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 19.06.2019.	Date of Availability of the European standard is 19.06.2019.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.060.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Dentistry - Polymer-based restorative materials (ISO
4049:2019)

Médecine bucco-dentaire - Produits de restauration à
base de polymères (ISO 4049:2019)

Zahnheilkunde - Polymerbasierende
Restaurationswerkstoffe (ISO 4049:2019)

This European Standard was approved by CEN on 18 May 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 4049:2019) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2019, and conflicting national standards shall be withdrawn at the latest by December 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 4049:2009.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 4049:2019 has been approved by CEN as EN ISO 4049:2019 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification	2
4.1 Type	2
4.2 Class	2
5 Requirements	2
5.1 Biocompatibility	2
5.2 Physical and chemical properties	2
5.2.1 General	2
5.2.2 Film thickness, luting materials	3
5.2.3 Working time, Class 1 and Class 3 restorative materials, excluding luting materials	3
5.2.4 Working time, Class 1 and Class 3 luting materials	3
5.2.5 Setting time, Class 1 materials	3
5.2.6 Setting time, Class 3 materials	3
5.2.7 Sensitivity to light, Class 2 materials	3
5.2.8 Depth of cure, Class 2 materials excluding luting materials	3
5.2.9 Flexural strength	4
5.2.10 Water sorption and solubility	4
5.3 Shade of restorative materials	4
5.4 Colour stability after irradiation and water sorption	4
5.5 Radio-opacity	4
6 Sampling	5
7 Test methods	5
7.1 General reagent — Water	5
7.2 Test conditions	5
7.3 Inspection	6
7.4 Preparation of test specimens	6
7.5 Measurement of film thickness of luting materials	6
7.5.1 Apparatus	6
7.5.2 Test procedure	8
7.5.3 Treatment of results	9
7.6 Working time, Class 1 and Class 3 restorative materials, excluding luting materials	10
7.6.1 Apparatus	10
7.6.2 Procedure	11
7.6.3 Treatment of results	11
7.7 Working time, Class 1 and Class 3 luting materials	12
7.7.1 Apparatus	12
7.7.2 Procedure	12
7.7.3 Treatment of results	12
7.8 Setting time, Class 1 and Class 3 materials	12
7.8.1 Apparatus for the determination of setting time of Class 1 and Class 3 restorative materials	12
7.8.2 Apparatus for the determination of setting time of Class 1 and Class 3 luting materials	12
7.8.3 Procedure	12
7.8.4 Treatment of results	13
7.9 Sensitivity to light, Class 2 materials	13
7.9.1 Apparatus	14

7.9.2	Procedure.....	14
7.9.3	Treatment of results.....	15
7.10	Depth of cure, Class 2 materials excluding luting materials.....	15
7.10.1	Apparatus.....	15
7.10.2	Procedure.....	15
7.10.3	Treatment of results.....	16
7.11	Flexural strength.....	16
7.11.1	Apparatus.....	16
7.11.2	Preparation of test specimens.....	17
7.11.3	Procedure.....	18
7.11.4	Treatment of results.....	18
7.12	Water sorption and solubility.....	19
7.12.1	Apparatus.....	19
7.12.2	Preparation of test specimens.....	20
7.12.3	Procedure.....	21
7.12.4	Treatment of results.....	21
7.13	Shade and colour stability after irradiation and water sorption.....	22
7.13.1	General.....	22
7.13.2	Apparatus.....	22
7.13.3	Preparation of test specimens.....	23
7.13.4	Procedure.....	23
7.13.5	Colour comparison for shade.....	23
7.13.6	Colour comparison for colour stability.....	23
7.14	Radio-opacity.....	23
7.14.1	Perform the test in accordance with ISO 13116.....	23
7.14.2	Preparation of test specimens.....	23
7.14.3	Treatment of results.....	23
8	Packaging, marking, instructions and information to be supplied by the manufacturer.....	24
8.1	Packaging.....	24
8.2	Marking and instructions for use.....	24
8.3	Declaration of components.....	28
	Bibliography.....	29

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 1, *Filling and restorative materials*.

This fifth edition cancels and replaces the fourth edition (ISO 4049:2009), which has been technically revised. The main changes compared to the previous edition are as follows:

- the test for sensitivity to ambient light has been changed because a filter used in the current test was not available;
- the test for radio-opacity has been updated to refer to ISO 13116;
- luting materials no longer have to conform to the requirement for depth of cure;
- the manufacturer is now required to publish details of material composition, see [Clause 8](#);
- several minor changes have been made to clarify content together with editorial changes.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Specific qualitative and quantitative requirements for freedom from biological hazard are not included in this document. It is recommended, however, that reference should be made to ISO 10993-1 and ISO 7405 when assessing possible biological or toxicological hazards.

This document is a preview generated by EVS

Dentistry — Polymer-based restorative materials

1 Scope

This document specifies requirements for dental polymer-based restorative materials supplied in a form suitable for mechanical mixing, hand-mixing, or intra-oral and extra-oral external energy activation, and intended for use primarily for the direct or indirect restoration of the teeth and for luting.

The polymer-based luting materials covered by this document are intended for use in the cementation or fixation of restorations and appliances such as inlays, onlays, veneers, crowns and bridges. This document does not cover those polymer-based luting materials that have an adhesive component within the structure of the material (see ISO/TS 16506).

The document does not cover polymer-based materials intended to prevent caries (see ISO 6874), core materials or those used for veneering metal sub-frames (see ISO 10477).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 7491:2000, *Dental materials — Determination of colour stability*

ISO 8601-1:2019, *Date and time — Representations for information interchange Part — 1: Basic rules*

ISO 8601-2:2019, *Date and time — Representations for information interchange Part — 2: Extensions*

ISO 13116:2014, *Dentistry — Test method for determining radio-opacity of materials*

ISO 17304:2013, *Dentistry — Polymerization shrinkage: Method for determination of polymerization shrinkage of polymer-based materials*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

opaque

shade of an intensely pigmented polymer-based restorative material of low translucency

3.2

outer pack

form of packaging used to combine a number of single dose containers or capsules