

Elektromagnetiline ühilduvus. Osa 4-15: Katsetus- ja mõõtetehnika. Värelusmõõtur. Talitluse ja ehituse iseloomustus

Electromagnetic compatibility (EMC) - Part 4-15: Testing and measurement techniques - Flickermeter - Functional and design specifications

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN 61000-4-15:2011 sisaldab Euroopa standardi EN 61000-4-15:2011 ingliskeelset teksti.

Standard on kinnitatud Eesti Standardikeskuse 30.06.2011 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.

Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on 10.06.2011.

Standard on kättesaadav Eesti standardiorganisatsioonist.

This Estonian standard EVS-EN 61000-4-15:2011 consists of the English text of the European standard EN 61000-4-15:2011.

This standard is ratified with the order of Estonian Centre for Standardisation dated 30.06.2011 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.

Date of Availability of the European standard text 10.06.2011.

The standard is available from Estonian standardisation organisation.

ICS 33.100.20

Inglisekeelsed võtmesõnad: design, electromagnetic compatibility, measuring instruments, performance evaluation, specification, tests,

Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: 605 5050; E-mail: info@evs.ee

**Electromagnetic compatibility (EMC) -
Part 4-15: Testing and measurement techniques -
Flickermeter -
Functional and design specifications
(IEC 61000-4-15:2010)**

Compatibilité électromagnétique (CEM) -
Partie 4-15: Techniques d'essai et de
mesure -
Flickermètre -
Spécifications fonctionnelles et de
conception
(CEI 61000-4-15:2010)

Elektromagnetische Verträglichkeit (EMV)
-
Teil 4-15: Prüf- und Messverfahren -
Flickermeter -
Funktionsbeschreibung und
Auslegungsspezifikation
(IEC 61000-4-15:2010)

This European Standard was approved by CENELEC on 2011-01-02. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 77A/722/FDIS, future edition 2 of IEC 61000-4-15, prepared by SC 77A, Low frequency phenomena, of IEC TC 77, Electromagnetic compatibility was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 61000-4-15 on 2011-01-02.

This European Standard supersedes EN 61000-4-15:1998 + A1:2003.

EN 61000-4-15:2011, in particular, adds or clarifies the definition of several directly measured parameters, so that diverging interpretations are avoided.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

- | | | |
|--|-------|------------|
| – latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement | (dop) | 2011-10-02 |
| – latest date by which the national standards conflicting with the EN have to be withdrawn | (dow) | 2014-01-02 |

Annex ZA has been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 61000-4-15:2010 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

IEC 61000-4-30	NOTE Harmonized as EN 61000-4-30.
----------------	-----------------------------------

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60068	Series	Environmental testing	EN 60068	Series
IEC 61000-3-3	-	Electromagnetic compatibility (EMC) – Part 3-3: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection	EN 61000-3-3	-
IEC 61000-3-11	-	Electromagnetic compatibility (EMC) - Part 3-11: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems - Equipment with rated current ≤ 75 A and subject to conditional connection	EN 61000-3-11	-
IEC 61010-1	-	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements	EN 61010-1	-
IEC 61326-1	-	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements	EN 61326-1	-

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope and object.....	7
2 Normative references.....	7
3 Parameters and symbols.....	8
3.1 Directly measured parameters and characteristics	8
3.1.1 General	8
3.1.2 Half period rms value of the voltage	8
3.1.3 Half period rms value characteristics.....	8
3.1.4 Relative half period rms value characteristics.....	8
3.1.5 Steady state voltage and voltage change characteristics	8
3.1.6 Steady state voltage change	9
3.1.7 Maximum voltage change during a voltage change characteristic	9
3.1.8 Maximum steady state voltage change during an observation period	9
3.1.9 Maximum absolute voltage change during an observation period	10
3.1.10 Voltage deviation	10
3.1.11 Centre voltage	10
3.2 Symbols	10
4 Description of the instrument	11
4.1 General	11
4.2 Block 1 – Input voltage adaptor.....	11
4.3 Block 2 – Squaring multiplier.....	11
4.4 Block 3 – Weighting filters	12
4.5 Block 4 – Squaring and smoothing	12
4.6 Block 5 – On-line statistical analysis	12
4.7 Outputs	13
4.7.1 General	13
4.7.2 P_{lin} output.....	13
4.7.3 P_{inst} output	13
4.7.4 P_{st} output	13
4.7.5 P_{lt} output	13
4.7.6 d-meter outputs	13
5 Specification.....	13
5.1 Response and accuracy.....	13
5.2 Input voltage ranges	18
5.3 Voltage adaptor	18
5.4 Weighting filters.....	18
5.5 Weighting filter response in block 3.....	18
5.6 Squaring multiplier and sliding mean filter	19
5.7 General statistical analysis procedure	19
5.7.1 General	19
5.7.2 Short-term flicker evaluation	19
5.7.3 Long-term flicker evaluation	20
6 Flickermeter tests	20
6.1 General	20
6.2 Sinusoidal/rectangular voltage changes	21

6.3	Rectangular voltage changes and performance testing	21
6.4	Combined frequency and voltage changes – Class F1 flickermeters	22
6.5	Distorted voltage with multiple zero crossings – Class F1 flickermeters	23
6.6	Bandwidth test using harmonic and inter-harmonic side band modulation	23
6.7	Phase jumps – Class F1 flickermeters	24
6.8	Rectangular voltage changes with 20 % duty cycle	24
6.9	d parameter test, d_c , $d_{\max}$, and $d(t) > 3,3\%$	25
7	Environmental and other requirements	27
7.1	General	27
7.2	Insulation, climatic, electromagnetic compatibility, and other tests	27
Annex A (normative)	Techniques to improve accuracy of flicker evaluation	30
Annex B (informative)	Meaning of $\Delta U/U$ and number of voltage changes, d_c , $d(t)$, $d_{\max}$ examples	32
Annex C (informative)	Sample protocols for type testing	36
Bibliography		40
Figure 1	– Illustration of 28 Hz modulated test voltage with 20 % duty cycle	25
Figure 2	– Functional diagram of IEC flickermeter	28
Figure 3	– Basic illustration of the time-at-level method for $P_{st} = 2,000$	29
Figure B.1	– Rectangular voltage change $\Delta U/U = 40\%$, 8,8 Hz, 17,6 changes/second	33
Figure B.2	– Illustration of “ d ” parameter definitions	35
Table 1a	– Normalized flickermeter response 120 V / 50 Hz and 120 V / 60 Hz for sinusoidal voltage fluctuations	14
Table 1b	– Normalized flickermeter response 230 V / 50 Hz and 230 V / 60 Hz for sinusoidal voltage fluctuations	15
Table 2a	– Normalized flickermeter response 120 V / 50 Hz and 120 V / 60 Hz for rectangular voltage fluctuations	16
Table 2b	– Normalized flickermeter response 230 V / 50 Hz and 230 V / 60 Hz for rectangular voltage fluctuations	17
Table 3	– Indicative values for the parameters of lamps	19
Table 4	– Test specifications for flickermeter	21
Table 5	– Test specification for flickermeter classifier	22
Table 6	– Test specification for combined frequency and voltage changes – Class F1 flickermeters	23
Table 7	– Test specification for distorted voltage with multiple zero crossings – Class F1 flickermeters	23
Table 8	– 8,8 Hz modulation depth for distorted voltage test – Class F 1 flickermeters	23
Table 9	– Test specification for Harmonics with side band – Class F1 flickermeters	24
Table 10	– Test specification for phase jumps – Class F1 flickermeters	24
Table 11	– Test specification for rectangular voltage changes with duty ratio	24
Table 12	– Test specification for d_c , $d_{\max}$, $t_{(d(t))} > 3,3\%$	25
Table 13	– Test specification for d_c , $d_{\max}$, $t_{(d(t))} > 3,3\%$	26
Table B.1	– Correction factor for other voltage/frequency combinations	33

INTRODUCTION

IEC 61000-4 is a part of the IEC 61000 series, according to the following structure:

- Part 1: General
 - General consideration (introduction, fundamental principles)
 - Definitions, terminology
- Part 2: Environment
 - Description of the environment
 - Classification of the environment
 - Compatibility levels
- Part 3: Limits
 - Emission limits
 - Immunity limits (in so far as they do not fall under the responsibility of the product committees)
- Part 4: Testing and measurement techniques
 - Measurement techniques
 - Testing techniques
- Part 5: Installation and mitigation guidelines
 - Installation guidelines
 - Mitigation methods and devices
- Part 6: Generic standards
- Part 9: Miscellaneous

Each part is further subdivided into several parts, published either as international standards, as technical specifications or technical reports, some of which have already been published as sections. Others are and will be published with the part number followed by a dash and completed by a second number identifying the subdivision (example: IEC 61000-6-1).

ELECTROMAGNETIC COMPATIBILITY (EMC) –

Part 4-15: Testing and measurement techniques – Flickermeter – Functional and design specifications

1 Scope and object

This part of IEC 61000 gives a functional and design specification for flicker measuring apparatus intended to indicate the correct flicker perception level for all practical voltage fluctuation waveforms. Information is presented to enable such an instrument to be constructed. A method is given for the evaluation of flicker severity on the basis of the output of flickermeters complying with this standard.

The flickermeter specifications in this part of IEC 61000 relate only to measurements of 120 V and 230 V, 50 Hz and 60 Hz inputs. Characteristics of some incandescent lamps for other voltages are sufficiently similar to the values in Table 1 and Table 2, that the use of a correction factor can be applied for those other voltages. Some of these correction factors are provided in the Annex B. Detailed specifications for voltages and frequencies other than those given above, remain under consideration.

The object of this part of IEC 61000 is to provide basic information for the design and the instrumentation of an analogue or digital flicker measuring apparatus. It does not give tolerance limit values of flicker severity.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60068 (all parts), *Environmental testing*

IEC 61000-3-3, *Electromagnetic compatibility (EMC) – Part 3-3: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection*

IEC 61000-3-11, *Electromagnetic compatibility (EMC) – Part 3-11: Limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems – Equipment with rated current ≤ 75 A and subject to conditional connection*

IEC 61010-1, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements*

IEC 61326-1, *Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements*