

---

---

**Water quality — Determination of  
ammonium nitrogen by flow analysis  
(CFA and FIA) and spectrometric detection**

*Qualité de l'eau — Détermination de l'azote ammoniacal par analyse en  
flux (CFA et FIA) et détection spectrométrique*



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 11732 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

Annexes A to D of this International Standard are for information only.

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet central@iso.ch  
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

## Introduction

Methods using flow analysis are automatized wet chemical procedures and are therefore particularly suitable for the processing of large sample series at a high analysis frequency (up to 100 samples per hour).

One differentiates between flow injection analysis (FIA) [1], [2] and continuous flow analysis (CFA) [3]. Both methods include the automatic dosage of the sample into a flow system (manifold) in which the analytes in the sample react with the reagent solutions on their way through the manifold. The sample preparation may be integrated in the manifold. The reaction product is analysed spectrometrically in a flow detector.

This document is a preview generated by EVS

This page intentionally left blank

# Water quality — Determination of ammonium nitrogen by flow analysis (CFA and FIA) and spectrometric detection

## 1 Determination of ammonium nitrogen by flow injection analysis (FIA) and spectrometric detection

### 1.1 Scope

#### 1.1.1 Field of application

This International Standard specifies a method suitable for the determination of ammonium nitrogen in various types of waters (such as ground, drinking, surface and waste waters) in mass concentrations ranging from 0,1 to 10 mg/l (in the undiluted sample). In particular cases, the range of application may be adapted by varying the operating conditions.

#### 1.1.2 Interferences

Volatile amines will diffuse through the membrane and lead to a pH shift. If the concentrations of the volatile amines (e.g. methylamine or ethylamine) are equal to those of the ammonium, erroneously high results may be expected [12]. In significant cases, prior to analysis an (online) distillation of the sample, adjusted to a pH of 5,8 may be necessary.

Interferences may occur in exceptional cases when the sample does not reach a pH at least 12 after the addition of the alkaline reagent, since then ammonium will not be converted quantitatively into ammonia. In particular, this may occur with strongly acidic or buffered samples. In such cases the pH of the sample should be adjusted to 3 to 5 by the addition of sodium hydroxide solution (1.4.1 or 1.4.2).

High concentrations of metal ions which may precipitate as hydroxides will give poorly reproducible results. The addition of a suitable complexing agent, such as (ethylenedinitrilo)tetraacetic acid, disodium salt, to the alkaline reaction solution (1.4.17) in a sufficiently large concentration will prevent interference by Cu, Zn, Fe, Ca, Mg and Al; up to individual metal concentrations of 0,2 mg/l, a concentration of 30 g/l of ethylenedinitrilotetraacetic acid, disodium salt, in solution R<sub>1</sub> (see 1.4.17) is adequate.

For samples containing particulate matter, see 1.6 (last paragraph).

Samples with a total salt concentration of > 10 g/l should be diluted prior to measurement.

### 1.2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3696:1987, *Water for analytical and laboratory use — Specification and test methods*.

ISO 5667-1:1980, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes*.

ISO 5667-2:1991, *Water quality — Sampling — Part 2: Guidance on sampling techniques*.

ISO 5667-3:1994, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples*.

### 1.3 Principle

The test sample containing ammonium is injected into a continuous carrier stream by means of an injection valve, and is mixed with a continuous flow of an alkaline solution. The ammonia formed is separated in a diffusion cell from the solution over a hydrophobic semipermeable membrane and taken up by a streaming recipient flow containing a pH indicator. Due to the resulting pH shift, the indicator solution will change colour; the colour change is monitored continuously in a flow spectrophotometer. Additional information concerning this analytical technique is given in [4], [5], [6], [7] and [8].

### 1.4 Reagents

Apart from the reagents listed in 1.4.4 to 1.4.6, use only reagents of analytical grade quality for the determination of nitrogen or, if not available, those of recognized analytical grade quality and water of grade 1 (in accordance with ISO 3696), freshly prepared. The ammonium content of the blank shall be checked regularly (see 1.7.3).

**1.4.1 Sodium hydroxide solution I**,  $c(\text{NaOH}) = 5 \text{ mol/l}$ .

**1.4.2 Sodium hydroxide solution II**,  $c(\text{NaOH}) = 0,01 \text{ mol/l}$ .

**1.4.3 Ethylenedinitrilotetraacetic acid (EDTA)**, disodium salt, monohydrate,  $\text{Na}_2\text{C}_{10}\text{H}_{14}\text{N}_2\text{O}_8 \cdot \text{H}_2\text{O}$ .

**1.4.4 Bromcresol purple**,  $\text{C}_{21}\text{H}_{16}\text{Br}_2\text{O}_5\text{S}$ .

**1.4.5 Bromthymol blue**,  $\text{C}_{27}\text{H}_{28}\text{Br}_2\text{O}_5\text{S}$ .

**1.4.6 Cresol red**,  $\text{C}_{21}\text{H}_{18}\text{O}_5\text{S}$ .

**1.4.7 Ammonium chloride**,  $\text{NH}_4\text{Cl}$ , dried at  $105^\circ\text{C}$  to constant weight.

**1.4.8 Potassium chloride**,  $\text{KCl}$ .

**1.4.9 Boric acid**,  $\text{H}_3\text{BO}_3$ .

**1.4.10 Ethanol**,  $\text{C}_2\text{H}_5\text{OH}$ , 95 % mass fraction.

**1.4.11 Hydrochloric acid I**,  $c(\text{HCl}) = 0,01 \text{ mol/l}$ .

**1.4.12 Hydrochloric acid II**,  $c(\text{HCl}) = 0,1 \text{ mol/l}$ .

**1.4.13 Hydrochloric acid III**,  $c(\text{HCl}) = 1,0 \text{ mol/l}$ .

**1.4.14 Sulfuric acid**,  $\rho(\text{H}_2\text{SO}_4) = 1,84 \text{ g/ml}$ .

**1.4.15 Mixed indicator.**

In a mortar prepare a dry mixture consisting of 10 g of Bromcresol purple (1.4.4), 5 g of Bromthymol blue (1.4.5), 2,5 g of Cresol red (1.4.6) and 45 g of potassium chloride (1.4.8).

The given quantities can be reduced (e.g. by one-tenth), maintaining the ratio.

**1.4.16 Carrier solution**, C (see figure 1).

Use grade 1 water (ISO 3696), degassed by reduced pressure.