

Soil improvers and growing media - Determination of the aerobic biological activity - Part 1: Oxygen uptake rate (OUR)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 16087-1:2020 sisaldab Euroopa standardi EN 16087-1:2020 ingliskeelset teksti.	This Estonian standard EVS-EN 16087-1:2020 consists of the English text of the European standard EN 16087-1:2020.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 15.01.2020.	Date of Availability of the European standard is 15.01.2020.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 65.080

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Soil improvers and growing media - Determination of the
aerobic biological activity - Part 1: Oxygen uptake rate
(OUR)

Amendements du sol et supports de culture -
Détermination de l'activité biologique aérobie - Partie
1 : Cinétique d'absorption de l'oxygène (OUR)

Bodenverbesserungsmittel und Substrate -
Bestimmung der aeroben biologischen Aktivität - Teil
1: Sauerstoffaufnahme (OUR)

This European Standard was approved by CEN on 21 October 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

Page

European Foreword.....	3
1 Scope.....	4
2 Normative references.....	4
3 Terms and Definitions.....	4
4 Principle	4
5 Apparatus.....	4
5.1 Testing facility.....	4
5.2 Pressure transducer.....	4
5.3 CO ₂ -absorbent containing unit.....	5
5.4 Reaction vessel.....	5
5.5 Mixing device	5
5.6 Balance.....	5
5.7 pH meter	5
5.8 Dispenser	5
5.9 Glassware	5
5.10 Sieve	5
6 Reagents	5
6.1 Water of class 3	5
6.2 pH buffer.....	5
6.3 Macronutrient solution.....	5
6.4 Micronutrient solution.....	5
6.5 Complete nutrient solution	6
6.6 Nitrification inhibitor	6
6.7 CO ₂ -absorbent	6
6.8 NaOH (0,5 mol/l)	6
6.9 HCl (0,5 mol/l).....	6
7 Procedure.....	6
7.1 Sample preparation.....	6
7.2 Determination of moisture content and organic matter content.....	6
7.3 Starting the procedure	6
7.4 Respiration measurement.....	7
8 Calculations.....	8
8.1 Theoretical background	8
8.2 Calculations.....	8
9 Test report.....	9
Annex A (informative) Validation	10
Annex B (informative) Specific information on the OUR-test	11
Bibliography.....	13

European Foreword

This document (EN 16087-1:2020) has been prepared by Technical Committee CEN/TC 223 “Soil improvers and growing media”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2020, and conflicting national standards shall be withdrawn at the latest by July 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 16087-1:2011.

The main changes compared with the previous edition are as follows:

- For the balance (5.6) requirements are added;
- Clarification of sample preparation (7.1) is added;
- Formula 3 and 5 are corrected;
- The figures in Annex B have been updated;
- The Bibliography has been corrected.

SAFETY PRECAUTIONS — Care should be taken when handling substances of caustic nature or samples that may contain sharps or is of a dusty nature.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This document describes a method to determine the aerobic biological activity of growing media and soil improvers or constituents thereof by measuring the oxygen uptake rate (OUR). The oxygen uptake rate is an indicator of the extent to which biodegradable organic matter is being broken down within a specified time period. The method is not suitable for material with a content of particle sizes > 10 mm exceeding 20 %.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13039, *Soil improvers and growing media – Determination of organic matter content and ash*

EN 13040, *Soil improvers and growing media – Sample preparation for chemical and physical tests, determination of dry matter content, moisture content and laboratory compacted bulk density*

EN 45501, *Metrological aspects of non-automatic weighing instruments*

EN ISO 3696, *Water for analytical laboratory use – Specification and test methods (ISO 3696)*

3 Terms and Definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp/ui>

4 Principle

The material is suspended in water. The respiration rate (i.e. oxygen uptake rate) is estimated by measuring the pressure drop in the headspace (i.e. gas phase in the closed space above the water phase). The produced CO₂ (carbon dioxide) is removed by a suitable alkaline absorbent. The measurements are performed under defined conditions.

5 Apparatus

5.1 Testing facility

Temperature controlled room, climate cabinet or water bath, temperature adjustable to $(30 \pm 2) ^\circ\text{C}$.

5.2 Pressure transducer

Operating range 0 kPa to 20 kPa (accuracy $\pm 0,1$ kPa) and record for measuring 2 to 4 times per hour for seven days.