

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment – Medical image display systems –
Part 1: Evaluation methods**

**Appareils électromédicaux – Systèmes d'imagerie médicale –
Partie 1: Méthodes d'évaluation**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2009 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch

Tel.: +41 22 919 02 11

Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch

Tél.: +41 22 919 02 11

Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment – Medical image display systems –
Part 1: Evaluation methods**

**Appareils électromédicaux – Systèmes d'imagerie médicale –
Partie 1: Méthodes d'évaluation**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

XA

ICS 11.040.55

ISBN 2-8318-1073-5

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms, definitions, symbols and abbreviations.....	7
3.1 Terms and definitions	7
3.2 Symbols	10
3.3 Abbreviations	11
4 General	11
5 Prerequisites	11
6 Equipment and tools.....	12
6.1 LUMINANCE meter	12
6.2 ILLUMINANCE meter.....	12
6.3 Colour meter	12
6.4 TEST PATTERNS	13
7 Evaluation methods	14
7.1 General.....	14
7.2 Evaluation method table overview	14
7.3 Visual evaluation methods.....	16
7.3.1 General	16
7.3.2 Overall image quality evaluation	16
7.3.3 Greyscale resolution evaluation	17
7.3.4 LUMINANCE response evaluation.....	18
7.3.5 LUMINANCE uniformity evaluation.....	19
7.3.6 Chromaticity evaluation	19
7.3.7 Pixel faults evaluation.....	19
7.3.8 VEILING GLARE evaluation	20
7.3.9 Geometrical image evaluation.....	20
7.3.10 Angular viewing evaluation	21
7.3.11 Clinical evaluation	22
7.4 Quantitative evaluation methods.....	22
7.4.1 Basic LUMINANCE evaluation.....	22
7.4.2 Basic LUMINANCE evaluation without ambient light.....	23
7.4.3 LUMINANCE response evaluation.....	23
7.4.4 LUMINANCE evaluation of multiple displays	26
7.4.5 Chromaticity evaluation	26
7.4.6 Chromaticity evaluation of multiple displays.....	26
7.4.7 LUMINANCE uniformity evaluation.....	26
7.4.8 Viewing angle evaluation	26
Annex A (informative) Sample test reports	28
Annex B (informative) LUMINANCE measurement methods	43
Annex C (informative) Description of TEST PATTERNS	46
Bibliography.....	55
Index of defined terms	57

Figure 1 – Overall image quality evaluation using the TG18-QC TEST PATTERN.....	16
Figure 2 – Overall image quality evaluation using the TG18-OIQ TEST PATTERN.....	17
Figure 3 – Magnified view of TG18-MP TEST PATTERN showing the 8-bit and 10-bit markers	18
Figure 4 – A close-up of the TG18-CT TEST PATTERN.....	19
Figure 5 – The TG18-GV TEST PATTERN is displayed (left), a close-up of the centre of the TEST PATTERN when covered with a mask (right)	20
Figure 6 – Geometrical evaluation using the GD pattern	21
Figure 7 – Visual evaluation of viewing angle response	22
Figure 8 – Example of the measured LUMINANCE in relation to the standard LUMINANCE response function according to GREYSCALE STANDARD DISPLAY FUNCTION (GSDF)	25
Figure 9 – An example of the CONTRAST response computed from 18 grey levels as related to the expected CONTRAST response associated with the DICOM 3.14 [2] standard LUMINANCE response with a given tolerance limit (e.g. 15 %) [10].....	25
Figure B.1 – Method A, telescopic method	43
Figure B.2 – Method B, near-range LUMINANCE meter in combination with an ILLUMINANCE meter	44
Figure B.3 – Method C, frontal integrated LUMINANCE meter in combination with ILLUMINANCE meter	44
Figure B.4 – Method D, back integrated LUMINANCE meter in combination with ILLUMINANCE meter	45
Figure C.1 – Example of TG-18 QC pattern for a matrix size of 1536 × 2048.....	54
Table 1 – Overview to the definitions of physical parameters	10
Table 2 – TEST PATTERNS used for display testing	13
Table 3 – List of the evaluation methods that can be used for testing medical IMAGE DISPLAY SYSTEMS	15
Table A.1 – Acceptance test sample report of a diagnostic display	29
Table A.2 – Constancy test sample report of a diagnostic display	33
Table A.3 – Acceptance test sample report of a monochrome reviewing display	35
Table A.4 – Constancy test sample report of a monochrome reviewing display	37
Table A.5 – Acceptance test sample report of a colour reviewing display	39
Table A.6 – Constancy test sample report of a colour reviewing display	41
Table C.1 – Description of multi-purpose TEST PATTERNS.....	47
Table C.2 – TG18-QC pattern: LUMINANCE levels with 8-bit and [12-bit] pixel values and CX ratings.....	50
Table C.3 – The blurring characteristics of the CX reference set utilized in TG18-QC TEST PATTERNS [16]	51
Table C.4 – Evaluation criteria for the examples of the CLINICAL REFERENCE IMAGES	52
Table C.5 – Example description of TG-18 QC pattern for a matrix size of 1536 × 2048.....	53

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT – MEDICAL IMAGE DISPLAY SYSTEMS –

Part 1: Evaluation methods

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62563-1 has been prepared by subcommittee 62B: Diagnostic imaging equipment of technical committee 62: Electrical equipment in medical practice.

The text of this standard is based on the following documents:

Enquiry draft	Report on voting
62B/743/CDV	62B/768/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THIS INTERNATIONAL STANDARD, OR AS NOTED: SMALL CAPITALS.

A list of all parts of the IEC 62563 series, published under the general title *Medical electrical equipment – Medical image display systems*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This International Standard provides evaluation methods for testing IMAGE DISPLAY SYSTEMS used in MEDICAL ELECTRICAL EQUIPMENT and medical electrical systems for diagnostic imaging.

On site or after installation, two types of testing can be carried out. An acceptance test is carried out after a new IMAGE DISPLAY SYSTEM has been installed, or major modifications have been made to the existing IMAGE DISPLAY SYSTEM. Since an IMAGE DISPLAY SYSTEM may degrade over time, the constancy test is carried out by the user in a periodic cycle to verify that the performance is maintained for the intended use.

The standard describes various evaluation methods without dictating what particular tests shall be used for acceptance and/or constancy tests.

Rather, it is the intention of this standard to be a reference for other standards and guidelines specific to each modality or to be defined by national authorities who will refer to the evaluation methods of this standard and mention limiting values and frequencies for acceptance and constancy tests. Annex A shows sample reports of such a reference.

To maintain the homogeneity in the IEC standards for MEDICAL ELECTRICAL EQUIPMENT, IEC 61223-2-5, *Evaluation and routine testing in medical imaging departments – Part 2-5: Constancy tests – Image display devices* should be reviewed.

MEDICAL ELECTRICAL EQUIPMENT – MEDICAL IMAGE DISPLAY SYSTEMS –

Part 1: Evaluation methods

1 Scope

This part of IEC 62563 describes the evaluation methods for testing medical IMAGE DISPLAY SYSTEMS.

The scope of this International Standard is directed to practical tests that can be visually evaluated or measured using basic test equipment. More advanced or more quantitative measurements can be performed on these devices, but these are beyond the scope of this document.

This standard applies to medical IMAGE DISPLAY SYSTEMS, which can display monochrome image information in the form of greyscale values on colour and greyscale IMAGE DISPLAY SYSTEMS (e.g. CATHODE RAY TUBE (CRT) monitors, FLAT PANEL DISPLAYS, PROJECTION SYSTEM). This standard applies to medical IMAGE DISPLAY SYSTEMS used for diagnostic (interpretation of medical images toward rendering clinical diagnosis) or viewing (viewing medical images for medical purposes other than for providing a medical interpretation) purposes and therefore having specific requirements in terms of image quality. Head mounted IMAGE DISPLAY SYSTEMS and IMAGE DISPLAY SYSTEMS used for confirming positioning and for operation of the system are not covered by this standard.

It is not in the scope of this standard to define the requirements of acceptance and constancy tests nor the frequencies of constancy tests.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60788:2004, *Medical electrical equipment – Glossary of defined terms*

ISO 11664-1:2007, *Colorimetry – Part 1: CIE standard colorimetric observers*

CIE S 010/E:2004 *Photometry – The CIE system of physical photometry*

3 Terms, definitions, symbols and abbreviations

3.1 Terms and definitions

For the purpose of this document, the terms and definitions given in IEC 60788:2004 and the following apply.

3.1.1

accuracy

closeness of agreement between a test result and the accepted reference value

[ISO 5725-1:1994, definition 3.6]