

**Radiation protection instrumentation - Passive
integrating dosimetry systems for environmental and
personal monitoring - Part 1: General characteristics
and performance requirements**

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 62387-1:2012 sisaldab Euroopa standardi EN 62387-1:2012 ingliskeelset teksti.	This Estonian standard EVS-EN 62387-1:2012 consists of the English text of the European standard EN 62387-1:2012.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 10.02.2012.	Date of Availability of the European standard is 10.02.2012.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.280

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:
Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

**Radiation protection instrumentation -
Passive integrating dosimetry systems for environmental and personal
monitoring -
Part 1: General characteristics and performance requirements
(IEC 62387-1:2007, modified)**

Instrumentation pour la radioprotection -
Systèmes dosimétriques intégrés passifs
pour la surveillance de l'environnement et
de l'individu -
Partie 1: Caractéristiques générales et
exigences de fonctionnement
(CEI 62387-1:2007, modifiée)

Strahlenschutz-Messgeräte -
Passive, integrierende Dosimetriesysteme
zur Umwelt- und Personenüberwachung -
Teil 1: Allgemeine Eigenschaften und
Leistungsanforderungen
(IEC 62387-1:2007, modifiziert)

This European Standard was approved by CENELEC on 2012-01-02. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Contents

Foreword	5
Introduction	6
1 Scope and object	7
2 Normative references	8
3 Terms and definitions	8
4 Units and symbols	16
5 General test procedures	16
5.1 Basic test procedures	16
5.2 Test procedures to be considered for every test	17
6 Performance requirements: summary	18
7 Capability of a dosimetry system	18
7.1 General	18
7.2 Measuring range and type of radiation	18
7.3 Rated ranges of the influence quantities	18
7.4 Maximum rated measurement time t_{\max}	18
7.5 Reusability	18
7.6 Model function	18
7.7 Example for the capabilities of a dosimetry system	19
8 Requirements for the design of the dosimetry system	19
8.1 General	19
8.2 Indication of the dose value (dosimetry system)	19
8.3 Assignment of the dose value to the dosimeter (dosimetry system)	20
8.4 Information given on the devices (reader and dosimeter)	20
8.5 Retention and removal of radioactive contamination (dosimeter)	20
8.6 Algorithm to evaluate the indicated value (dosimetry system)	20
8.7 Use of dosimeters in mixed radiation fields (dosimetry system)	20
9 Instruction manual	21
9.1 General	21
9.2 Specification of the technical data	21
10 Software, data and interfaces of the dosimetry system	22
10.1 General	22
10.2 Requirements	22
10.3 Method of test	25
11 Radiation performance requirements and tests (dosimetry system)	27
11.1 General	27
11.2 Coefficient of variation	27
11.3 Non-linearity	28
11.4 Overload characteristics, after-effects, and reusability	29
11.5 Radiation energy and angle of incidence for $H_p(10)$ or $H^*(10)$ dosimeters	30
11.6 Radiation energy and angle of incidence for $H_p(0,07)$ dosimeters	32
11.7 Over response to radiation incidence from the side of an $H_p(10)$ or $H_p(0,07)$ dosimeter	34
11.8 Indication of the presence of beta dose for $H_p(0,07)$ whole body dosimeters	35
12 Response to mixed irradiations (dosimetry system)	35
12.1 Requirements	35
12.2 Method of test	35

12.3	Interpretation of the results	36
13	Environmental performance requirements and tests	37
13.1	General.....	37
13.2	Ambient temperature and relative humidity (dosemeter).....	37
13.3	Light exposure (dosemeter)	38
13.4	Dose build-up, fading, self-irradiation, and response to natural radiation (dosemeter)	39
13.5	Sealing (dosemeter).....	40
13.6	Reader stability (reader)	40
13.7	Ambient temperature (reader)	41
13.8	Light exposure (reader).....	41
13.9	Primary power supply (reader)	42
14	Electromagnetic performance requirements and tests (dosimetry system).....	43
14.1	General.....	43
14.2	Requirement	43
14.3	Method of test.....	44
14.4	Interpretation of the results	44
15	Mechanical performance requirements and tests	44
15.1	General requirement	44
15.2	Drop (dosemeter).....	45
16	Documentation	45
16.1	Type test report	45
16.2	Certificate issued by the laboratory performing the type test	45
Annex A (normative)	Confidence limits	55
A.1	General.....	55
A.2	Confidence interval for the mean, \bar{x}	56
A.3	Confidence interval for a combined quantity	56
Annex B (informative)	Causal connection between readout signals, indicated value and measured value.....	58
Annex C (informative)	Overview of the necessary actions that have to be performed for a type test according to this standard	59
Annex D (informative)	Usage categories of passive dosimeters	61
Annex ZA (normative)	Normative references to international publications with their corresponding European publications	62
Annex ZB (informative)	Uncertainty of dosimetry systems	64
Annex ZC (informative)	Conversion coefficients $h_{pK}(0.07;S,\alpha)$ and $h_{pK}(0.07;R,\alpha)$ from air kerma, K_a , to the dose equivalent $H_p(0.07)$ for radiation qualities defined in ISO 4037-1 and the rod, pillar and slab phantom	65
Annex ZD (informative)	Computational method of test for mixed irradiations.....	66
Bibliography	68

Figures

Figure A.1 – Test for confidence interval	55
Figure B.1 – Data evaluation in dosimetry systems.....	58
Figure ZD.1 – Flow chart of a computer program to perform tests according to 12.2.....	67

Tables

Table 1 – Symbols.....	47
Table 2 – Reference conditions and standard test conditions	49
Table 3 – Performance requirements for $H_p(10)$ dosimeters	50
Table 4 – Performance requirements for $H_p(0,07)$ dosimeters	51
Table 5 – Performance requirements for $H^*(10)$ dosimeters.....	52
Table 6 – Environmental performance requirements for dosimeters and readers	53
Table 7 – Electromagnetic disturbance performance requirements for dosimetry systems according to Clause 14.....	54
Table 8 – Mechanical disturbances performance requirements for dosimeters.....	54
Table A.1 – Student's t -value for a double sided 95 % confidence interval.....	56
Table C.1 – Schedule for a type test of a dosimeter for $H_p(10)$ fulfilling the requirements within the mandatory ranges.....	59
Table D.1 – Usage categories of passive dosimeters.....	61
Table ZC.1 – Conversion coefficients $h_{pK}(0,07;S,\alpha)$ and $h_{pK}(0,07;R,\alpha)$ from air kerma, K_a , to the dose equivalent $H_p(0.07)$ for radiation qualities defined in ISO 4037-1 and for the rod, pillar, and slab phantom	65

Foreword

This document (EN 62387-1:2012) consists of the text of IEC 62387-1:2007 prepared by IEC/SC 45B, "Radiation protection instrumentation", of IEC/TC 45, "Nuclear instrumentation", together with the common modifications prepared by CLC/TC 45B, "Radiation protection instrumentation".

The following dates are fixed:

- latest date by which this document has to be implemented (dop) 2013-01-02
at national level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2015-01-02

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Clauses, subclauses, notes, tables, figures and annexes which are additional to those in IEC 62387-1:2007 are prefixed "Z".

In this document, the common modifications to IEC 62387-1:2007 are indicated by a vertical line in the left margin of the text.

The main objectives of EN 62387-1 are to

- specify performance requirements for complete dosimetry systems including detectors, dosimeters, readers, and additional equipment. In addition, the corresponding methods of test to check that these requirements are met are given in detail,
- harmonize requirements for all types of passive dosimetry systems detecting external photon and beta radiation,
- specify the use of the operational quantities according to ICRU 51,
- harmonize tests using radiation with relevant ISO standards on reference radiation and calibration: ISO 4037 for photon radiation, ISO 6980 for beta radiation and ISO 8529 for neutron radiation. For this reason, no conversion coefficients from air kerma (or absorbed dose or fluence) to the operational quantities are given in this standard, except in case the necessary conversion coefficients are not included in the respective ISO standard. Those given in the ISO-standards are applicable,
- incorporate basic terms of the concept that a result of a measurement essentially consists of a value and an associated uncertainty, as laid down in the introductions of IEC 311 and EN 60359 and refer the reader to an IEC technical report for complete uncertainty analysis in radiation protection measurements and to the GUM,
- align CENELEC performance requirements on dosimetry systems for measuring personal dose equivalents with the recommendations on accuracy stated in the ICRP Publication 75: *General Principles for the Radiation Protection of Workers*. Further information is given in the informative Annex ZB.

Introduction

A dosimetry system may consist of the following elements:

- a) a passive device, referred to here as a *detector*, which, after the presence of radiation, provides and stores a signal for use in measuring one or more quantities of the incident radiation field;
- b) a *dosemeter*, that incorporates some means of identification and contains one or more detectors;
- c) a *reader* which is used to readout the stored information (signal) from the detector, in order to determine the radiation dose;
- d) a *computer* with appropriate *software* to control the reader, store the signals transmitted from the reader, calculate, display and store the evaluated dose in the form of an electronic file or paper copy;
- e) *additional equipment* and documented procedures (instruction manual) for performing associated processes such as deleting stored dose information, cleaning dosimeters, or those needed to ensure the effectiveness of the whole system.

1 Scope and object

This European Standard applies to all kinds of passive dosimetry systems that are used for measuring

- the personal dose equivalent $H_p(10)$ (for whole body dosimetry),
- the personal dose equivalent $H_p(0,07)$ (for both whole body and extremity dosimetry), or
- the ambient dose equivalent $H^*(10)$ (for environmental dosimetry).

It applies to dosimetry systems that measure external photon or beta radiation in the dose range between 0,01 mSv and 10 Sv and in the energy ranges given in the following Table. All the energy values are mean energies with respect to the prevailing dose quantity. The dosimetry systems usually use electronic devices for the data evaluation and thus are often computer controlled.

Measuring quantity	Mandatory energy range for photon radiation	Maximum energy range for testing photon radiation	Mandatory energy range for beta-particle radiation	Maximum energy range for testing beta-particle radiation
$H_p(10)$, $H^*(10)$	80 keV to 1,25 MeV	12 keV to 7 MeV	---	---
$H_p(0,07)$	30 keV to 250 keV	8 keV to 7 MeV	0,8 MeV almost equivalent to an E_{max} of 2,27 MeV	0,07 MeV ^a to 1,2 MeV almost equivalent to E_{max} from 0,225 MeV to 3,54 MeV
^a For beta-particle radiation, an energy of 0,07 MeV is required to penetrate the dead layer of skin of 0,07 mm (almost equivalent to 0,07 mm of ICRU tissue).				

NOTE 1 In this standard, “dose” means personal or ambient dose equivalent, unless otherwise stated.

NOTE 2 For $H_p(10)$ and $H^*(10)$ no beta radiation is considered. Reasons: 1) $H_p(10)$ and $H^*(10)$ are a conservative estimate for the effective dose which is not a suitable quantity for beta radiation. 2) No conversion coefficients are available in ICRU 56, ICRU 57 or ISO 6980.

NOTE 3 The maximum energy ranges are the energy limits within which type tests according to this standard are possible.

In addition, this standard can be applied for testing neutron dosimetry systems concerning the design (Clause 8), the instruction manual (Clause 9), the software (Clause 10), environmental influences (Clause 13), electromagnetic influences (Clause 14), mechanical influences (Clause 15), and the documentation (Clause 16). The test utilizing radiation (Clauses 13 to 15) shall be done with neutron reference radiation qualities according to the ISO 8529 series.

In some countries the presence of beta dose has to be indicated by dosimeters worn on the trunk. Such an indication of the presence of beta dose is not a measurement. For that reason, a specific subclause (11.8) deals with the indication of the presence of beta dose.

This standard is intended to be applied to dosimetry systems that are capable of evaluating doses in the required quantity and unit (Sv) from readout signals in any quantity and unit. The only correction that may be applied to the evaluated dose (indicated value) is the one resulting from natural background radiation using extra dosimeters.

NOTE 4 The correction due to natural background may be made before or after the dose calculation.

Usually, a dosimetry system is not able to measure all quantities given above. Thus, the systems shall only be tested with regards to those quantities and types of radiation it is intended to be used for. Annex D gives further guidelines to define specific usage categories.

Full compliance with this standard is given if the requirements for the mandatory ranges given in Tables 3 to 5 are fulfilled. If the customer or manufacturer requires extended ranges then the test should also be performed as specified in this standard, i.e. the requirements given in Tables 3 to 5 apply, too. The range of any influence quantity stated by the manufacturer is called rated range. Thus, dosimetry systems can be classified by stating a set of ranges (for example, for dose, for energy, for temperature) within which the requirements stated in this standard are met (Capabilities of the system, see Clause 7). In addition, usage categories are given in Annex D with respect to different measuring capabilities.

For the dosimetry systems described above, this standard specifies general characteristics, general test procedures and performance requirements, radiation characteristics as well as environmental, electrical, mechanical, software and safety characteristics.

A dosimetry system may be tested with regards to different quantities at different times. In case the dosimetry system was changed since the previous test, a new test with regards to quantities tests formerly may be necessary.

The absolute calibration of the dosimetry system is not checked during a type test according to this standard as only system properties are of interest. The absolute calibration is checked during a routine test.

2 Normative references

For normative references, see the normative Annex ZA.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

For definitions related to measurements in general, definitions were taken from IEC 60050-300, Part 311, from IEC 60050-393 and from IEC 60050-394. A very limited number of definitions was taken from ISO 4037-3 and the ISO Guide to the Expression of Uncertainty in Measurement (GUM).

The references are given in brackets []. The information following the brackets is specific to this standard and is not originating from the given source.

A word between parentheses () in the title of a definition is a qualifier that may be skipped if there is no danger of confusion with a similar term.

The terms are listed in alphabetical order.

3.1

ambient dose equivalent

$H^*(d)$

at a point in a radiation field, dose equivalent that would be produced by the corresponding expanded and aligned field, in the ICRU sphere at a depth, d , on the radius opposing the direction of the aligned field

[SOURCE: ICRU 51]

Note 1 to entry: The recommended depth, d , for environmental monitoring in terms of $H^*(d)$ is 10 mm, and $H^*(d)$ may be written as $H^*(10)$. [IEV 393-14-95]

3.2

calibration factor

N_0

quotient of the conventional true value of a quantity $C_{r,0}$ and the indicated value $G_{r,0}$ at the point of test for a reference radiation under reference conditions. It is expressed as

$$N_0 = \frac{C_{r,0}}{G_{r,0}}$$

Note 1 to entry: The reciprocal of the calibration factor is equal to the response under reference conditions. In contrast to the calibration factor, which refers to the reference conditions only, the response refers to any conditions prevailing at the time of measurement.

[SOURCE: ISO 4037-3, Definition 3.2.12, modified]

Note 2 to entry: This definition is of special importance for non-linear dosimeters.

Note 3 to entry: The reference value $C_{r,0}$ for the dose is given in Table 2.