

KIIRGUSKAITSE

Dosimeetrite ja doosi kiiruse mõõtseseadmete kalibreerimiseks ning nende footoni energiast sõltuva koste määramiseks kasutatav röntgeni- ja gammareferentskiirgus

Osa 2: Kiirguskaitseline dosimeetria energiavahemikes 8 keV kuni 1,3 MeV ja 4 MeV kuni 9 MeV

Radiological protection

X and gamma reference radiation for calibrating dosemeters and doserate meters and for determining their response as a function of photon energy

Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV (ISO 4037-2:2019, identical)

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

<p>See Eesti standard EVS-ISO 4037-2:2019 „Kiirguskaitse. Dosimeetrite ja doosi kiiruse mõõteseadmete kalibreerimiseks ning nende footoni energiast sõltuva koste määramiseks kasutatav röntgeni- ja gammareferentskiirgus. Osa 2: Kiirguskaitsetiline dosimeetria energiavahemikes 8 keV kuni 1,3 MeV ja 4 MeV kuni 9 MeV“ sisaldab rahvusvahelise standardi ISO 4037-2:2019 „Radiological protection. X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy. Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV“ identset ingliskeelset teksti.</p>	<p>This Estonian Standard EVS-ISO 4037-2:2019 consists of the identical English text of the International Standard ISO 4037-2:2019 „Radiological protection. X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy. Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV“.</p>
<p>Ettepaneku rahvusvahelise standardi ümbertrüki meetodil ülevõtuks on esitanud EVS/TK 28, standardi avaldamist on korraldanud Eesti Standardikeskus.</p>	<p>Proposal to adopt the International Standard by reprint method has been presented by EVS/TK 28, the Estonian Standard has been published by the Estonian Centre for Standardisation.</p>
<p>Standard EVS-ISO 4037-2:2019 on jõustunud sellekohase teate avaldamisega EVS Teataja 2019. aasta juunikuu numbris.</p>	<p>Standard EVS-ISO 4037-2:2019 has been endorsed with a notification published in the June 2019 issue of the official bulletin of the Estonian Centre for Standardisation.</p>
<p>Standard on kättesaadav Eesti Standardikeskusest.</p>	<p>The standard is available from the Estonian Centre for Standardisation.</p>

Käsitlusala

See dokument kirjeldab röntgeni- ja gammareferentskiirguse dosimeetria protseduure kiirguskaitse mõõteseadmete kalibreerimiseks energiavahemikus ligikaudu 8 keV kuni 1,3 MeV ja 4 MeV kuni 9 MeV ning üle 1 $\mu\text{Gy}/\text{h}$ õhukerma kiiruste jaoks. Käsitletavad mõõtesuurused on vabalt õhus tekkiv kerma K_a ja rahvusvahelise kiirgusühikute ja kiirgusmõõtmiste komisjoni (International Commission on Radiation Units and Measurement, ICRU^[2]) fantomiga seotud operatiivsuurused $H^*(10)$, $H_p(10)$, $H'(3)$, $H_p(3)$, $H'(0,07)$ ja $H_p(0,07)$ koos vastavate doosi kiirustega. Nende saamismeetodid on toodud standardis ISO 4037-1.

Seda dokumenti võib kasutada ka standardi ISO 4037-1:2019 lisades A, B ja C määratletud kiirgusparameetrite puhul, mis aga ei tähenda seda, et nendes lisades kirjeldatud kiirgusparameetrite kalibreerimistunnistus oleks vastavuses standardisarja ISO 4037 nõuetega.

Selles dokumendis esitatud nõuete ja meetodite eesmärk on doosi (kiiruse) üldise määramatuse, ligikaudu 6 % kuni 10 % ($k = 2$) saavutamine fantomiga seotud referentsväljade operatiivsuuruste korral. Selle nõude saavutamiseks on standardis ISO 4037-1 esitatud kaks referentsväljade saamismeetodit.

Esimene meetod seisneb selliste „vastavuses olevate referentsväljade“ tekitamises, mis järgivad nõudeid niivörd täpselt, et on võimalik kasutada soovitatud teisendustegureid. „Vastavuses olevate referentsväljade“ spektraaljaotuste ainult väikeste erinevuste, võrreldes nominaalse tekitamiseks referentsväljadega, olemasolu on kinnitatud toimingutega, mis on antud ja üksikasjalikult kirjeldatud selles dokumendis. Vastavuses olevate kiirguse referentsväljade puhul on standardis ISO 4037-3 soovitatavad teisendustegurid toodud ainult kindlate allika ja dosimeetri vahekauguste korral, nt 1,0 m ja 2,5 m. Muude vahekauguste korral peab kasutaja ise otsustama, kas neid teisendustegureid võib kasutada.

Teine meetod on tekitada „kirjeldatud referentsvälju“. Seda tehakse kas teisendustegurite määramisel spektromeetria abil või vajalik väärthus mõõdetakse vahetult sekundaarsete standardsete dosimeetrite abil. Seda meetodit rakendatakse mis tahes kiirgusparameetri korral, mis tahes mõõdetavale suurusele ja kui see on rakendatav, siis ka iga fantoomi ja kiirguse vahelisele langemisnurgale. Teisendustegureid on võimalik määräta suvalise vahekauguse korral, kui on tagatud, et õhukerma kiirus ei ole alla 1 $\mu\text{Gy}/\text{h}$.

Mõlemad meetodid vajavad referentsvälja jaoks laetud osakeste tasakaalu. Kusjuures see pole alati kindlaks määratud töökohal olevas väljas, mille jaoks dosimeeter tuleb kalibreerida. See kehtib eriti footoni energia korral, mil puudub referentssügavusel d sellele omane laetud osakeste tasakaal, mis omakorda sõltub energia ja referentssügavuse d tegelikust kombinatsioonist. Elektronid, mille energia on üle 65 keV, 0,75 MeV ja 2,1 MeV, võivad läbida vastavalt 0,07 mm, 3 mm ja 10 mm ICRU kudet ja nendest väärustumest suuremate footoni energiate korral loetakse kiirgusparameetrid sellel sügavusel defineeritud suuruste jaoks sisemise tasakaaluta laetud osakeste kiirgusparameetriteks.

See dokument ei ole rakendatav pulseerivate referentsväljade dosimeetria korral.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 17.240

Standardite reproduutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact the Estonian Centre for Standardisation:
Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 Standard instrument	3
4.1 General	3
4.2 Calibration of the standard instrument	3
4.3 Energy dependence of the response of the standard instrument	3
5 Conversion from the measured quantity air kerma, K_a, to the required phantom related measuring quantity	4
5.1 General	4
5.2 Determination of conversion coefficients	6
5.2.1 General	6
5.2.2 Calculation of conversion coefficients from spectral fluence	6
5.3 Validation of reference fields and of listed conversion coefficients using dosimetry	7
6 Direct calibration of the reference field in terms of the required phantom related measuring quantity	8
7 Measurement procedures applicable to ionization chambers	8
7.1 Geometrical conditions	8
7.2 Chamber support and stem scatter	8
7.3 Location and orientation of the standard chamber	8
7.4 Measurement corrections	8
7.4.1 General	8
7.4.2 Corrections for air temperature, pressure and humidity variation from reference calibration conditions	9
7.4.3 Corrections for radiation-induced leakage, including ambient radiation	9
7.4.4 Incomplete ion collection	10
7.4.5 Beam non-uniformity	10
8 Additional procedures and precautions specific to gamma radiation dosimetry using radionuclide sources	10
8.1 Use of certified source output	10
8.2 Use of electron equilibrium caps	10
8.3 Radioactive source decay	10
8.4 Radionuclide impurities	10
8.5 Interpolation between calibration positions	10
9 Additional procedures and precautions specific to X-radiation dosimetry	11
9.1 Variation of X-radiation output	11
9.2 Monitor	11
9.3 Adjustment of air kerma rate	11
10 Dosimetry of reference radiation at photon energies between 4 MeV and 9 MeV	12
10.1 Dosimetric quantities	12
10.2 Measurement of the dosimetric quantities	12
10.2.1 General	12
10.2.2 Air kerma (rate)	13
10.2.3 Phantom related operational quantities $H^*(10)$, $H_p(10)$, $H'(3)$ and $H_p(3)$	13
10.3 Measurement geometry	13
10.4 Monitor	13
10.5 Determination of air kerma (rate) free-in-air	14
10.5.1 General	14

10.5.2	Measurement conditions	14
10.5.3	Direct measurement with an ionization chamber	14
10.5.4	Determination of air kerma (rate) from photon fluence (rate)	17
11	Uncertainty of measurement	18
11.1	General	18
11.2	Components of uncertainty	18
11.2.1	General	18
11.2.2	Uncertainties in the calibration of a secondary standard	18
11.2.3	Uncertainties in the measurements of the reference radiation due to the standard instrument and its use	19
11.3	Statement of uncertainty	19
Annex A (normative)	Technical details of the instruments and their operation	20
Annex B (informative)	Measurement of photon spectra	23
Bibliography		26

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 4037-2:1997), which has been technically revised.

A list of all the parts in the ISO 4037 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO 4037-2:2019 incorporates the following corrections:

- In 10.5.2.2, the subscripts to the values have been reapplied;
- In Table 5, the headers in columns 4 and 5 have been reinserted.

Introduction

The maintenance release of this document incorporates the improvements to high voltage generators from 1996 to 2017 (e.g., the use of high frequency switching supplies providing nearly constant potential), and the spectral measurements at irradiation facilities equipped with such generators (e.g., the catalogue of X-ray spectra by Ankerhold^[1]). It also incorporates all published information with the aim to adjust the requirements for the technical parameters of the reference fields to the targeted overall uncertainty of about 6 % to 10 % for the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[2]. It does not change the general concept of the existing ISO 4037.

ISO 4037, focusing on photon reference radiation fields, is divided into four parts. ISO 4037-1 gives the methods of production and characterization of reference radiation fields in terms of the quantities spectral photon fluence and air kerma free-in-air. This document describes the dosimetry of the reference radiation qualities in terms of air kerma and in terms of the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[2]. ISO 4037-3 describes the methods for calibrating and determining the response of dosemeters and doserate meters in terms of the phantom related operational quantities of the ICRU^[2]. ISO 4037-4 gives special considerations and additional requirements for calibration of area and personal dosemeters in low energy X reference radiation fields, which are reference fields with generating potential lower or equal to 30 kV.

In this document, two methods are given to determine the phantom related operational quantities. Both methods need a reference field according to ISO 4037-1. The first method requires the dosimetry with respect to air kerma free-in-air and after that the selected operational quantity is derived by the application of a conversion coefficient that relates the air kerma free-in-air to the selected operational quantity. For matched reference fields, this conversion coefficient is taken from ISO 4037-3, for characterized reference fields the conversion coefficient is determined using spectrometry. The second method, applicable for characterized reference fields, requires the direct dosimetry with respect to the selected operational quantity. For all calibrations secondary standard instruments are required, which have a nearly constant energy dependence of the response to the selected quantity.

The general procedures described in ISO 29661 are used as far as possible in this document. Also, the used symbols are in line with ISO 29661.

Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV

1 Scope

This document specifies the procedures for the dosimetry of X and gamma reference radiation for the calibration of radiation protection instruments over the energy range from approximately 8 keV to 1,3 MeV and from 4 MeV to 9 MeV and for air kerma rates above 1 $\mu\text{Gy}/\text{h}$. The considered measuring quantities are the air kerma free-in-air, K_a , and the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[2], $H^*(10)$, $H_p(10)$, $H'(3)$, $H_p(3)$, $H'(0,07)$ and $H_p(0,07)$, together with the respective dose rates. The methods of production are given in ISO 4037-1.

This document can also be used for the radiation qualities specified in ISO 4037-1:2019, Annexes A, B and C, but this does not mean that a calibration certificate for radiation qualities described in these annexes is in conformity with the requirements of ISO 4037.

The requirements and methods given in this document are targeted at an overall uncertainty ($k = 2$) of the dose(rate) of about 6 % to 10 % for the phantom related operational quantities in the reference fields. To achieve this, two production methods of the reference fields are proposed in ISO 4037-1.

The first is to produce "*matched reference fields*", which follow the requirements so closely that recommended conversion coefficients can be used. The existence of only a small difference in the spectral distribution of the "*matched reference field*" compared to the nominal reference field is validated by procedures, which are given and described in detail in this document. For matched reference radiation fields, recommended conversion coefficients are given in ISO 4037-3 only for specified distances between source and dosimeter, e.g., 1,0 m and 2,5 m. For other distances, the user has to decide if these conversion coefficients can be used.

The second method is to produce "*characterized reference fields*". Either this is done by determining the conversion coefficients using spectrometry, or the required value is measured directly using secondary standard dosimeters. This method applies to any radiation quality, for any measuring quantity and, if applicable, for any phantom and angle of radiation incidence. The conversion coefficients can be determined for any distance, provided the air kerma rate is not below 1 $\mu\text{Gy}/\text{h}$.

Both methods require charged particle equilibrium for the reference field. However this is not always established in the workplace field for which the dosimeter shall be calibrated. This is especially true at photon energies without inherent charged particle equilibrium at the reference depth d , which depends on the actual combination of energy and reference depth d . Electrons of energies above 65 keV, 0,75 MeV and 2,1 MeV can just penetrate 0,07 mm, 3 mm and 10 mm of ICRU tissue, respectively, and the radiation qualities with photon energies above these values are considered as radiation qualities without inherent charged particle equilibrium for the quantities defined at these depths.

This document is not applicable for the dosimetry of pulsed reference fields.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4037-1, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 1: Radiation characteristics and production methods*

ISO 4037-3, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

ISO 4037-4, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 4: Calibration of area and personal dosimeters in low energy X reference radiation fields*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4037-1, ISO 29661, ISO 80000-10, ISO/IEC Guide 99, and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

ionization chamber

ionization detector consisting of a chamber filled with a suitable gas, in which an electric field, insufficient to induce gas multiplication, is provided for the collection at the electrodes of charges associated with the ions and the electrons produced in the sensitive volume of the detector by the ionizing radiation^[3]

Note 1 to entry: The ionization chamber includes the sensitive volume, the collecting and polarizing electrodes, the guard electrode, if any, the chamber wall, the parts of the insulator adjacent to the sensitive volume and any necessary caps to ensure electron equilibrium.

3.2

ionization chamber assembly

ionization chamber (3.1) and all other parts to which the chamber is permanently attached, except the measuring assembly

Note 1 to entry: For a cable-connected chamber, it includes the stem, the electrical fitting and any permanently attached cable or pre-amplifier. For a thin-window chamber, it includes any block of material in which the ionization chamber is permanently embedded.