

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-83: Particular requirements for the basic safety and essential performance
of home light therapy equipment**

**Appareils électromédicaux –
Partie 2-83: Exigences particulières pour la sécurité de base et les performances
essentielles des appareils de luminothérapie à domicile**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-83: Particular requirements for the basic safety and essential
performance of home light therapy equipment**

**Appareils électromédicaux –
Partie 2-83: Exigences particulières pour la sécurité de base et les
performances essentielles des appareils de luminothérapie à domicile**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.60

ISBN 978-2-8322-6953-4

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

| | |
|--|----|
| FOREWORD..... | 3 |
| INTRODUCTION..... | 5 |
| 201.1 Scope, object and related standards | 6 |
| 201.2 Normative references..... | 7 |
| 201.3 Terms and definitions..... | 8 |
| 201.4 General requirements | 10 |
| 201.5 General requirements for testing ME EQUIPMENT | 10 |
| 201.6 Classification of ME EQUIPMENT and ME SYSTEMS | 10 |
| 201.7 ME EQUIPMENT identification, marking and documents | 12 |
| 201.8 Protection against electrical HAZARDS from ME EQUIPMENT..... | 15 |
| 201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS..... | 15 |
| 201.10 Protection against unwanted and excessive radiation HAZARDS | 15 |
| 201.11 Protection against excessive temperatures and other HAZARDS | 17 |
| 201.12 Accuracy of controls and instruments and protection against hazardous outputs | 18 |
| 201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT | 18 |
| 201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)..... | 18 |
| 201.15 Construction of ME EQUIPMENT | 18 |
| 201.16 ME SYSTEMS | 18 |
| 201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS | 18 |
| 202 Electromagnetic disturbances – Requirements and tests | 18 |
| 206 Usability | 19 |
| 211 Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT..... | 19 |
| Annexes | 20 |
| Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS | 21 |
| Annex D (informative) Symbols on marking..... | 24 |
| Annex AA (informative) Particular guidance and rationale | 25 |
| Annex BB (informative) Protective eyewear for HOME LIGHT THERAPY EQUIPMENT | 28 |
| Bibliography..... | 29 |
| Index of defined terms used in this particular standard..... | 30 |
| Table 201.101 – EMISSION LIMITS for risk groups of HOME LIGHT THERAPY EQUIPMENT | 11 |
| Table 201.102 – Time criteria for risk groups of HOME LIGHT THERAPY EQUIPMENT | 12 |
| Table 201.103 – Applicable ANGLE OF ACCEPTANCE for the assessment of emitted OPTICAL RADIATION from HOME LIGHT THERAPY EQUIPMENT..... | 12 |
| Table 201.C.101 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts..... | 21 |
| Table 201.C.102 –Safety signs per HAZARD and per risk group..... | 22 |
| Table 201.C.103 – ACCOMPANYING DOCUMENTS, instructions for use | 22 |
| Table 201.C.104 –Caution statements per HAZARD and per risk group..... | 23 |
| Table 201.D.1 – General symbols | 24 |
| Table 201.D.2 – Safety signs | 24 |

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-83: Particular requirements for the basic safety and essential performance of home light therapy equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-83 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this International Standard is based on the following documents:

| FDIS | Report on voting |
|---------------|------------------|
| 62D/1682/FDIS | 62D/1687/RVD |

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;

- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This part of IEC 60601 has been prepared to provide safety requirements for HOME LIGHT THERAPY EQUIPMENT, based on IEC 60601-1 and its collateral standards. This equipment is intended to be used in the HOME HEALTHCARE ENVIRONMENT and is typically used by a LAY OPERATOR, who is familiar with this environment and the specific characteristics of lamps. Some requirements of IEC 60601-1-11 are amended to better suit this type of ME EQUIPMENT and the environment in which it is used.

HOME LIGHT THERAPY EQUIPMENT provides light therapy by means of eye-mediated photobiological effects (which can be visual or non-visual) and skin-mediated photobiological effects (non-visual only). Possible applications include pain relief, psoriasis treatment, and treatment of winter depression (seasonal affective disorder, SAD).

This document is developed because IEC 60601-2-57 [2]¹ only covers light source equipment providing light therapy by means of non-visual photobiological effects, which excludes an important group of light source equipment creating visual photobiological effects. Further, IEC 60601-2-57 focuses on radiation aspects and related markings but hardly provides any product-specific safety requirements. IEC 60335-2-113 [1] provides such specific requirements for household appliances with light sources for cosmetic and beauty care, but does not apply to equipment with medical purposes. IEC 60601-2-83 addresses all safety requirements for HOME LIGHT THERAPY EQUIPMENT and has taken over relevant requirements from [1] and [2].

This document is the first edition of IEC 60601-2-83. It is aligned with IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, IEC 60601-1-2:2014, IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013, and IEC 60601-1-11:2015.

¹ Numbers in square brackets refer to the Bibliography.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-83: Particular requirements for the basic safety and essential performance of home light therapy equipment

201.1 Scope, object and related standards

Clause 1 of the general standard² applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 60601 is applicable to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HOME LIGHT THERAPY EQUIPMENT, intended for use in the HOME HEALTHCARE ENVIRONMENT. HOME LIGHT THERAPY EQUIPMENT is typically used by a LAY OPERATOR.

The scope of this document includes all light sources except laser.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular requirements for the BASIC SAFETY and ESSENTIAL PERFORMANCE of HOME LIGHT THERAPY EQUIPMENT.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2014, IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013 and IEC 60601-1-11:2015 apply as modified in Clauses 202, 206 and 211, respectively. IEC 60601-1-3, IEC 60601-1-8, IEC 60601-1-10 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME

² The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x", where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered from 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
IEC 60601-1-6:2010/AMD1:2013

ISO 15223-1:2016, *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements*

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 62471:2006, *Photobiological safety of lamps and lamp systems*

ISO 3864-1:2011, *Graphical symbols – Safety colours and safety signs – Part 1: Design principles for safety signs and safety markings*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1, IEC 60601-1-2, IEC 60601-1-6 and IEC 60601-1-11 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE An index of defined terms is found on page 30.

Addition:

201.3.201

*** ANGLE OF ACCEPTANCE**

γ

plane angle within which a detector responds to OPTICAL RADIATION

Note 1 to entry: The ANGLE OF ACCEPTANCE can be controlled by apertures or optical elements.

Note 2 to entry: The ANGLE OF ACCEPTANCE is sometimes referred to as the field-of-view.

[SOURCE: IEC 60601-2-57:2011, 201.3.201, modified – Unit deleted.]

201.3.202

*** ANGULAR SUBTENSE**

α

visual angle subtended by the source or apparent source at the eye of an observer or at the point of measurement

Note 1 to entry: In this particular standard ANGULAR SUBTENSE is denoted by the full included angle, not the half angle.

[SOURCE: IEC 60601-2-57:2011, 201.3.202, modified – Unit deleted.]