

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Underwater acoustics – Hydrophones – Calibration of hydrophones –
Part 2: Procedures for low frequency pressure calibration**

**Acoustique sous-marine – Hydrophones – Étalonnage des hydrophones –
Partie 2: Procédures pour l'étalonnage à basse pression de fréquence**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Underwater acoustics – Hydrophones – Calibration of hydrophones –
Part 2: Procedures for low frequency pressure calibration**

**Acoustique sous-marine – Hydrophones – Étalonnage des hydrophones –
Partie 2: Procédures pour l'étalonnage à basse pression de fréquence**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-7346-3

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	8
2 Normative references	8
3 Terms and definitions	8
4 Symbols	12
5 Procedures for calibration.....	13
5.1 Principles	13
5.2 Field limitations.....	14
5.3 Schematic survey of procedures	14
5.4 Reporting of results.....	14
5.5 Recalibration periods	15
5.6 Temperature and pressure considerations for calibration	15
5.7 Preparation of transducers.....	15
6 Electrical measurements.....	15
6.1 Signal type.....	15
6.2 Earthing	15
6.3 Measurement of hydrophone output voltage.....	15
6.3.1 General	15
6.3.2 Electrical loading by measuring instrument.....	15
6.3.3 Electrical loading by extension cable	16
6.3.4 Cross-talk and acoustic interference	16
6.3.5 Integral pre-amplifier	16
6.4 Measurement of projector current	16
7 Calibration by hydrostatic excitation	16
7.1 General.....	16
7.2 Principle	16
7.2.1 Determination of the alternating pressure	16
7.2.2 Determination of the correction factor	18
7.2.3 Determination of the equivalent height.....	18
7.2.4 Calculation of the pressure sensitivity of hydrophone.....	20
7.3 Design of vibration system	20
7.4 Alternative method for hydrostatic excitation	20
7.5 Uncertainty	21
8 Calibration by piezoelectric compensation	21
8.1 General.....	21
8.2 Principle	21
8.2.1 Determination of the sound pressure	21
8.2.2 Determination of the characteristic constant	22
8.2.3 Calculation of the pressure sensitivity of hydrophone.....	23
8.3 Design of the calibration chamber.....	23
8.3.1 General	23
8.3.2 Low frequency chamber.....	23
8.3.3 High frequency chamber	23
8.4 Practical limitations of the piezoelectric compensation method	24
8.5 Relative calibration method.....	24

8.6	Uncertainty	25
9	Calibration by acoustic coupler reciprocity	25
9.1	General	25
9.2	Principle	25
9.2.1	Theory of acoustic coupler reciprocity	25
9.2.2	Procedures for the reciprocity calibration	26
9.2.3	Calculation of transfer impedance	27
9.2.4	Determination of acoustic compliance	27
9.3	Limitation of acoustic coupler reciprocity	27
9.3.1	Frequency limit	27
9.3.2	Hydrophone limit	28
9.4	Measurement	28
9.4.1	General	28
9.4.2	Evidence of interference effects	28
9.4.3	Reciprocity verification	28
9.4.4	Linearity verification	29
9.5	Uncertainty	29
10	Calibration by pistonphone	29
10.1	General	29
10.2	Principle	29
10.2.1	Determination of the sound pressure	29
10.2.2	Determination of the compliance of the medium	30
10.2.3	Calculation of the pressure sensitivity	30
10.3	Limitations	30
10.4	Relative calibration	31
10.4.1	Comparison with a reference transducer	31
10.4.2	Comparison using air–water pistonphone	31
10.5	Uncertainty	32
11	Calibration by vibrating column	32
11.1	General	32
11.2	Principle	32
11.2.1	General	32
11.2.2	Expression for the pressure	33
11.2.3	Determination of the sensitivity	34
11.3	Conditions of measurement	35
11.3.1	Mechanical	35
11.3.2	Acoustical	36
11.4	Relative calibration method	36
11.5	Uncertainty	37
12	Calibration by static pressure transducer	37
12.1	General	37
12.2	Principle	38
12.2.1	Theory of static pressure calibration	38
12.2.2	Determination of the sensitivity of static pressure transducer	38
12.2.3	Calculation of the pressure sensitivity	39
12.3	Limitations	39
12.4	Uncertainty	39
Annex A (informative)	Advanced acoustic coupler calibration methods	40

A.1	General.....	40
A.2	Acoustic-coupler calibration using a reference coupler with two reciprocal transducers and an auxiliary coupler with the same two transducers and a hydrophone to be calibrated.....	41
A.2.1	General	41
A.2.2	Theory	42
A.3	Acoustic-coupler calibration using a reference coupler with two reciprocal transducers and an auxiliary coupler with the same two transducers, a hydrophone to be calibrated, and a sound source	43
A.3.1	General	43
A.3.2	Theory	44
A.4	Acoustic-coupler calibration using a coupler, a reciprocal transducer, a projector, a hydrophone to be calibrated, and a subsidiary body of known compliance	45
A.4.1	General	45
A.4.2	Theory	45
Annex B (informative)	Assessment of uncertainty in the low frequency pressure calibration of hydrophone	48
B.1	General.....	48
B.2	Type A evaluation of uncertainty	48
B.3	Type B evaluation of uncertainty	48
B.4	Reported uncertainty.....	48
B.5	Common sources of uncertainty	49
Bibliography.....		51
Figure 1	– Diagram of calibration by hydrostatic excitation.....	17
Figure 2	– Schematic drawing of calibration by piezoelectric compensation	22
Figure 3	– Diagram of the chamber for high frequency	24
Figure 4	– Reciprocity coupler with three transducers: a projector P, a reciprocal transducer T, and a hydrophone H to be calibrated	26
Figure 5	– Comparison calibration using pistonphone with calibrated hydrophone.....	32
Figure 6	– Diagram of calibration by vibrating column	33
Figure 7	– Diagram of calibration by vibrating column using comparison	37
Figure 8	–Diagram of comparison calibration using static pressure transducer.....	38
Figure A.1	– Reference coupler with two transducers: a projector P and a reciprocal transducer T	42
Figure A.2	– Auxiliary coupler with three transducers: a projector P, a reciprocal transducer T and a hydrophone H to be calibrated	42
Figure A.3	– Auxiliary coupler with four transducers: a projector P, a reciprocal transducer T, a sound source S and a hydrophone H to be calibrated	44
Figure A.4	– Schematic drawing of the measuring system.....	46

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**UNDERWATER ACOUSTICS – HYDROPHONES –
CALIBRATION OF HYDROPHONES –****Part 2: Procedures for low frequency pressure calibration****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60565-2 has been prepared by IEC technical committee 87: Ultrasonics.

This first edition of IEC 60565-2, together with IEC 60565-1, replaces the second edition of IEC 60565 published in 2006. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- 1) IEC 60565 has been divided into two parts:
 - Part 1: Procedures for free-field calibration;
 - Part 2: Procedures for low frequency pressure calibration (this document).
- 2) A relative calibration method has been added to Clause 8: Calibration by piezoelectric compensation.

- 3) A relative calibration method has been added to Clause 11: Calibration by **vibrating column**.
- 4) Clause 12: Calibration by static pressure transducer, has been added.
- 5) Annex A: Equivalent circuit of the excitation system for calibration with a **vibrating column**, has been deleted.
- 6) Subclauses 9.6, 9.7 and 9.8 have been moved to form a new Annex A: Advanced acoustic coupler calibration methods.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
87/720/FDIS	87/723/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

NOTE Words in **bold** in the text are terms defined in Clause 3.

A list of all parts in the IEC 60565 series, published under the general title *Underwater acoustics – Hydrophones – Calibration of hydrophones*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

The purpose of this document is to establish the procedures for low frequency pressure calibrations of **hydrophones** in the frequency range from 0,01 Hz to several kilohertz.

To ensure the correctness of the calibrations, the **hydrophones** to be calibrated are "rigid" **hydrophones** with small size compared to the acoustic wavelength, and are not sensitive to vibration when calibrated.

Principles, procedures, and **uncertainties** of physical calibrations such as hydrostatic excitation, piezoelectric compensation, **pistonphone**, **vibrating column**, static pressure transducer, etc., and reciprocity calibrations in acoustic **couplers** are given in this document. Calibrations are carried out using one of these methods, depending on the different principles to be used, and its limitations to the sound field and the frequency range.

UNDERWATER ACOUSTICS – HYDROPHONES – CALIBRATION OF HYDROPHONES –

Part 2: Procedures for low frequency pressure calibration

1 Scope

This part of IEC 60565 specifies the methods for low frequency pressure calibration of **hydrophones** at frequencies from 0,01 Hz to several kilohertz depending on calibration method.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-801, *International Electrotechnical Vocabulary – Chapter 801: Acoustics and electroacoustics* (available at <http://www.electropedia.org/>)

IEC 60500:2017, *Underwater acoustics – Hydrophones – Properties of the hydrophone in the frequency range 1 Hz to 500 kHz*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60050-801, IEC 60500:2017 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

coupler

apparatus comprising a rigid fluid-filled chamber into which transducers and **hydrophones** can be inserted whose largest dimension is small compared to the wavelength

Note 1 to entry: In this document, the term small chamber is used interchangeably with **coupler**.

[SOURCE: IEC 60565:2006 [1]¹, 3.3, modified – In the definition, "small dimensions" has been replaced by "whose largest dimension is small compared to the wavelength".]

¹ Numbers in square brackets refer to the Bibliography.