

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Electroacoustics – Hearing aids –  
Part 13: Requirements and methods of measurement for electromagnetic  
immunity to mobile digital wireless devices**

**Électroacoustique – Appareils de correction auditive–  
Partie 13: Exigences et méthodes de mesure de l'immunité électromagnétique  
aux appareils numériques mobiles sans fil**





## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembé  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Electroacoustics – Hearing aids –  
Part 13: Requirements and methods of measurement for electromagnetic  
immunity to mobile digital wireless devices**

**Électroacoustique – Appareils de correction auditive–  
Partie 13: Exigences et méthodes de mesure de l'immunité électromagnétique  
aux appareils numériques mobiles sans fil**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 17.140.50; 33.100.20

ISBN 978-2-8322-7330-2

**Warning! Make sure that you obtained this publication from an authorized distributor.**

**Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD .....	3
INTRODUCTION .....	5
1 Scope .....	6
2 Normative references .....	6
3 Terms and definitions .....	6
4 Operation and function of the hearing aid .....	8
5 Requirements for electromagnetic immunity .....	8
5.1 General .....	8
5.2 Compliance criteria .....	9
6 Test procedures for immunity to radiated RF electromagnetic fields .....	11
6.1 General .....	11
6.2 Test setup .....	11
6.3 Hearing aid test setting .....	11
6.4 Determination of gain .....	12
6.5 Measurement of the input related ambient noise ( <i>IRAN</i> ) .....	12
6.6 Hearing aid output coupling during immunity test .....	13
6.7 Position of the hearing aid during immunity test .....	13
6.8 Measurement of the output related interference level ( <i>ORIL</i> ) .....	14
6.9 Calculation of the input related interference level ( <i>IRIL</i> ) .....	15
6.10 Report .....	15
7 Measurement uncertainty for immunity to radiated RF electromagnetic fields .....	16
Annex A (informative) Background for establishing test methods, performance criteria and test levels .....	17
A.1 General .....	17
A.2 Radiated RF electromagnetic fields – History of the test method .....	17
A.3 Performance criteria .....	19
A.4 Test field strengths – Bystander compatibility .....	19
A.5 Test field strengths – User compatibility .....	20
Bibliography .....	21
Figure 1 – Example of a test arrangement for hearing aid immunity measurements using a GTEM cell .....	11
Figure 2 – Examples of input-output response curves at 1 kHz and the determination of gain at an input SPL of 55 dB .....	12
Figure 3 – Hearing aid test positions for BTE (upper) and ITE (lower) .....	14
Figure A.1 – Ratio of 1:2 between field strength and interference level in dB .....	18
Figure A.2 – Example of test arrangement for hearing aid immunity measurements using dipole antenna .....	20
Table 1 – Field strengths of RF test signals to be used to establish immunity for bystander compatible and user compatible hearing aids .....	10

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**ELECTROACOUSTICS –  
HEARING AIDS –****Part 13: Requirements and methods of measurement  
for electromagnetic immunity to mobile digital wireless devices****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60118-13 has been prepared by IEC technical committee 29: Electroacoustics.

This fifth edition cancels and replaces the fourth edition published in 2016 and constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) it introduces a new measurement method and set of EMC requirements for hearing aids immunity to mobile digital wireless devices;
- b) generic EMC requirements for hearing aids are no longer included – should be covered by other standards as appropriate.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
29/1024/FDIS	29/1031/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 60118 series, published under the general title *Electroacoustics – Hearing aids*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

This part of IEC 60118 specifies methods of measurement and requirements for hearing aid immunity to digital wireless devices. Most hearing aids contain digital signal processors and some can contain wireless transceivers.

Experience in connection with the use of hearing aids in recent times has identified digital wireless devices, such as wireless telephones and GSM mobile phones, as potential sources of disturbance for hearing aids. Interference in hearing aids depends on the emitted power from the digital wireless device as well as the immunity of the hearing aid. The performance criteria in this document will not totally ensure hearing aid users' interference- and noise-free use of wireless telephones, but will establish useable conditions in most situations.

In practice, a hearing aid user, when using a wireless telephone, will seek, if possible, to find a position on the ear, which gives minimum or no interference in the hearing aid. Various test methods have been considered for determining the immunity of hearing aids. When a digital wireless device is used close to a hearing aid, there is an RF near-field illumination of the hearing aid. However, validation investigations in preparing this document have shown that it is possible to establish a correlation between the measured far-field immunity level and the immunity level experienced by an actual hearing aid used in conjunction with a digital wireless device. The use of a far-field test has shown high reproducibility and is considered sufficient to verify and express the immunity of hearing aids. Near-field illumination of the hearing aid (i.e. by generating an RF field using a dipole antenna) could however provide valuable information during the design and development of hearing aids.

It is recognized that the new wireless products introduced have to coexist with existing spectra, potential networks and other wireless products (medical as well as non-medical). This revision does not address the issue of coexistence, and the user of this document shall consult applicable entities for guidance.

In this fifth edition of IEC 60118-13, the field strengths and hearing aid positioning during measurements have been updated for consistency with IEEE C63.19 [1]<sup>1</sup> and ANSI C63.19 [2]. The field strength levels used since the first edition of IEC 60118-13 was published in 1997 have demonstrated, through measurements of more than 1 000 hearing aid models (ref. European Hearing Instrument Manufacturers Association – EHIMA), to be sufficiently high to ensure well-functioning hearing aids in everyday use, with only a small expectation of a few complaints regarding interference from digital wireless devices.

Hearing aids where the outputs are non-acoustic, for example bone conduction hearing aids, are not directly included in this document, but this document can be used if precise descriptions of measurement setups for these types of hearing aids are given by the manufacturer.

---

<sup>1</sup> Numbers in square brackets refer to the Bibliography

## ELECTROACOUSTICS – HEARING AIDS –

### Part 13: Requirements and methods of measurement for electromagnetic immunity to mobile digital wireless devices

#### 1 Scope

This part of IEC 60118 covers the relevant EMC phenomena for hearing aids. Hearing aid immunity to high frequency fields originating from digital wireless devices such as mobile phones was identified as one of the most relevant EMC phenomena impacting hearing aids.

#### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60118-0:2015, *Electroacoustics – Hearing aids – Part 0: Measurement of the performance characteristics of hearing aids*

IEC 60318-5, *Electroacoustics – Simulators of human head and ear – Part 5: 2 cm<sup>3</sup> coupler for the measurement of hearing aids and earphones coupled to the ear by means of ear inserts*

IEC 61000-4-3, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-20, *Electromagnetic compatibility (EMC) – Part 4-20: Testing and measurement techniques – Emission and immunity testing in transverse electromagnetic (TEM) waveguides*

#### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

##### 3.1

##### hearing aid

wearable instrument intended to aid a person with impaired hearing, usually consisting of a microphone, amplifier, signal processor and earphone, powered by a low-voltage battery, and possibly also containing an induction pick-up coil and which is fitted using audiometric and prescriptive methods

Note 1 to entry: Hearing aids can be placed on the body (BW), behind the ear (BTE), in the ear (ITE) or in the canal (ITC).