Acoustics and vibration - Laboratory measurement of vibro-acoustic transfer properties of resilient elements - Part 2: Direct method for determination Of Chick of the Ch of the dynamic stiffness of resilient supports for translatory motion



EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN ISO 10846-2:2008 sisaldab Euroopa standardi EN ISO 10846-2:2008 ingliskeelset teksti.

Standard on kinnitatud Eesti Standardikeskuse 25.09.2008 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.

Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on 15.08.2008.

Standard on kättesaadav Eesti standardiorganisatsioonist.

This Estonian standard EVS-EN ISO 10846-2:2008 consists of the English text of the European standard EN ISO 10846-2:2008.

This standard is ratified with the order of Estonian Centre for Standardisation dated 25.09.2008 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.

Date of Availability of the European standard text 15.08.2008.

The standard is available from Estonian standardisation organisation.

ICS 17.140.01

Võtmesõnad: acoustic measurements, acoustics, determination, dynamic stiffness, generalities, mechanical properties, resilient devices, tests, vibration, vibration isolators

Standardite reprodutseerimis- ja levitamisõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

EUROPEAN STANDARD

EN ISO 10846-2

NORME EUROPÉENNE EUROPÄISCHE NORM

August 2008

ICS 17.140.01

Supersedes EN ISO 10846-2:1998

English Version

Acoustics and vibration - Laboratory measurement of vibroacoustic transfer properties of resilient elements - Part 2: Direct method for determination of the dynamic stiffness of resilient supports for translatory motion (ISO 10846-2:2008)

Acoustique et vibrations - Mesurage en laboratoire des propriétés de transfert vibro-acoustique des éléments élastiques - Partie 2: Méthode directe pour la détermination de la raideur dynamique en translation des supports élastiques (ISO 10846-2:2008)

Akustik und Schwingungstechnik - Laborverfahren zur Messung der vibro-akustischen Transfereigenschaften elastischer Elemente - Teil 2: Direktes Verfahren zur Ermittlung der dynamischen Steifigkeit elastischer Stützelemente bei Anregung in translatorischer Richtung (ISO 10846-2:2008)

This European Standard was approved by CEN on 12 April 2008.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

This document (EN ISO 10846-2:2008) has been prepared by Technical Committee ISO/TC 43 "Acoustics" in collaboration with Technical Committee CEN/TC 211 "Acoustics" the secretariat of which is held by DS.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2009, and conflicting national standards shall be withdrawn at the latest by February 2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10846-2:1998.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 10846-2:2008 has been approved by CEN as a EN ISO 10846-2:2008 without any modification.

Page

Contents

Forew	ord	iv
	uction	
introa		
1	Scope	1
2	Normative references	2
3	Terms and definitions	3
4	Principle	6
5	Requirements for apparatus	6
5.1	Normal translations	
5.2	Transverse translations	
5.3	Suppression of unwanted vibrations	11
6	Criteria for adequacy of the test arrangement	12
6.1	Frequency range	12
6.2	Measurement of blocking force	12
6.3	Flanking transmission	
6.4	Unwanted input vibrations	
6.5	Accelerometers	
6.6	Force transducers	
6.7	Summation of signals	
6.8	Analysers	15
7	Test procedures	15
7.1	Installation of the test elements	15
7.2	Selection of force measurement system and output-force distribution plates	15
7.3	Mounting and connection of accelerometers	15
7.4	Mounting and connections of the vibration exciter	16
7.5	Source signal	16
7.6	Measurements	
7.7	Test for linearity	17
8	Evaluation of test results	18
8.1	Calculation of dynamic transfer stiffness	18
8.2	One-third-octave-band values of the frequency-averaged dynamic transfer stiffness	
8.3	Presentation of one-third-octave-band results	
8.4	Presentation of narrow-band data	
9	Information to be recorded	21
10	Test report	22
Annex	A (informative) Static load-deflection curve	
A	B (informative) Measurement uncertainty	
Bibliog	graphy	28

Introduction

Passive resilient elements of various kinds are used to reduce the transmission of vibrations. Examples are automobile engine mounts, resilient supports for buildings, resilient mounts and flexible shaft couplings for shipboard machinery and small isolators in household appliances.

This part of ISO 10846 specifies a direct method for measuring the dynamic transfer stiffness function of linear resilient supports. This includes resilient supports with non-linear static load-deflection characteristics, as long as the elements show an approximate linearity for vibration behaviour for a given static preload. This part of ISO 10846 belongs to a series of International Standards on methods for the laboratory measurement of vibro-acoustic properties of resilient elements, which also includes documents on measurement principles, on an indirect method and on a driving point method. ISO 10846-1 provides guidance for the selection of the appropriate International Standard.

The laboratory conditions described in this part of ISO 10846 include the application of static preload.

The results of the method described in this part of ISO 10846 are useful for resilient supports that are used to prevent low-frequency vibration problems and to attenuate structure-borne sound in the lower part of the audible frequency range. However, for complete characterization of resilient elements that are used to a. Jrsic attenuate low- frequency vibration or shock excursions, additional information is needed, which is not provided by this method.

Acoustics and vibration — Laboratory measurement of vibroacoustic transfer properties of resilient elements —

Part 2:

Direct method for determination of the dynamic stiffness of resilient supports for translatory motion

1 Scope

This part of ISO 10846 specifies a method for determining the dynamic transfer stiffness for translations of resilient supports, under specified preload. The method concerns the laboratory measurement of vibrations on the input side and blocking output forces and is called "the direct method". The method is applicable to test elements with parallel flanges (see Figure 1).

Resilient elements, which are the subject of this part of ISO 10846, are those which are used to reduce

- the transmission of vibration in the lower part of the audible frequency range (typically 20 Hz to 500 Hz) to a structure which may, for example, radiate unwanted fluid-borne sound (airborne, waterborne or others), and
- the transmission of low-frequency vibrations (typically 1 Hz to 80 Hz), which may, for example, act upon human subjects or cause damage to structures of any size when vibration is too severe.
- NOTE 1 In practice, the size of the available test rig(s) can restrict the use of very small or very large resilient supports.


NOTE 2 Samples of continuous supports of strips and mats are included in this method. Whether or not the sample describes the behaviour of the complex system sufficiently is the responsibility of the user of this part of ISO 10846.

Measurements for translations normal and transverse to the flanges are covered in this part of ISO 10846.

The direct method covers the frequency range from 1 Hz up to a frequency $f_{\rm UL}$, which is usually determined by the test rig.

NOTE 3 Because of the large variety of test rigs and test elements, $f_{\rm UL}$ is variable. In this part of ISO 10846, the adequacy of the test rig is not defined for a fixed frequency range, but on the basis of measured data, as described in 6.1 to 6.4.

© ISO 2008 – All rights reserved



NOTE 1 When a resilient support has no parallel flanges, an auxiliary fixture is included as part of the test element to arrange for parallel flanges.

NOTE 2 The arrows indicate the load direction.

Figure 1 — Example of resilient supports with parallel flanges

The data obtained according to the method specified in this part of ISO 10846 can be used for the following:

- product information provided by manufacturers and suppliers;
- information during product development;
- quality control;
- calculation of the transfer of vibration energy through isolators.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 266, Acoustics — Preferred frequencies

ISO 2041:—1), Mechanical vibration, shock and condition monitoring — Vocabulary

ISO 5348, Mechanical vibration and shock — Mechanical mounting of accelerometers

ISO 7626-1, Vibration and shock — Experimental determination of mechanical mobility — Part 1: Basic definitions and transducers

ISO 10846-1, Acoustics and vibration — Laboratory measurement of vibro-acoustic transfer properties of resilient elements — Part 1: Principles and guidelines

ISO 16063-21, Methods for the calibration of vibration and shock transducers — Part 21: Vibration calibration by comparison to a reference transducer

ISO/IEC Guide 98-3²), Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM 1995)

.

¹⁾ To be published. (Revision of ISO 2041:1990)

²⁾ ISO/IEC Guide 98-3 will be published as a re-issue of the *Guide to expression of uncertainty in measurement* (GUM), 1995.