

Sterilization of medical devices - Low temperature steam and formaldehyde - Requirements for development, validation and routine control of a sterilization process for medical devices

Sterilization of medical devices - Low temperature steam and formaldehyde - Requirements for development, validation and routine control of a sterilization process for medical devices

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN 15424:2007 sisaldab Euroopa standardi EN 15424:2007 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 31.05.2007 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN 15424:2007 consists of the English text of the European standard EN 15424:2007.</p> <p>This document is endorsed on 31.05.2007 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

<p>Käsitlusala: This European Standard specifies requirements for the development, validation and routine control of a Low Temperature Steam and Formaldehyde (LTSF) sterilization process for medical devices.</p>	<p>Scope: This European Standard specifies requirements for the development, validation and routine control of a Low Temperature Steam and Formaldehyde (LTSF) sterilization process for medical devices.</p>
--	--

ICS 11.080.01

Võtmesõnad:

English Version

Sterilization of medical devices - Low temperature steam and formaldehyde - Requirements for development, validation and routine control of a sterilization process for medical devices

Stérilisation des dispositifs médicaux - Développement, validation et contrôle de routine des procédés de stérilisation - Vapeur d'eau à basse température et au formaldéhyde

Sterilisation von Medizinprodukten - Niedertemperatur-Dampf-Formaldehyd - Anforderungen an die Entwicklung, Validierung und Routineüberwachung von Sterilisationsverfahren für Medizinprodukte

This European Standard was approved by CEN on 17 February 2007.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	4
Introduction.....	5
1 Scope.....	7
1.1 Inclusions.....	7
1.2 Exclusions.....	7
2 Normative references.....	8
3 Terms and definitions.....	8
4 Quality Management System Elements.....	14
4.1 Documentation.....	14
4.2 Management responsibility.....	14
4.3 Product realization.....	14
4.4 Control of non-conforming product.....	15
5 Sterilizing agent characterization.....	15
5.1 General.....	15
5.2 Sterilizing agent.....	15
5.3 Microbicidal effectiveness.....	15
5.4 Material effects.....	15
5.5 Environmental considerations.....	15
6 Process and equipment characterization.....	16
6.1 General.....	16
6.2 Process.....	16
6.3 Equipment.....	16
7 Product definition.....	17
8 Process definition.....	18
9 Validation.....	19
9.1 General.....	19
9.2 Installation qualification.....	19
9.2.1 General.....	19
9.2.2 Installation.....	19
9.2.3 Equipment.....	20
9.3 Operational qualification.....	20
9.4 Performance qualification.....	20
9.4.1 General.....	20
9.4.2 Performance qualification – physical.....	21
9.4.3 Performance qualification – microbiological.....	21
9.4.4 Performance qualification – desorption and drying.....	22
9.5 Review and approval of validation.....	22
10 Routine monitoring and control.....	23
10.1 General.....	23
10.2 Biological indicators.....	23
10.3 Chemical indicators.....	23
10.4 Records.....	23
11 Product release from sterilization.....	24
12 Maintaining process effectiveness.....	24
12.1 General.....	24
12.2 Maintenance of equipment.....	24
12.3 Requalification.....	24
12.4 Assessment of change.....	25
Annex A (normative) Process definition based on inactivation of reference microorganisms and knowledge of bioburden on product items to be sterilized.....	26

A.1	General	26
A.2	Procedure	26
Annex B (normative) Process definition based on inactivation of reference microorganisms		27
B.1	General	27
B.1.1	Overkill approach	27
B.1.2	Penetration characteristics into medical devices	27
B.2	Test procedure	28
B.2.1	General	28
B.2.2	Biological indicators	28
B.2.3	Test systems	28
B.2.4	Load configuration	28
B.2.5	Testing	28
Annex C (informative) Guidance on application of this European Standard		30
C.1	Scope	30
C.2	Normative references	30
C.3	Terms and definitions	30
C.4	Quality management system elements	30
C.5	Sterilizing agent characterization	30
C.5.1	Neutralization	30
C.5.2	Studies of microbial inactivation	31
C.6	Process and equipment characterization	31
C.7	Product definition	31
C.7.1	General	31
C.7.2	Design considerations for medical devices intended for sterilization	32
C.7.3	Packaging considerations	32
C.8	Process definition	32
C.8.1	General	32
C.8.2	Influence on product and packaging	33
C.8.3	Determination of process effectiveness	33
C.9	Validation	34
C.9.1	General	34
C.9.2	Installation qualification (IQ)	34
C.9.3	Operational qualification (OQ)	35
C.9.4	Performance qualification (PQ)	36
C.9.5	Documentation and approval of validation	37
C.10	Routine monitoring and control	37
C.11	Product release from sterilization	38
C.11.1	General	38
C.11.2	Product release using biological indicators	38
C.11.3	Parametric release	38
C.12	Maintaining process effectiveness	38
C.12.1	Change control	38
C.12.2	Requalification	39
Annex D (informative) Environmental aspects regarding development, validation and routine control of Low Temperature Steam and Formaldehyde processes		40
D.1	General	40
D.2	Formaldehyde (brief description)	40
D.3	Environmental impact of formaldehyde	40
D.4	Other environmental burdens	41
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC Medical devices		44
Bibliography		45

Foreword

This document (EN 15424:2007) has been prepared by Technical Committee CEN/TC 204 "Sterilization of medical devices", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2007, and conflicting national standards shall be withdrawn at the latest by October 2007.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Introduction

A sterile medical device is one which is free of viable microorganisms. European Standards, which specify requirements for validation and routine control of a sterilization process require, when it is necessary to supply a sterile medical device, that adventitious microbiological contamination of a medical device prior to sterilization be minimized. Even so, medical devices produced under standard manufacturing conditions in accordance with the requirements for quality management systems (see EN ISO 13485) or which have been subjected to a cleaning process as part of their reprocessing in a health care establishment may, prior to sterilization, have microorganisms on them, albeit in low numbers. Such products are non-sterile. The purpose of sterilization is to inactivate the microbiological contaminants and thereby transform the non-sterile products into sterile ones.

The kinetics of inactivation of a pure culture of microorganisms by physical and/or chemical agents used to sterilize medical devices can generally best be described by an exponential relationship between the numbers of microorganisms surviving and the extent of treatment with the sterilizing agent; inevitably this means that there is always a finite probability that a microorganism may survive regardless of the extent of treatment applied. For a given treatment, the probability of survival is determined by the number and resistance of microorganisms and by the environment in which the organisms exist during treatment. It follows that the sterility of any one product in a population subjected to sterilization cannot be guaranteed and the sterility of a processed population is defined in terms of the probability of there being a viable microorganism present on a product item.

This standard describes requirements which will enable the demonstration that a low temperature steam and formaldehyde sterilization process intended to sterilize medical devices has appropriate microbicidal activity, and that this activity is both reliable and reproducible, such that the relationship for the inactivation of microorganisms can be extrapolated with reasonable confidence to low levels of probability of there being a viable microorganism present on a product after sterilization. This standard does not specify the maximal value to be taken by this probability; specification of this probability is given in EN 556-1.

Requirements of the quality management system for medical device design/development, production, installation and servicing are given in EN ISO 13485. The standard for quality management systems recognizes that, for certain processes used in manufacturing or reprocessing, the effectiveness cannot be fully verified by subsequent inspection and testing of the product. Sterilization is an example of such a process. For this reason, sterilization processes are validated for use, the performance of the sterilization process monitored routinely and the equipment maintained.

Exposure to a properly validated, accurately controlled sterilization process is not the only factor associated with the provision of reliable assurance that the product is sterile and, in this regard, suitable for its intended use. Attention is therefore given to a number of factors including:

- a) the microbiological status of incoming raw materials and/or components,
- b) the validation and routine control of any cleaning and disinfection procedures used on the product,
- c) the control of the environment in which the product is manufactured, assembled and packaged,
- d) the control of equipment and processes,
- e) the control of personnel and their hygiene,
- f) the manner and materials in which the product is packaged, and,
- g) the conditions under which the product is transported and stored.

The type of contamination on a product to be sterilized varies and this impacts upon the effectiveness of a sterilization process. Products that have been used in a health care setting and are being presented for re-sterilization in accordance with the manufacturer's instructions (see EN ISO 17664) should be regarded as a special case. There is the potential for such products to possess a wide range of contaminating microorganisms and residual inorganic and/or organic contamination in spite of the application of a cleaning process. Hence, particular attention has to be given to the validation and control of the cleaning and disinfection processes used during reprocessing.

The requirements are the normative parts of this standard with which compliance is claimed. The guidance given in the informative annexes is not normative and is not provided as a checklist for auditors. The guidance provides explanations as well as methods that are accepted as being suitable means for complying with the requirements. Approaches other than those given in the guidance may be used, if they are effective in achieving compliance with the requirements of this European Standard.

The development, validation and routine control of a sterilization process comprise a number of discrete but interrelated activities, for example calibration, maintenance, product definition, process definition, installation qualification, operational qualification, and performance qualification. While the activities required by this standard have been grouped together and are presented in a particular order, this European Standard does not require that the activities be performed in the order that they are presented. The activities required are not necessarily sequential, as the programs of development and validation may be iterative. It is possible that performing these different activities will involve a number of separate individuals and/or organizations, each of whom undertake one or more of these activities. This European Standard does not specify the particular individuals or organizations to carry out the activities.

Activities required by this standard might also give rise to an environmental burden that should be considered and minimized, e.g. by utilizing flexibility in planning. Environmental aspects are addressed in Annex D of this standard.

1 Scope

1.1 Inclusions

1.1.1 This European Standard specifies requirements for the development, validation and routine control of a Low Temperature Steam and Formaldehyde (LTSF) sterilization process for medical devices.

NOTE Although the scope of this standard is limited to medical devices, it specifies requirements and provides guidance that may be applicable to other products and equipment.

1.1.2 This European Standard is intended to be applied by process developers, manufacturers of sterilization equipment, manufacturers of medical devices to be sterilized and the organizations with responsibility for sterilizing medical devices. (See EN ISO 14937:2000, Table E.1)

1.1.3 This European Standard covers sterilization processes which use a mixture of low temperature steam and formaldehyde as sterilant, and which are working below ambient pressure only.

1.2 Exclusions

1.2.1 Sterilization processes validated and controlled in accordance with the requirements of this standard should not be assumed to be effective in inactivating the causative agents of spongiform encephalopathies such as scrapie, bovine spongiform encephalopathy and Creutzfeld-Jakob disease. Specific recommendations have been produced in particular countries for the processing of materials potentially contaminated with these agents.

1.2.2 This standard does not specify requirements for designating a medical device as "STERILE". Such requirements are given in EN 556-1.

1.2.3 This standard does not specify a quality management system for the control of all stages of production of medical devices.

NOTE It is not a requirement of this standard to have a complete quality management system during manufacture or reprocessing, but those elements of such a system that are required are normatively referenced at appropriate places in the text. Attention is drawn to the standards for quality management systems (see EN ISO 13485) that control all stages of production or reprocessing of medical devices including the sterilization process. Further guidance is given in E.2 of EN ISO 14937:2000.

1.2.4 This standard does not specify requirements for occupational safety associated with the design and operation of LTSF sterilization facilities.

NOTE 1 Safety requirements for sterilizers are specified in EN 61010-2-040.

NOTE 2 Attention is also drawn to the existence in some countries of regulations stipulating safety requirements.

1.2.5 This European Standard does not cover analytical methods for determining levels or residues of formaldehyde and/or its reaction products.

NOTE 1 Attention is drawn to EN 14180.

NOTE 2 Attention is drawn to the possible existence in some countries of statutory regulation specifying limits for the level of formaldehyde residues on medical devices and products.

1.2.6 This European Standard does not cover preparatory measures that may be necessary before sterilization such as cleaning, disinfection and packing.

NOTE For re-sterilizable medical devices, the manufacturer(s) of these devices should supply information on the preparatory measures (see EN ISO 17664).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 14180:2003, *Sterilizers for medical purposes – Low temperature steam and formaldehyde sterilizers – Requirements and testing*

EN ISO 11138-1, *Sterilization of health care products - Biological indicators - Part 1: General requirements (ISO 11138-1:2006)*

EN ISO 11138-5:2006, *Sterilization of health care products - Biological indicators - Part 5: Biological indicators for low-temperature steam and formaldehyde sterilization processes (ISO 11138-5:2006)*

EN ISO 11140, *Sterilization of health care products - Chemical indicators (Parts as appropriate)*

EN ISO 11737-1, *Sterilization of medical devices - Microbiological methods - Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2006)*

EN ISO 11737-2, *Sterilization of medical devices - Microbiological methods - Part 2: Tests of sterility performed in the validation of a sterilization process (ISO 11737-2:1998)*

EN ISO 13485, *Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2003)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

adjustment

correction of a measurement device or system to indicate the value as established by calibration

3.2

aeration

part or parts of the sterilization process in which defined conditions are used such that formaldehyde and its reaction products are desorbed from the medical device, and which can be performed within the sterilizer, within a separate room or chamber, or by a combination of the two

[3.3 of EN 14180:2003]

3.3

air removal

removal of air from the sterilizer chamber and sterilization load to facilitate sterilant penetration

[3.3 of EN 14180:2003]

3.4

bioburden

population of viable microorganisms on or in product and/or sterile barrier system

[2.2 of ISO/TS 11139:2006]