

**Health informatics - Device interoperability - Part
20701: Point-of-care medical device communication -
Service oriented medical device exchange architecture
and protocol binding (ISO/IEEE 11073-20701:2020)**

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

See Eesti standard EVS-EN ISO 11073-20701:2020 sisaldab Euroopa standardi EN ISO 11073-20701:2020 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 11073-20701:2020 consists of the English text of the European standard EN ISO 11073-20701:2020.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 01.04.2020.	Date of Availability of the European standard is 01.04.2020.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 35.240.80

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 11073-20701

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2020

ICS

English Version

Health informatics - Device interoperability - Part 20701:
Point-of-care medical device communication - Service
oriented medical device exchange architecture and
protocol binding (ISO/IEEE 11073-20701:2020)

Informatique de santé - Interopérabilité des dispositifs
- Partie 20701: Communication entre dispositifs
médicaux sur le site des soins - Architecture d'échange
orientée services entre dispositifs médicaux et liaison
par protocole (ISO/IEEE 11073-20701:2020)

Medizinische Informatik - Kommunikation
patientennaher medizinischer Geräte - Teil 20701:
Service-Orientierte Architektur und Protokoll für
Medizingeräte-Kommunikation (ISO/IEEE 11073-
20701:2020)

This European Standard was approved by CEN on 18 February 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 11073-20701:2020) has been prepared by Technical Committee ISO/TC 215 "Health informatics" in collaboration with Technical Committee CEN/TC 251 "Health informatics" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2020, and conflicting national standards shall be withdrawn at the latest by October 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO/IEEE 11073-20701:2020 has been approved by CEN as EN ISO 11073-20701:2020 without any modification.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted (see www.iso.org/directives).

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO/IEEE 11073-20701 was prepared by the IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society (as IEEE Std 11073-20701-2018) and drafted in accordance with its editorial rules. It was adopted, under the "fast-track procedure" defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE, by Technical Committee ISO/TC 215, *Health informatics*.

A list of all parts in the ISO/IEEE 11073 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This introduction is not part of IEEE Std 11073-20701-2018, Health Informatics—Point-of-care medical device communication—Part 20701: Service-Oriented Medical Device Exchange Architecture and Protocol Binding.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. They provide automatic and detailed electronic data capture of patient vital signs information and device operational data. The primary goals are to:

- Provide real-time plug-and-play interoperability for medical devices
- Facilitate the efficient exchange of vital signs and medical device data, acquired at the Point-of-Care (PoC), in all health care environments

“Real-time” means that data from multiple devices can be retrieved, time correlated, and displayed or processed in fractions of a second. “Plug-and-play” means that all the clinician has to do is to make the connection—the Participants automatically detect, configure, and communicate without any other human interaction.

“Efficient exchange of medical device data” means that information that is captured at the PoC (e.g., patient vital signs data) can be received, parsed, and interpreted by many different types of applications without unnecessary loss of information. The standards are especially targeted at acute, surgical, and continuing care devices, such as patient monitors, ventilators, infusion pumps, ECG devices, endoscopic camera system, insufflators, endoscopic light sources, dissectors, etc. They comprise a family of standards that can be bound to one another to provide optimized connectivity for devices at the Point-of-Care.

Within the context of the ISO/IEEE 11073 family of standards for PoC medical device communication, this standard defines an architecture for service-oriented distributed PoC medical devices and medical IT systems. It defines a binding of the Participant, Discovery, and Communication Model defined in IEEE Std 11073-10207 to the profile for transport over Web Services defined in IEEE Std 11073-20702. Moreover, a binding to Network Time Protocol (NTP) and Differentiated Services (DiffServ) is defined to satisfy time synchronization and transport Quality of Service requirements.

Contents

1. Overview	10
1.1 Scope	10
1.2 Purpose	10
2. Normative references.....	10
3. Definitions	11
4. Notational conventions.....	15
4.1 XML schema namespaces	15
5. Introduction	16
6. Service-oriented medical device exchange architecture	17
7. Service-oriented device connectivity (SDC) participant model binding	18
7.1 Coded values	18
7.2 Remote-control capabilities	18
7.3 Retrievability of containment tree entries.....	20
7.4 Dynamic containment tree changes	21
7.5 MDIB versioning.....	22
7.6 Types	22
8. Communication model binding	23
8.1 Service	23
8.2 Message	25
9. Discovery binding	27
9.1 Discovery mechanism.....	27
9.2 Complex device component based discovery	27
9.3 SDC PARTICIPANT KEY PURPOSE based discovery	28
9.4 Context-based discovery.....	29
9.5 Announcing absence.....	31
10. Non-functional quality attributes.....	31
10.1 Cybersecurity.....	31
10.2 Patient safety.....	33
10.3 Clinical effectiveness.....	34
11. Conformance	36
11.1 General format.....	36
11.2 ICS tables.....	37
Annex A (normative) Constants	39
Annex B (normative) SDC service provider WSDL service descriptions.....	40
B.1 Get Service.....	40
B.2 Set Service	40
B.3 Description Event Service.....	40
B.4 State Event Service	41
B.5 Context Service.....	41
B.6 Waveform Service	41

B.7 Containment Tree Service.....	42
B.8 Archive Service.....	42
B.9 Localization Service.....	42
Annex C (informative) Bibliography.....	43

This document is a preview generated by EVS

Health informatics—Point-of-care medical device communication

Part 20701: Service-Oriented Medical Device Exchange Architecture and Protocol Binding

1. Overview

1.1 Scope

The scope of this standard is a service-oriented medical device architecture and communication protocol specification for distributed system of Point-of-Care (PoC) medical devices and medical IT systems that need to exchange data or safely control networked PoC medical devices. It identifies the functional components, their communication relationships as well as the binding of the components and communication relationships to protocol specifications.

1.2 Purpose

This standard defines an architecture for service-oriented distributed PoC medical devices and medical IT systems. It describes a binding of the Participant and Communication model as defined in IEEE Std 11073-10207™ to Medical Devices Communication Profile for Web Services (MDPWS) as defined in IEEE Std 11073-20702™ for transport over Web Services.¹ Moreover, a binding to the Network Time Protocol (NTP) and Differentiated Services (DiffServ) is specified for time synchronization and transport Quality of Service requirements.

2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used, so each referenced document is cited in text and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 11073-10207-2017, IEEE Health informatics—Point-of-care medical device communication—Part 10207: Domain Information and Service Model for Service-Oriented Point-of-Care Medical Device Communication.^{2, 3}

¹ Information on references can be found in Clause 2.

² IEEE publications are available from the Institute of Electrical and Electronics Engineers (<http://standards.ieee.org>).

IEEE Std 11073-20701-2018
 Health informatics—Point-of-care medical device communication
 Part 20701: Service-Oriented Medical Device Exchange Architecture and Protocol Binding

IEEE Std 11073-20702-2016, IEEE Health informatics—Point-of-care medical device communication—Part 20702: Medical Devices Communication Profile for Web Services.

IETF RFC 1305, Network Time Protocol (Version 3) Specification, Implementation and Analysis, D. Mills, March 1992. Available at <https://tools.ietf.org/html/rfc1305>.

IETF RFC 1769, Simple Network Time Protocol (SNTP), D. Mills, March 1995. Available at <https://tools.ietf.org/html/rfc1769>.

IETF RFC 2474, Definition of the Differentiated Services Field (DS Field) in the IPv4 and IPv6 Headers, K. Nichols et al., December 1998. Available at <https://tools.ietf.org/html/rfc2474>.

IETF RFC 2475, An Architecture for Differentiated Services, S. Blake et al., December 1998. Available at <https://tools.ietf.org/html/rfc2475>.

IETF RFC 4330, Simple Network Time Protocol (SNTP) Version 4 for IPv4, IPv6 and OSI, D. Mills, January 2006. Available at <https://tools.ietf.org/html/rfc4330>.

IETF RFC 5227, IPv4 Address Conflict Detection, S. Cheshire, July 2008. Available at <https://tools.ietf.org/html/rfc5227>.

IETF RFC 5246, The Transport Layer Security (TLS) Protocol Version 1.2, T. Dierks et al., August 2008. Available at <https://tools.ietf.org/html/rfc5246.txt>.

IETF RFC 5905, Network Time Protocol Version 4: Protocol and Algorithms Specification, D. Mills et al., June 2010. Available at <https://tools.ietf.org/html/rfc5905>.

ISO/IEEE Std 11073-10101:2004, Health informatics—Point-of-care medical device communication—Part 10101: Nomenclature.⁴

3. Definitions

For the purposes of this document, the following terms and definitions apply. The *IEEE Standards Dictionary Online* should be consulted for terms not defined in this clause.⁵

ALERT: Synonym for the combination of patient-related physiological alarms, technical alarms and equipment user advisory signals.

NOTE 1—Patient-related physiological alarms as well as technical alarms are ALARM SIGNALS indicating the presence of ALARM CONDITIONS. An ALERT CONDITION is an ALARM CONDITION if the priority is LOW PRIORITY, MEDIUM PRIORITY, or HIGH PRIORITY and where the origin is a Physiological or Technical. See IEC 60601-1-8:2006+AMD1:2012 [B3].^{6, 7}

NOTE 2—Equipment user advisory signals are INFORMATION SIGNALS indicating the presence of conditions that are not ALARM CONDITIONS. An ALERT CONDITION is not an ALARM CONDITION if the priority is neither LOW PRIORITY nor MEDIUM PRIORITY nor HIGH PRIORITY. See IEC 60601-1-8:2006+AMD1:2012 [B3].

³ The IEEE standards or products referred to in Clause 2 are trademarks owned by the Institute of Electrical and Electronics Engineers, Incorporated.

⁴ ISO/IEEE publications are available from the International Organization for Standardization (<http://www.iso.org/>) and the Institute of Electrical and Electronics Engineers (<http://standards.ieee.org/>).

⁵ *IEEE Standards Dictionary Online* is available at: <http://dictionary.ieee.org>.

⁶ The numbers in brackets correspond to those of the bibliography in Annex C.

⁷ Notes in text, tables, and figures of a standard are given for information only and do not contain requirements needed to implement this standard.