
Dentistry — Colour tabs for intraoral tooth colour determination

*Médecine bucco-dentaire — Plaquettes de teinte pour la
détermination de la teinte dentaire intra-buccale*



This document is a preview generated by ERS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	2
4.1 Materials and processing.....	2
4.2 Appearance.....	2
4.3 Uniformity.....	3
4.4 Cleaning and disinfection.....	3
4.5 Colour stability.....	3
5 Test method for colour measurement	3
5.1 Test devices and test equipment.....	3
5.1.1 Spectrophotometers.....	3
5.1.2 Colour tab.....	4
5.1.3 Countershape.....	4
5.2 Procedure.....	4
5.2.1 Positioning of the colour tab.....	4
5.2.2 Measurement parameters to be set.....	4
5.2.3 Number of measurements.....	4
5.3 Expression of results.....	4
6 Information and instructions	5
6.1 Information.....	5
6.2 Instructions.....	5
7 Marking	5
Bibliography	6

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthodontic materials*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

In dentistry, colour matching, colour communication among dentist and dental technician (indirect restorations) and colour reproduction and verification are essential elements for a successful aesthetic restoration. Other important appearance attributes include translucency/opacity, gloss, opalescence and fluorescence.

Questions regarding the principles of chromatics and the individual colour perception have been controversially discussed for years between those involved. Today, various systems are available for the purpose of objectivizing tooth colour perception. The most common method is the colour comparison using reference colours being supplied in the form of shade guides consisting of tooth shaped colour tabs made of ceramic, for a dental colour system. The objective of this document is to specify test methods for such colour tabs. This document does not cover colour samples, which are produced by each manufacturer to demonstrate colour effects usually of all components of a given individually designed assortment. These colour samples are not used to determine the colour *in vivo* and are not provided in a tooth-shaped form.

Dentistry — Colour tabs for intraoral tooth colour determination

1 Scope

This document specifies requirements for tooth-like colour representations made of ceramic materials used to determine the tooth colour in the patient's mouth or to check the colour of dental prosthesis, which are referred to as colour tabs in this document.

The colour coordinates of colour tabs are left to the manufacturers' discretion.

Resources for visualizing the colours of ceramic and other masses, e.g. mass shade guides and colour patterns for certain ceramic and other masses, do not fall into the scope of this document. They can be manufactured from any materials and serve solely to illustrate the colour effect; they do not serve to determine colour inside the mouth.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 8601-1, *Date and time — Representations for information interchange — Part 1: Basic rules*

ISO 8601-2, *Date and time — Representations for information interchange — Part 2: Extensions*

ISO 11664-2, *Colorimetry — Part 2: CIE standard illuminants*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO/TR 28642, *Dentistry — Guidance on colour measurement*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 colour difference

single number or metric expressing the distance from complete match in colour or shade

Note 1 to entry: A colour distance metric defined by the International Commission on Illumination (CIE) is called delta E (ΔE).

Note 2 to entry: Two formulae for calculating ΔE are recommended in this document: CIELAB or CIE 76 (denoted ΔE^*_{ab}) and CIEDE2000 (denoted ΔE_{00}). CIEDE2000 formula is currently the most advanced one and is recommended by CIE as it exhibits better correlation with visual findings compared with CIELAB.