

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Medical electrical equipment –  
Part 1-9: General requirements for basic safety and essential performance –  
Collateral Standard: Requirements for environmentally conscious design**

**Appareils électromédicaux –  
Partie 1-9: Exigences générales pour la sécurité de base et les performances  
essentiels – Norme collatérale: Exigences pour une conception éco-  
responsable**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 electrotechnical terminology entries in English and French extracted from the Terms and definitions clause of IEC publications issued between 2002 and 2015. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et définitions des publications IEC parues entre 2002 et 2015. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.



IEC 60601-1-9

Edition 1.2 2020-07  
CONSOLIDATED VERSION

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Medical electrical equipment –  
Part 1-9: General requirements for basic safety and essential performance –  
Collateral Standard: Requirements for environmentally conscious design**

**Appareils électromédicaux –  
Partie 1-9: Exigences générales pour la sécurité de base et les performances  
essentielle – Norme collatérale: Exigences pour une conception éco-  
responsable**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.01; 13.020.01

ISBN 978-2-8322-8707-1

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**



# REDLINE VERSION

## VERSION REDLINE



**Medical electrical equipment –  
Part 1-9: General requirements for basic safety and essential performance –  
Collateral Standard: Requirements for environmentally conscious design**

**Appareils électromédicaux –  
Partie 1-9: Exigences générales pour la sécurité de base et les performances  
essentielle – Norme collatérale: Exigences pour une conception éco-  
responsable**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
INTRODUCTION to Amendment 1 .....	7
INTRODUCTION to Amendment 2 .....	7
1 Scope, object and related standards.....	8
1.1 * Scope .....	8
1.2 Object .....	8
1.3 Related standards .....	8
2 Normative references .....	9
3 Terms and definitions .....	9
4 Protection of the ENVIRONMENT.....	11
4.1 * Identification of ENVIRONMENTAL ASPECTS.....	11
4.2 * Determination of significant ENVIRONMENTAL ASPECTS.....	11
4.3 * Information from the SUPPLY CHAIN .....	11
4.4 * Reduction of adverse ENVIRONMENTAL IMPACTS .....	12
4.5 Environmental information.....	12
Annex A (informative) General guidance and rationale.....	14
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT .....	29
Bibliography.....	30
Index of defined terms used in this collateral standard.....	31
Table A.1 – Example product LIFE-CYCLE stages .....	15
Table A.2 – Examples of ENVIRONMENTAL IMPACTS and their cause .....	21
Table A.3 – ENVIRONMENTAL ASPECTS and typical ENVIRONMENTAL IMPACTS .....	24
Table B.1 – ACCOMPANYING DOCUMENTS, General .....	29
Table B.2 – Other information .....	29

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

### MEDICAL ELECTRICAL EQUIPMENT –

#### Part 1-9: General requirements for basic safety and essential performance –

#### Collateral Standard: Requirements for environmentally conscious design

### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

**This consolidated version of the official IEC Standard and its amendments has been prepared for user convenience.**

**IEC 60601-1-9 edition 1.2 contains the first edition (2007-07) [documents 62A/571/FDIS and 62A/575/RVD], its amendment 1 (2013-06) [documents 62A/874/FDIS and 62A/881/RVD] and its amendment 2 (2020-07) [documents 62A/1393/FDIS and 62A/1408/RVD].**

**In this Redline version, a vertical line in the margin shows where the technical content is modified by amendments 1 and 2. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.**

International standard IEC 60601-1-9 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

The first edition of this publication constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the four numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 4 includes subclauses 4.1, 4.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 4.1, 4.5 and 4.5.1 are all subclauses of Clause 4).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (\*).

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.



The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.**

## INTRODUCTION

The objective of this collateral standard is to improve the ENVIRONMENTAL IMPACT for the entire range of MEDICAL ELECTRICAL EQUIPMENT, taking into account all stages of the product LIFE CYCLE:

- product specification;
- design;
- manufacturing;
- sales, logistics, installation;
- use;
- END OF LIFE management.

This means protecting the ENVIRONMENT and human health from HAZARDOUS SUBSTANCES, conserving raw materials and energy, minimizing the generation of WASTE, as well as minimizing the adverse ENVIRONMENTAL IMPACTS associated with WASTE. The criteria needed to reach this goal must be integrated into all stages of the MEDICAL ELECTRICAL EQUIPMENT LIFE CYCLE from the specification stage to END OF LIFE management.

The ENVIRONMENTAL IMPACTS of ME EQUIPMENT through all LIFE-CYCLE stages are determined from the MEDICAL ELECTRICAL EQUIPMENT'S ENVIRONMENTAL ASPECTS defined during the identification of need, product planning, and design stages (see Table A.1). Consideration of ENVIRONMENTAL ASPECTS as early as possible in these stages can produce numerous benefits that might include lower costs, stimulation of innovation and creativity, and increased knowledge about the product. It can also provide new business opportunities, and improved product quality as well as reduction of adverse ENVIRONMENTAL IMPACTS. The assessment of the ENVIRONMENTAL ASPECTS and IMPACTS of MEDICAL ELECTRICAL EQUIPMENT is a developing science and it is anticipated that this collateral standard will require periodic updating as the science develops.

The requirements given in this collateral standard do not replace national or international laws and regulations.

Environmental protection is one element of the overall RISK MANAGEMENT PROCESS as required by the general standard.

The acceptability of MEDICAL ELECTRICAL EQUIPMENT'S ENVIRONMENTAL IMPACTS are balanced against other factors, such as the product's intended function, performance, safety, cost, marketability, quality, legal and regulatory requirements. This balance can differ depending on the intended function of the MEDICAL ELECTRICAL EQUIPMENT. For example, a solution appropriate for life-saving or life-supporting MEDICAL ELECTRICAL EQUIPMENT might not be appropriate for a device intended to correct a minor ailment. A MANUFACTURER of MEDICAL ELECTRICAL EQUIPMENT might have to justify, as a result of RISK MANAGEMENT, that a medical benefit outweighs the associated adverse ENVIRONMENTAL IMPACTS.

## INTRODUCTION to Amendment 1

The first edition of IEC 60601-1-9 was published in 2007. This amendment is intended to update the references to IEC 60601-1:2005 to include Amendment 1:2012 and to make a few minor editorial updates.

## INTRODUCTION to Amendment 2

The first edition of IEC 60601-1-9 was published in 2007 and amended in 2013. Since the publication of IEC 60601-1-9:2007+A1:2013, the IEC Subcommittee (SC) 62A Secretariat has been collecting issues from a variety of sources including comments from National Committees. At the November 2015 meeting of IEC/SC 62A in Kobe, Japan, the subcommittee initiated a process to identify high-priority issues that need to be considered in an amendment and should not wait until the second edition of IEC 60601-1-9, which is presently targeted for publication sometime after 2024.

As directed in item 1 of Kobe Resolution 1, the IEC/SC 62A Chairman Advisory Group (CAG) considered the 7 issues collected by the SC/62A Secretariat for IEC 60601-1-9:2007 and determined that none met the selection criteria stated in Kobe Resolution 1.

However, an amendment is needed to update the reference to IEC 60601-1:2005+A1:2012 +A2:2020. In London in 2018, SC 62A approved the development of an administrative amendment to IEC 60601-1-9:2007+A1:2013.

Because this is an amendment to IEC 60601-1-9:2007, the style in force at the time of publication of IEC 60601-1-9 has been applied to this amendment. The specified in ISO/IEC Directives Part 2:2018 has only been applied when implementing the new style guidance would not result in additional editorial changes.

Users of this document should note that when constructing the dated references to specific elements in a standard, such as definitions, amendments are only referenced if they modified the text being cited. For example, if a reference is made to a definition that has not been modified by an amendment, then the reference to the amendment is not included in the dated reference.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 1-9: General requirements for basic safety and essential performance – Collateral Standard: Requirements for environmentally conscious design

## 1 Scope, object and related standards

### 1.1 \* Scope

This International Standard applies to the reduction of adverse ENVIRONMENTAL IMPACTS of MEDICAL ELECTRICAL EQUIPMENT, hereafter referred to as ME EQUIPMENT.

MEDICAL ELECTRICAL SYSTEMS are excluded from the scope of this collateral standard.

### 1.2 Object

The object of this collateral standard is to specify general requirements, in addition to those of the general standard, for the reduction of the adverse ENVIRONMENTAL IMPACT of ME EQUIPMENT, and to serve as the basis for particular standards.

### 1.3 Related standards

#### 1.3.1 IEC 60601-1

For ME EQUIPMENT, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone ~~(IEC 60601-1:2005+A1:2012)~~, including any amendments;
- "this collateral standard" designates IEC 60601-1-9 alone ~~(IEC 60601-1-9:2007+A1:2013)~~, including any amendments;
- "this standard" designates the combination of the general standard and this collateral standard.

#### 1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

#### 1.3.3 Environmental standards

This standard takes into account the ISO 14000 series of environmental standards with particular emphasis on ISO 14062 [8]<sup>1)</sup>.

<sup>1)</sup> Figures in square brackets refer to the Bibliography.

## 2 Normative references

The following ~~referenced~~ documents, in whole or in part, are normatively referenced in this document and are indispensable for its application ~~of this document~~. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*  
Amendment 1:2012  
Amendment 2:2020

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005+A1:2012+A2:2020 and the following definitions apply.

NOTE An index of defined terms is found beginning on page 30.

### 3.1

#### DESIGN AND DEVELOPMENT

set of PROCESSES that transforms requirements into specified characteristics or into the specification of a product, PROCESS or system

NOTE 1 The terms “design” and “development” are sometimes used synonymously and sometimes used to define different stages of the overall PROCESS of turning an idea into a product.

NOTE 2 Product development is the PROCESS of taking a product idea from planning to market launch and post-market review of the product, in which business strategies, marketing considerations, research methods and design aspects are used to take a product to a point of practical use. It includes improvements or modifications to existing products or PROCESSES

NOTE 3 The integration of ENVIRONMENTAL ASPECTS into product DESIGN AND DEVELOPMENT can also be termed design for the ENVIRONMENT (DFE), eco-design, the environmental part of product stewardship, etc.

[ISO/TR 14062:2002, definition 3.3]

### 3.2

#### END OF LIFE

#### EOL

state of a ME EQUIPMENT when it is finally removed from its INTENDED USE

NOTE Adapted from IEC Guide 109:2003, Definition 3.1.

### 3.3

#### ENVIRONMENT

surroundings in which an ORGANIZATION operates, including air, water, land, natural resources, flora, fauna, humans and their interrelation

NOTE Surroundings in this context extend from within an ORGANIZATION to the global system.

[ISO 14001:2004, definition 3.5]

### 3.4

#### \* ENVIRONMENTAL ASPECT

element of an ORGANIZATION'S activities, products or services that can interact with the ENVIRONMENT

NOTE A significant ENVIRONMENTAL ASPECT has or can have a significant ENVIRONMENTAL IMPACT.

[ISO 14001:2004, definition 3.6]