IEC 60601-1-11

Edition 2.1 2020-07
CONSOLIDATED VERSION

INTERNATIONAL STANDARD

NORME INTERNATIONALE

colour inside

This document is a preview generated by EVS

Medical electrical equipment –
Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

Appareils électromédicaux –
Partie 1-11: Exigences générales pour la sécurité de base et les performances essentielles – Norme Collatérale: Exigences pour les appareils électromédicaux et les systèmes électromédicaux utilisés dans l'environnement des soins à domicile



IEC IEC 60601-1-11:2015-01+AMD1:2020-07 CSV(en-fr)

THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembé
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC
The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications
The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform
The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,…). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished
Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc
If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org
The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary
67 000 electrotechnical terminology entries in English and French extracted from the Terms and definitions clause of IEC publications issued between 2002 and 2015. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC
La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC
Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC - webstore.iec.ch/advsearchform
La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,…). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished
Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc
Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org
Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary
67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et définitions des publications IEC parues entre 2002 et 2015. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

This document is a preview generated by EVS

IEC 60601-1-11

Edition 2.1 2020-07
CONSOLIDATED VERSION

INTERNATIONAL STANDARD

NORME INTERNATIONALE

colour inside

This document is a preview generated by EVS

Medical electrical equipment –
Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

Appareils électromédicaux –
Partie 1-11: Exigences générales pour la sécurité de base et les performances essentielles – Norme Collatérale: Exigences pour les appareils électromédicaux et les systèmes électromédicaux utilisés dans l'environnement des soins à domicile

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.020.10; 11.040.01

ISBN 978-2-8322-8709-5

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

This document is a preview generated by EVS

IEC 60601-1-11

Edition 2.1 2020-07

REDLINE VERSION

VERSION REDLINE

This document is a preview generated by EVS

colour inside

Medical electrical equipment –
Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

Appareils électromédicaux –
Partie 1-11: Exigences générales pour la sécurité de base et les performances essentielles – Norme Collatérale: Exigences pour les appareils électromédicaux et les systèmes électromédicaux utilisés dans l'environnement des soins à domicile



IEC IEC 60601-1-11:2015-01+AMD1:2020-07 CSV(en-fr)

CONTENTS

FOREWORD 5
INTRODUCTION 8
INTRODUCTION to Amendment 1 8
1 Scope, object and related standards 10
1.1 * Scope 10
1.2 Object 10
1.3 Related standards 10
1.3.1 IEC 60601-1 10
1.3.2 Particular standards 11
2 Normative references 11
3 Terms and definitions 12
4 General requirements 13
4.1 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS 13
4.2 * Environmental conditions for ME EQUIPMENT 14
4.2.1 General 14
4.2.2 * Environmental conditions of transport and storage between uses 14
4.2.3 * Environmental operating conditions 15
5 * General requirements for testing ME EQUIPMENT 18
6 * Classification of ME EQUIPMENT and ME SYSTEMS 18
7 ME EQUIPMENT identification, marking and documents 19
7.1 * USABILITY of the ACCOMPANYING DOCUMENTS 19
7.2 * Additional requirements for marking of IP classification 19
7.3 ACCOMPANYING DOCUMENTS 20
7.3.1 Contact information 20
7.3.2 LAY OPERATOR briefing information 20
7.4 Instructions for use 20
7.4.1 Additional requirements for warning and safety notices 20
7.4.2 * Additional requirements for an electrical power source 21
7.4.3 Additional requirements for ME EQUIPMENT description 21
7.4.4 Additional requirements for ME EQUIPMENT start-up PROCEDURE 21
7.4.5 Additional requirements for operating instructions 22
7.4.6 Additional requirements for ME EQUIPMENT messages 22
7.4.7 * Additional requirements for cleaning, disinfection and sterilization 22
7.4.8 Additional requirements for maintenance 23
7.4.9 Additional requirements for environmental protection 23
7.4.10 Additional requirements for ME EQUIPMENT and ME SYSTEMS 23
7.5 Technical description 23
7.5.1 PERMANENTLY INSTALLED CLASS I ME EQUIPMENT 23
7.5.2 Additional requirements for professional hygienic maintenance 24
8 Protection against excessive temperatures and other HAZARDS 24
8.1 * Additional requirements for cleaning, disinfection of ME EQUIPMENT and ME SYSTEMS 24
8.2 * Additional requirements for sterilization of ME EQUIPMENT and ME SYSTEMS 24
8.3 Additional requirements for ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS 24

This document is a preview generated by EVS

8.3.1 * Ingress of water or particulate matter into ME EQUIPMENT 24
8.3.2 * Ingress of water or particulate matter into ME SYSTEMS 25
8.4 Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT and ME SYSTEM 25
8.5 Additional requirements for an INTERNAL ELECTRICAL POWER SOURCE 26
8.5.1 * Indication of state 26
8.5.2 Accessibility of small INTERNAL ELECTRICAL POWER SOURCES 26
8.5.3 * Additional requirements for separation of parts 26
9 Accuracy of controls and instruments and protection against hazardous outputs 27
10 Construction of ME EQUIPMENT 27
10.1 * Additional requirements for mechanical strength 27
10.1.1 General requirements for mechanical strength 27
10.1.2 * Requirements for mechanical strength for non-TRANSIT-OPERABLE ME EQUIPMENT 29
10.1.3 * Requirements for mechanical strength for TRANSIT-OPERABLE ME EQUIPMENT 30
10.2 Additional requirements for actuating parts of controls of ME EQUIPMENT 31
11 * Protection against strangulation or asphyxiation 32
12 Additional requirements for ELECTROMAGNETIC EMISSIONS of ME EQUIPMENT and ME SYSTEMS 32
13 Additional requirements for ALARM SYSTEMS of ME EQUIPMENT and ME SYSTEMS 32
13.1 * Additional requirement for generation of ALARM SIGNALS 32
13.2 * Additional requirement for ALARM SIGNAL volume 32
Annex A (informative) General guidance and rationale 33
A.1 General guidance 33
A.2 Rationale for particular clauses and subclauses 34
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS 54
B.1 Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts 54
B.2 Accompanying documents, general 54
B.3 ACCOMPANYING DOCUMENTS, instructions for use 54
B.4 ACCOMPANYING DOCUMENTS, technical description 56
Annex C (informative) Symbols on marking 57
Bibliography 59
Index of defined terms used in this collateral standard 61

Figure 1 – Small finger probe Ø 5,6 18
Figure A.1 – Saturation water vapour pressure as function of temperature 38

Table 1 – Mechanical strength test applicability, non-TRANSIT-OPERABLE 28
Table 2 – Mechanical strength test applicability, TRANSIT-OPERABLE 29
Table A.1 – Saturation water vapour pressure as function of temperature 39
Table A.2 – Summary by use of HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT ENCLOSURE ingress of water and particulate matter requirements 48
Table A.3 – Qualitative assessment of HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT subjected to shock and vibration 50
Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts 54
Table B.2 – ACCOMPANYING DOCUMENTS, general 54

This document is a preview generated by EVS

Table B.3 – ACCOMPANYING DOCUMENTS, instructions for use 55
Table B.4 – ACCOMPANYING DOCUMENTS, technical description 56
Table C.1 – General symbols *(1 of 2)* 57

This document is a preview generated by EVS

INTERNATIONAL ELECTROTECHNICAL COMMISSION

_____

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 60601-1-11 edition 2.1 contains the second edition (2015-01) [documents 62A/959/FDIS and 62A/978/RVD] and its amendment 1 (2020-07) [documents 62A/1395/FDIS and 62A/1410/RVD].

In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.

This document is a preview generated by EVS

International standard IEC 60601-1-11 has been prepared by a joint working group of IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice and ISO subcommittee SC3: Lung ventilators and related devices, of ISO technical committee 121: Anaesthetic and respiratory equipment.

It is published as a double logo standard.

This second edition constitutes a collateral standard to IEC 60601-1 (third edition, including Amendment 1): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance,* hereafter referred to as the general standard.

This second edition cancels and replaces the first edition of IEC 60601-1-11, published in 2010, and constitutes a technical revision.

The most significant changes with respect to the previous edition include the following modifications:

– correction of test method for relative humidity control at temperatures above 35 °C;
– redrafting of subclauses that altered instead of adding to the general standard or other collateral standards; and
– harmonizing with the changes to the amendments to the general standard and other collateral standards.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the IEC 60601 series of publications, collateral standards specify general requirements for safety applicable to:

– a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
– a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

– requirements and definitions: roman type;
– *test specifications: italic type;*
– informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
– TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this collateral standard, the term

– “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
– “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.3.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

This document is a preview generated by EVS

– “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;

– “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

– “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

This document is a preview generated by EVS

INTRODUCTION

Medical practice is increasingly using MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for monitoring, treatment or diagnosis of PATIENTS in the HOME HEALTHCARE ENVIRONMENT (see 3.1). The safety of MEDICAL ELECTRICAL EQUIPMENT in this uncontrolled environment with regard to the electrical installation and its related safety and protection means is a cause for concern.

The potential lack of training of the LAY OPERATOR and possibly of those supervising the use of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM and their level of education need to be addressed in the development of the ACCOMPANYING DOCUMENTS and in the relevant marking on the equipment itself so that this material can be understood. This collateral standard gives special guidance on how this should be addressed in the instructions for use.

This collateral standard was developed with contributions from clinicians, engineers and regulators. The terminology, requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for the development of particular standards.

INTRODUCTION to Amendment 1

The second edition of IEC 60601-1-11 was published in 2015. Since the publication of IEC 60601-1-11:2015, the IEC Subcommittee (SC) 62A Secretariat has been collecting issues from a variety of sources including comments from National Committees. At the November 2015 meeting of IEC/SC 62A in Kobe, Japan, the subcommittee initiated a process to identify high-priority issues that need to be considered in an amendment and should not wait until the third edition of IEC 60601-1-11, which is presently targeted for publication sometime after 2024.

Those issues selected for inclusion on the final "short list" to be addressed in Amendment 1 were those approved by a 2/3 majority of the National Committees present and voting at the Frankfurt meeting of SC 62A. At the meeting held on 10 October 2016, four items were presented to the National Committees present. All four items received the required 2/3 majority of the National Committees present and voting and have been included in the "short list" for consideration in preparing Amendment 1. All remaining issues have been placed on a "long list" for consideration in the third edition of IEC 60601-1-11.

The "short list" of issues was documented in the design specification for Amendment 1. As IEC 60601-1-11 was jointly developed with ISO/TC 121/SC 3, the work was assigned to IEC/SC 62A-ISO/TC 121/SC 3 Joint Working Group (JWG) 6. JWG 6 was directed to consider each issue described in Clause 6 of the design specification and develop an appropriate solution for the identified problem. That final solution in this amendment can encompass any technical solution proposed by the author of the issue or it can involve a different solution developed by the expert group. The expert group can also have recommended that no change to the standard was justified by the problem statement.

Because this is an amendment to IEC 60601-1-11:2015, the style in force at the time of publication of IEC 60601-1-11 has been applied to this amendment. The style specified in ISO/IEC Directives Part 2:2018 has only been applied when implementing the new style guidance would not result in additional editorial changes.

Users of this document should note that when constructing the dated references to specific elements in a standard, such as definitions, amendments are only referenced if they modified the text being cited. For example, if a reference is made to a definition that has not been

This document is a preview generated by EVS

modified by an amendment, then the reference to the amendment is not included in the dated reference.

This document is a preview generated by EVS

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment

This document is a preview generated by EVS

1 Scope, object and related standards

1.1 * Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for use in the HOME HEALTHCARE ENVIRONMENT, as defined in 3.1, and specified by the MANUFACTURER in the instructions for use. This International Standard applies regardless of whether the ME EQUIPMENT or ME SYSTEM is intended for use by a LAY OPERATOR or by trained healthcare personnel.

The HOME HEALTHCARE ENVIRONMENT includes:

– the dwelling place in which a PATIENT lives;

– other places where PATIENTS are present both indoors and outdoors, excluding professional healthcare facility environments where OPERATORS with medical training are continually available when PATIENTS are present.

This International Standard does not apply to ME EQUIPMENT and ME SYSTEMS intended solely for use in the EMERGENCY MEDICAL SERVICES ENVIRONMENT, covered by IEC 60601-1-12 or solely for use in professional healthcare facilities covered by IEC 60601-1 without the additions of IEC 60601-1-12 or this collateral standard. Nonetheless, ME EQUIPMENT or ME SYSTEMS can be intended for multiple use environments, and as such, if also intended for use in the HOME HEALTHCARE ENVIRONMENT, are within the scope of this standard.

EXAMPLE ME EQUIPMENT or ME SYSTEMS intended for both the HOME HEALTHCARE ENVIRONMENT and the professional healthcare facility environment.

NOTE HOME HEALTHCARE ENVIRONMENT ME EQUIPMENT and ME SYSTEMS can frequently be used in locations with unreliable electrical sources and poor electrical grounding.

1.2 Object

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

– "the general standard" designates IEC 60601-1 alone, including any amendments;

– "this collateral standard" designates IEC 60601-1-11 alone, including any amendments;

– "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

This document is a preview generated by EVS

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography on page 59.

CISPR 11:2009, *Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement*

IEC 60068-2-27:2008, *Environmental testing – Part 2-27: Tests – Test Ea and guidance: Shock*

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-64:2008, *Environmental testing Part 2-64: Tests – Test Fh: Vibration, broadband random and guidance*

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*
IEC 60529:1989/AMD1:1999
IEC 60529:1989/AMD2:2013 [1)]

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012 ~~2)~~
IEC 60601-1:2005/AMD2:2020

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*
IEC 60601-1-2:2014/AMD1:2020

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
IEC 60601-1-6:2010/AMD1:2013 ~~3)~~
IEC 60601-1-6:2010/AMD2:2020

---

1) There exists a consolidated edition 2.2 (2013) including IEC 60529:1989 and its Amendment 1 (1999) and Amendment 2 (2013).

~~2) There exists a consolidated edition 3.1(2012) including IEC 60601-1:2005 and its Amendment 1 (2012).~~

~~3) There exists a consolidated edition 3.1 (2013) including IEC 60601-1-6:2010 and its Amendment 1 (2013).~~

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*
IEC 60601-1-8:2006/AMD1:2012 4)
IEC 60601-1-8:2006/AMD2:2020

IEC 60601-1-12:2014, *Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*
IEC 60601-1-12:2014/AMD1:*2020*

~~IEC 62366:2007, *Medical devices – Application of usability engineering to medical devices*~~
~~IEC 62366:2007/AMD1:2014 5)~~

IEC 62366-1:2015, *Medical devices – Part 1: Application of usability engineering to medical devices*
IEC 62366-1:2015/AMD1:2020

ISO 7000, *Graphical symbols for use on equipment — Registered symbols*. Available from: http://www.graphical-symbols.info/equipment

~~ISO 7010:2011, *Graphical symbols – Safety colours and safety signs – Registered safety signs*~~
~~ISO 7010:2011/AMD1:2012~~
~~ISO 7010:2011/AMD2:2012~~
~~ISO 7010:2011/AMD3:2012~~
~~ISO 7010:2011/AMD4:2013~~
~~ISO 7010:2011/AMD5:2014~~

ISO 7010:2019, *Graphical symbols – Safety colours and safety signs – Registered safety signs*

ISO 15223-1:~~2012~~2016, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

3 Terms and definitions

~~For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 and IEC 60601-1:2005:AMD1:2012, IEC 60601-1-2:2014, IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-12:2014, IEC 62366:2007 and the following definitions apply.~~

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 and IEC 60601-1:2005:AMD1:2012 + IEC 60601-1:2005/AMD2:2020, IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013 + IEC 60601-1-6:2010/AMD2:2020, IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012 + IEC 60601-1-8:2006/AMD2:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020, IEC 62366-1:2015 and IEC 62366-1:2015/AMD1:2020, and the following definitions apply.

NOTE 1 Where the terms “voltage” and “current” are used in this document, they mean the r.m.s. values of an alternating, direct or composite voltage or current unless stated otherwise.

~~4) There exists a consolidated edition 2.1 (2012) including IEC 60601-1-8:2006 and its Amendment 1 (2012).~~

~~5) There exists a consolidated edition 2.1 (2014) including IEC 62366:2007 and Amendment 1 (2014).~~

This document is a preview generated by EVS