

Dosimetry for exposures to cosmic radiation in civilian aircraft - Part 1: Conceptual basis for measurements
(ISO 20785-1:2020)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 20785-1:2020 sisaldab Euroopa standardi EN ISO 20785-1:2020 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 20785-1:2020 consists of the English text of the European standard EN ISO 20785-1:2020.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 05.08.2020.	Date of Availability of the European standard is 05.08.2020.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.280, 49.020

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Dosimetry for exposures to cosmic radiation in civilian
aircraft - Part 1: Conceptual basis for measurements (ISO
20785-1:2020)**

Dosimétrie pour l'exposition au rayonnement
cosmique à bord d'un avion civil - Partie 1: Fondement
théorique des mesurages (ISO 20785-1:2020)

Dosimetrie für die Belastung durch kosmische
Strahlung in Zivilluftfahrzeugen - Teil 1:
Konzeptionelle Grundlage für Messungen (ISO 20785-
1:2020)

This European Standard was approved by CEN on 1 July 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 20785-1:2020) has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" in collaboration with Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2021, and conflicting national standards shall be withdrawn at the latest by February 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 20785-1:2017.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20785-1:2020 has been approved by CEN as EN ISO 20785-1:2020 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
3.1 General terms	1
3.2 Quantities and units	2
3.3 Atmospheric radiation field	4
4 General considerations	6
4.1 The cosmic radiation field in the atmosphere	6
4.2 General calibration considerations for the dosimetry of cosmic radiation fields in aircraft	7
4.2.1 Approach	7
4.2.2 Considerations concerning the measurement	7
4.2.3 Considerations concerning the radiation field	8
4.2.4 Considerations concerning calibration	8
4.2.5 Simulated aircraft fields	9
4.3 Conversion coefficients	9
5 Dosimetric devices	10
5.1 Introduction	10
5.2 Active devices	10
5.2.1 Devices to determine all field components	10
5.2.2 Devices for low LET/non-neutron	11
5.2.3 Devices for high-LET/neutron component	12
5.3 Passive devices	13
5.3.1 General considerations	13
5.3.2 Etched track detectors	14
5.3.3 Fission foil detectors	14
5.3.4 Superheated emulsion neutron detectors (bubble) detectors	14
5.3.5 Thermoluminescent detectors	15
5.3.6 Photoluminescent detectors	15
Annex A (informative) Representative particle fluence rate energy distributions for the cosmic radiation field at flight altitudes for solar minimum and maximum conditions and for minimum and maximum vertical cut-off rigidity	16
Bibliography	22

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiation protection*.

This third edition cancels and replaces the second edition (ISO 20785-1:2012), which has been technically revised. The main changes are as follows:

- revision of the terms and definitions;
- updated references.

A list of all the parts in the ISO 20785 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Aircraft crews are exposed to elevated levels of cosmic radiation of galactic and solar origin and secondary radiation produced in the atmosphere, the aircraft structure and its contents. Following recommendations of the International Commission on Radiological Protection (ICRP) in Publication 60^[1], confirmed by Publication 103^[2], the European Union (EU) introduced a revised Basic Safety Standards Directive^[3] and International Atomic Energy Agency (IAEA)^[4] issued a revised Basic Safety Standards. Those standards included exposure to natural sources of ionizing radiation, including cosmic radiation, as occupational exposure. The EU Directive requires account to be taken of the exposure of aircraft crews liable to receive more than 1 mSv per year. It then identifies the following four protection measures:

- a) to assess the exposure of the crew concerned;
- b) to take into account the assessed exposure when organizing working schedules with a view to reducing the doses of highly exposed crews;
- c) to inform the workers concerned of the health risks their work involves; and
- d) to apply the same special protection during pregnancy to female crews in respect of the "child to be born" as to other female workers.

The EU Council Directive has already been incorporated into laws and regulations of EU Member States and is being included in the aviation safety standards and procedures of the Joint Aviation Authorities and the European Air Safety Agency. Other countries such as Canada and Japan have issued advisories to their airline industries to manage aircraft crew exposure.

For regulatory and legislative purposes, the radiation protection quantities of interest are the equivalent dose (to the foetus) and the effective dose. The cosmic radiation exposure of the body is essentially uniform and the maternal abdomen provides no effective shielding to the foetus. As a result, the magnitude of equivalent dose to the foetus can be set equal to that of the effective dose received by the mother. Doses on board aircraft are generally predictable, and events comparable to unplanned exposure in other radiological workplaces cannot normally occur (with the rare exceptions of extremely intense and energetic solar particle events). Personal dosimeters for routine use are not considered necessary. The preferred approach for the assessment of doses of aircraft crews, where necessary, is to calculate directly the effective dose per unit time, as a function of geographic location, altitude and solar cycle phase, and to combine these values with flight and staff roster information to obtain estimates of effective doses for individuals. This approach is supported by guidance from the European Commission and the ICRP in Publications 75^[5] and 132^[6].

The role of calculations in this procedure is unique in routine radiation protection and it is widely accepted that the calculated doses should be validated by measurement. The effective dose is not directly measurable. The operational quantity of interest is ambient dose equivalent, $H^*(10)$. In order to validate the assessed doses obtained in terms of effective dose, calculations can be made of ambient dose equivalent rates or route doses in terms of ambient dose equivalent, and values of this quantity determined from measurements. Traceability should be provided for a reasonable number of particle types and energies of the atmospheric radiation field, corrections included for differences between the calibration fields and the total atmospheric radiation field, and related uncertainties properly taken into account. The validation of calculations of ambient dose equivalent for a particular calculation method may be taken as a validation of the calculation of the effective dose by the same computer code, but this step in the process may need to be confirmed. The alternative is to establish a priori that the operational quantity ambient dose equivalent is a good estimator of effective dose and equivalent dose to the foetus for the radiation fields being considered, in the same way that the use of the operational quantity personal dose equivalent is justified for the estimation of effective dose for ground-based radiation workers.

The radiation field in aircraft at altitude is complex, with many types of ionizing radiation present, with energies ranging up to many GeV. The determination of ambient dose equivalent for such a complex radiation field is difficult. In many cases, the methods used for the determination of ambient dose

equivalent in aircraft are similar to those used at high-energy accelerators in research laboratories. Therefore, it is possible to recommend dosimetric methods and methods for the calibration of dosimetric devices, as well as the techniques for maintaining the traceability of dosimetric measurements to national standards. Dosimetric measurements made to evaluate ambient dose equivalent should be performed using accurate and reliable methods that ensure the quality of readings provided to workers and regulatory authorities. This document gives a conceptual basis for the characterization of the response of instruments for the determination of ambient dose equivalent in aircraft.

Requirements for the determination and recording of the cosmic radiation exposure of aircraft crews have been introduced into the national legislation of EU Member States and other countries. Harmonization of methods used for determining ambient dose equivalent and for calibrating instruments is desirable to ensure the compatibility of measurements performed with such instruments.

This document is intended for the use of primary and secondary calibration laboratories for ionizing radiation, by radiation protection personnel employed by governmental agencies, and by industrial corporations concerned with the determination of ambient dose equivalent for aircraft crews.

Dosimetry for exposures to cosmic radiation in civilian aircraft —

Part 1: Conceptual basis for measurements

1 Scope

This document specifies the conceptual basis for the determination of ambient dose equivalent for the evaluation of exposure to cosmic radiation in civilian aircraft and for the calibration of instruments used for that purpose.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 General terms

3.1.1 calibration

operation that, under specified conditions, establishes a relation between the conventional quantity, H_0 , and the indication, G

Note 1 to entry: A calibration can be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it can consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called "self-calibration", or with verification of calibration.

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

3.1.2 response response characteristic

R

quotient of the indication, G , or the corrected indication, G_{corr} , and the conventional quantity value to be measured

Note 1 to entry: To avoid confusion, it is necessary to specify which of the quotients, given in the definition of the response (to G or to G_{corr}) is applied. Furthermore, it is necessary, in order to avoid confusion, to state the quantity to be measured, for example: the response with respect to fluence, R_Φ , the response with respect to kerma, R_K , the response with respect to absorbed dose, R_D .