

**TECHNICAL** 

REPORT

# IEC TR 63099-2

Edition 1.0 2020-07



Transmitting equipment for radiocommunication – Radio-over-fibre technologies for electromagnetic-field measurement – Part 2: Radio-over-fibre technologies for electric-field sensing



## THIS PUBLICATION IS COPYRIGHT PROTECTED Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

**IEC Central Office** 3, rue de Varembé CH-1211 Geneva 20 Switzerland

Tel.: +41 22 919 02 11 info@iec.ch www.iec.ch

#### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

#### Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.



Edition 1.0 2020-07





Transmitting equipment for radiocommunication - Radio-over-fibre technologies for electromagnetic-field measurement -Part 2: Radio-over-fibre technologies for electric-field sensing

4

**INTERNATIONAL** ELECTROTECHNICAL COMMISSION

ISBN 978-2-8322-8493-3

ICS 33.060.20; 33.180.99

Warning! Make sure that you obtained this publication from an authorized distributor.

# CONTENTS

INTRODUCTION	FOR	EWORD	3
2       Normative references       6         3       Terms, definitions and abbreviated terms       6         3.1       Terms and definitions       6         3.2       Abbreviated terms       7         4       Practical examples of electric-field sensing system using RoF technologies       7         4.1       Overview       7         4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16       Figure 2 - Structure of the sensor head unit       9         Figure 3 - 3-axis electric-field sensing system using LN optical modulator       10         Figure 5 - Evaluation results of sensitivity and measurement dynamic range       11         Figure 6 - Measurement results of sensitivity and measurement dynamic range       11         Figur	INTR	RODUCTION	5
3       Terms, definitions and abbreviated terms       6         3.1       Terms and definitions       6         3.2       Abbreviated terms       7         4       Practical examples of electric-field sensing system using RoF technologies       7         4.1       Overview       7         4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       4.4.1         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 5 – Evaluation of sensor isotropy of the conventional diode-type electric-field sensor using LN modulator       12         Figure 6 – Measurement results of sensitivity pattern	1	Scope	6
3.1       Terms and definitions       6         3.2       Abbreviated terms       7         4       Practical examples of electric-field sensing system using RoF technologies       7         4.1       Overview       7         4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 5 – Evaluation of sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor u	2	Normative references	6
3.1       Terms and definitions       6         3.2       Abbreviated terms       7         4       Practical examples of electric-field sensing system using RoF technologies       7         4.1       Overview       7         4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 5 – Evaluation of sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor u			
3.2       Abbreviated terms       7         4       Practical examples of electric-field sensing system using RoF technologies       7         4.1       Overview       7         4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Ceil up to 1 GHz       11         Figure 6 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor using LN modulator       12 <t< td=""><td></td><td></td><td></td></t<>			
4.1       Overview       7         4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4       System configuration       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 5 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor using LN modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor using LN optical modulator       13	3.1		
4.2       Features of electric-field sensing system using RoF technologies       7         4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 5 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Ceil up to 1 GHz       11         Figure 6 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor using LN modulator       12         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 9 – Schematic representation of the electric-field sensor using VCSEL co	4	Practical examples of electric-field sensing system using RoF technologies	7
4.3       List of implementation examples       7         4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy of the conventional diode-type electric-field sensor using LN modulator       12         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor using LN optical modulator       13         <	4.	1 Overview	7
4.4       3-axis electric-field sensor using LN optical modulators       7         4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy of the conventional diode-type electric-field sensor using LN modulator       12         Figure 7 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusi	4.	2 Features of electric-field sensing system using RoF technologies	7
4.4.1       System configuration       7         4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 10 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15	4.	3 List of implementation examples	7
4.4.2       Specifications       9         4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor using LN modulator according to       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specifica	4.	4 3-axis electric-field sensor using LN optical modulators	7
4.4.3       Example of measurement results       10         4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-	4	4.4.1 System configuration	7
4.5       Bulk-type electric-field sensor using ZnTe optical modulators       13         4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       12         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	4		
4.6       Electric-field probes using VCSEL       14         Bibliography       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 5 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	4		
Bibliography.       16         Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range.       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz.       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators.       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	4.		
Figure 1 – System diagram of the optical E-field sensor       8         Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9			
Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Biblic	ography	16
Figure 2 – Structure of the sensor head unit       9         Figure 3 – 3-axis electric-field sensing system using LN optical modulator       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9			
Figure 3 – 3-axis electric-field sensing system using LN optical modulator.       10         Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz.       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	-		
Figure 4 – Evaluation results of sensitivity and measurement dynamic range       11         Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type electric-field sensor and electric-field sensor using LN modulator according to       12         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe optical modulators       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Figur	re 2 – Structure of the sensor head unit	9
Figure 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz.       11         Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       15         Consisting of a miniature sensor head that is exclusively linked via fibre optics to a       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Figur	re 3 – 3-axis electric-field sensing system using LN optical modulator	10
Figure 6 – Measurement setup for isotropy of the conventional diode-type electric-field       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type       12         electric-field sensor and electric-field sensor using LN modulator according to       12         IEEE Std. 1309       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type       13         electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       15         consisting of a miniature sensor head that is exclusively linked via fibre optics to a       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Figur	re 4 – Evaluation results of sensitivity and measurement dynamic range	11
sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type       12         electric-field sensor and electric-field sensor using LN modulator according to       13         IEEE Std. 1309       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type       13         electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         consisting of a miniature sensor head that is exclusively linked via fibre optics to a       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Figur	re 5 – Evaluation of sensor isotropy in the TEM-Cell up to 1 GHz	11
sensor and electric-field sensor using LN modulator       12         Figure 7 – Measurement results of sensitivity pattern of the conventional diode-type       12         electric-field sensor and electric-field sensor using LN modulator according to       13         IEEE Std. 1309       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type       13         electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         consisting of a miniature sensor head that is exclusively linked via fibre optics to a       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Figur	re 6 – Measurement setup for isotropy of the conventional diode-type electric-field	
electric-field sensor and electric-field sensor using LN modulator according to IEEE Std. 1309			12
IEEE Std. 1309       13         Figure 8 – Frequency characteristics of isotropy of the conventional diode-type       13         electric-field sensor and electric-field sensor using LN optical modulator       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9	Figur	re 7 – Measurement results of sensitivity pattern of the conventional diode-type	
Figure 8 – Frequency characteristics of isotropy of the conventional diode-type       13         Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         Consisting of a miniature sensor head that is exclusively linked via fibre optics to a       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9			10
electric-field sensor and electric-field sensor using LN optical modulator			13
Figure 9 – Schematic representation of the bulk-type electric-field sensor using ZnTe       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL       14         consisting of a miniature sensor head that is exclusively linked via fibre optics to a       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9			13
optical modulators.       14         Figure 10 – Schematic representation of the electric-field sensor using VCSEL consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit       15         Table 1 – Specification of 3-axis electric-field sensing system using LN optical modulator       9         Table 2 – Specification of 3-axis electric-field sensing system using LN optical       9			
consisting of a miniature sensor head that is exclusively linked via fibre optics to a remote unit			14
<ul> <li>remote unit</li></ul>	Figur	re 10 – Schematic representation of the electric-field sensor using VCSEL	
Table 1 – Specification of 3-axis electric-field sensing system using LN optical         modulator         Table 2 – Specification of 3-axis electric-field sensing system using LN optical			45
modulator	remo	DTE UNIT	15
modulator			
Table 2 – Specification of 3-axis electric-field sensing system using LN optical			0
			9
			12

#### INTERNATIONAL ELECTROTECHNICAL COMMISSION

# TRANSMITTING EQUIPMENT FOR RADIOCOMMUNICATION – RADIO-OVER-FIBRE TECHNOLOGIES FOR ELECTROMAGNETIC-FIELD MEASUREMENT –

#### Part 2: Radio-over-fibre technologies for electric-field sensing

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC TR 63099-2, which is a Technical Report, has been prepared by IEC technical committee 103: Transmitting equipment for radiocommunication.

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
103/184/DTR	103/186A/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 63099 series, published under the general title *Transmitting* equipment for radiocommunication – Radio-over-fibre technologies for electromagnetic-field measurement, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

Orozioz Orozioz O

#### INTRODUCTION

This document provides information on the current and latest applications for electric-field sensing using radio-over-fibre technology. Electric-field measurement systems are covered, which are practically in use or will be used soon. It will be beneficial to system developers and h biguire. system users in the fields of electric-field measurement. As a Technical Report, this document contains no requirements and is informative only.

# TRANSMITTING EQUIPMENT FOR RADIOCOMMUNICATION – RADIO-OVER-FIBRE TECHNOLOGIES FOR ELECTROMAGNETIC-FIELD MEASUREMENT –

Part 2: Radio-over-fibre technologies for electric-field sensing

# 1 Scope

The purpose of this part of IEC 63099 is to provide information about the current and latest applications for electric-field measurement that use radio-over-fibre technologies. System configurations, specifications, and measurement examples of each electric-field measurement system are included. The theoretical background of electric-field measurement and calibration method of electric-field sensors are beyond the scope of this document.

#### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEEE Std. 145-2013, IEEE Standard for Definitions of Terms for Antennas

## 3 Terms, definitions and abbreviated terms

#### 3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in IEEE Std. 145-2013 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at http://www.iso.org/obp

#### 3.1.1 O/E converter optical to electrical converter converter which directly converts optical signals into electrical signal

Note 1 to entry: A photo-diode is generally used as an O/E converter device

# 3.1.2

**isotropy** uniform sensitivity for all spherical direction

Note 1 to entry: Sometimes it is misunderstood as equivalence of each axis sensor of 3-axis.

3.1.3

#### TEM-Cell

transverse electromagnetic field cell waveguide which can generate a certain level of uniform electric-fields in the cell