

This document is a preview generated by EVS

**General requirements of tissue-
engineered medical products**



This document is a preview generated by ELS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements for tissue-engineered medical products	4
5 Materials	6
5.1 General.....	6
5.2 Starting materials.....	6
5.2.1 Cells and tissues.....	6
5.2.2 Scaffolds and other substances.....	7
5.3 Ancillary materials.....	7
5.4 Excipients.....	8
6 Manufacturing process	8
7 Characterization and control strategy	8
8 Developmental studies	9
9 Container closure system	9
10 Transportation	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 7, *Tissue-engineered medical products*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Advances in the field of biological sciences have made possible the generation of a new type of medical product that when administered to the human body, may repair, replace, regenerate or enhance the function of impaired tissues or organs.

Extensive experience acquired through the administration of living human cells has yielded solid knowledge about the quality requirements and the risks associated with their use.

However, the development of tissue-engineered medical products (TEMPs) that are not simply obtained from a human donor or by separating living tissues, but rather are grown from various cell sources and are manipulated during manufacture to meet the medical needs of the patient, introduces new challenges with regard to quality requirements and risk management for the benefit of patients.

TEMPs utilizing human material are quite diverse but share a set of common quality requirements for their safe use. These kinds of products require special attention for contamination control, such as infectious agents transmitting disease (e.g. hepatitis, HIV, TSE) and harmful chemicals, unintended decomposition or degradation induced by inappropriate handling at any stage of the manufacturing process, tumorigenic potential, induction of an immunogenic reaction in the recipient, traceability of cells, critical materials, and the final product are key to product quality and its safe use.

This document has been developed with the objective of assisting interested parties, such as manufacturers and regulators, establish suitable quality parameters and specifications for the final TEMP as well as cells, critical materials, processing steps and appropriate controls ensuring the safety of TEMP.

General requirements of tissue-engineered medical products

1 Scope

This document specifies general requirements for tissue-engineered medical products (TEMPs), which are used in regenerative medicine. With regard to safety, this document outlines requirements for materials, manufacture, quality control, and unintentional biological effects elicited by TEMP. This document does not address requirements for clinical trials and efficacy.

This document is not applicable to tissue-engineered products used for diagnosis, *ex-vivo* testing or extracorporeal treatments of patients (e.g. dialysis with TEMP components). TEMP containing viable xenogenic cells, genetically modified cells, or cells derived from abnormal cells or tissues (e.g. cancerous tissues) are also excluded from the scope. The combination of TEMP with medical devices, with the exception of scaffolds comprised of synthetic and/or naturally-derived (e.g. animal sourced) materials, is also excluded from the scope.

NOTE International, national or regional regulations or requirements, or the Pharmacopeia also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 13022, *Medical products containing viable human cells — Application of risk management and requirements for processing practices*

ISO/TS 20399-1, *Biotechnology — Ancillary materials present during the production of cellular therapeutic products — Part 1: General requirements*

ISO/TS 20399-2, *Biotechnology — Ancillary materials present during the production of cellular therapeutic products — Part 2: Best practice guidance for ancillary material suppliers*

ISO/TS 20399-3, *Biotechnology — Ancillary materials present during the production of cellular therapeutic products — Part 3: Best practice guidance for ancillary material users*

ISO 22442-1, *Medical devices utilizing animal tissues and their derivatives — Part 1: Application of risk management*

ISO 22442-2, *Medical devices utilizing animal tissues and their derivatives — Part 2: Controls on sourcing, collection and handling*

ISO 22442-3, *Medical devices utilizing animal tissues and their derivatives — Part 3: Validation of the elimination and/or inactivation of viruses and transmissible spongiform encephalopathy (TSE) agents*

ISO/TR 22442-4, *Medical devices utilizing animal tissues and their derivatives — Part 4: Principles for elimination and/or inactivation of transmissible spongiform encephalopathy (TSE) agents and validation assays for those processes*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.