

ANESTEESIA- JA HINGAMISSEADMED.
LARÜNGOSKOOBID TRAHHEA INTUBEERIMISEKS

Anaesthetic and respiratory equipment - Laryngoscopes
for tracheal intubation (ISO 7376:2020)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 7376:2020 sisaldab Euroopa standardi EN ISO 7376:2020 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 7376:2020 consists of the English text of the European standard EN ISO 7376:2020.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 23.09.2020.	Date of Availability of the European standard is 23.09.2020.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 7376

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2020

ICS 11.040.10

Supersedes EN ISO 7376:2009

English Version

Anaesthetic and respiratory equipment - Laryngoscopes for tracheal intubation (ISO 7376:2020)

Matériel d'anesthésie et de réanimation respiratoire -
Laryngoscopes pour intubation trachéale (ISO
7376:2020)

Anästhesie- und Beatmungsgeräte - Laryngoskope für
Trachealintubation (ISO 7376:2020)

This European Standard was approved by CEN on 26 June 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 7376:2020) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2021, and conflicting national standards shall be withdrawn at the latest by March 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 7376:2009.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 7376:2020 has been approved by CEN as EN ISO 7376:2020 without any modification.

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 *Scope	1
2 Normative references	1
3 *Terms and definitions	1
4 General requirements	2
5 Materials	2
6 Design requirements	2
6.1 General.....	2
6.2 *Environmental conditions.....	2
7 Performance requirements	3
7.1 <i>Handles</i>	3
7.1.1 *Electrical safety.....	3
7.1.2 Electrical <i>contact</i>	3
7.1.3 Dimensions.....	3
7.1.4 *Optical output.....	5
7.1.5 *Strength.....	5
7.2 <i>Laryngoscope blades</i>	5
7.2.1 Dimensions.....	5
7.2.2 *Rigidity.....	5
7.2.3 *Strength.....	5
7.2.4 *Illumination.....	6
7.2.5 Electrical interface.....	6
7.3 <i>Hinged combinations</i>	6
7.4 <i>Single-piece laryngoscopes</i>	7
8 Cleaning, disinfection and sterilization	8
9 Information supplied by the manufacturer	8
9.1 General.....	8
9.2 Marking.....	8
9.3 Designated size for the intended patient demographic.....	8
9.4 Instructions for use.....	8
Annex A (informative) Rationale	10
Annex B (normative) Test method for illumination, strength and rigidity	12
Annex C (informative) Common terms for parts of a laryngoscope blade	15
Bibliography	16

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 2, *Airways and related equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 7376:2009), which has been technically revised.

The main changes compared to the previous edition are as follows:

- the layout has changed to align with ISO 18190:2016;
- removal of the requirements for filament *luminaires*, which are now almost entirely replaced by LED *luminaires*;
- additional requirement to prevent current surges from batteries with a rated voltage exceeding 3,6 V;
- test methods have been rationalized;
- each requirement is now followed by a conformance requirement;
- the strength requirement has been extended to *handles*;
- environmental conditions during transport, storage and normal use are now referenced to IEC 60601-1-12.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Laryngoscopes are manufactured in several forms. There are, for example, *single-piece laryngoscopes*, *non-user-detachable hinged laryngoscope blades*, or detachable hinged laryngoscope blades and *handles*. In the latter case, the light source for illuminating the larynx during use is either a *luminaire* fixed in the laryngoscope blade or a *luminaire* in the *handle* with a light guide in the laryngoscope blade. The minimum light output from the laryngoscope is specified.

Laryngoscope blades are defined and marked by size designation (see [Table 2](#)), suitable for the patient demographic for which they are designed and ranges from small, premature infants to extra-large adults. Because there are so many variations, the length of laryngoscope blade is disclosed on the packaging to allow an informed decision by the operator to select the most appropriate instrument for intubation. [Annex B](#) describes test methods for illumination, rigidity and strength.

Throughout this document the following print types are used:

- Requirements and definitions: roman type;
- *Conformance tests: italic type;*
- Informative material appearing outside of tables, such as notes, examples and references: smaller type. The Normative text of tables is also in smaller type;
- *Terms defined in [Clause 3](#): italics.*

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in [Annex A](#).

Anaesthetic and respiratory equipment — Laryngoscopes for tracheal intubation

1 *Scope

This document, which is device-specific, specifies requirements for laryngoscopes with non-flexible blades, with internal battery-operated power sources, used for illuminating the larynx during intubation. It also specifies critical dimensions for those *handles* and laryngoscope blades with interchangeable *hook-on fittings*.

It is not applicable to the following:

- flexible laryngoscopes;
- laryngoscopes designed for surgery;
- laryngoscopes powered from mains electricity supply;
- laryngoscopes connected by light-transmitting cables to external light sources;
- video laryngoscopes designed to work with an external, integral or attached video system.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4135, *Anaesthetic and respiratory equipment — Vocabulary*

ISO 18190:2016, *Anaesthetic and respiratory equipment — General requirements for airways and related equipment*

IEC 60601-1-12:2014, *Medical electrical equipment — Part 1-12: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the emergency medical services environment*

3 *Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4135, ISO 18190 and the following apply:

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

contact

metallic part that provides the electrical connection between a hinged laryngoscope blade and *handle* (3.4)

3.2

directly illuminating laryngoscope blade

laryngoscope blade incorporating a *luminaire* (3.8)