

MITTEPURUSTAV KATSETAMINE.
ULTRAHELIKATSESEADMETE MÄÄRATLEMINE JA
KONTROLLIMINE. OSA 1: SEADMED

Non-destructive testing - Characterization and
verification of ultrasonic test equipment - Part 1:
Instruments (ISO 22232-1:2020)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 22232-1:2020 sisaldab Euroopa standardi EN ISO 22232-1:2020 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 22232-1:2020 consists of the English text of the European standard EN ISO 22232-1:2020.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 19.08.2020.	Date of Availability of the European standard is 19.08.2020.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 19.100

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Non-destructive testing - Characterization and verification
of ultrasonic test equipment - Part 1: Instruments (ISO
22232-1:2020)**

Essais non destructifs - Caractérisation et vérification
de l'appareillage de contrôle par ultrasons - Partie 1:
Appareils (ISO 22232-1:2020)

Zerstörungsfreie Prüfung - Charakterisierung und
Verifizierung der Ultraschall-Prüfausrüstung - Teil 1:
Prüfgeräte (ISO 22232-1:2020)

This European Standard was approved by CEN on 4 July 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 22232-1:2020) has been prepared by Technical Committee ISO/TC 135 "Non-destructive testing" in collaboration with Technical Committee CEN/TC 138 "Non-destructive testing" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2021, and conflicting national standards shall be withdrawn at the latest by February 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 12668-1:2010.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 22232-1:2020 has been approved by CEN as EN ISO 22232-1:2020 without any modification.

Contents

Page

Foreword	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols	3
5 General requirements of conformity	4
6 Manufacturer's technical specification for ultrasonic instruments	4
7 Performance requirements for ultrasonic instruments	7
8 Group 1 tests	9
8.1 Equipment required for group 1 tests	9
8.2 Battery operational time	9
8.2.1 Procedure	9
8.2.2 Acceptance criterion	10
8.3 Stability after warm-up time	10
8.3.1 Procedure	10
8.3.2 Acceptance criteria	10
8.4 Stability against temperature	10
8.4.1 Procedure	10
8.4.2 Acceptance criterion	12
8.5 Stability against voltage variation	12
8.5.1 Procedure	12
8.5.2 Acceptance criterion	13
8.6 Time base deviation	13
8.6.1 Procedure	13
8.6.2 Acceptance criterion	15
8.7 Transmitter pulse parameters	15
8.7.1 General	15
8.7.2 Pulse repetition frequency	15
8.7.3 Effective output impedance	15
8.8 Receiver	16
8.8.1 General	16
8.8.2 Cross talk from transmitter to receiver during transmission	16
8.8.3 Dead time after transmitter pulse	17
8.8.4 Dynamic range and maximum input voltage	19
8.8.5 Receiver input impedance	20
8.8.6 Time-corrected gain (TCG)	21
8.9 Gates	22
8.9.1 General	22
8.9.2 Gates with value output	23
8.9.3 Gates with analogue output	25
8.9.4 Gates with alarm output	27
8.10 Highest digitized frequency	28
8.10.1 Procedure	28
8.10.2 Acceptance criterion	29
8.11 Response time of digital ultrasonic instruments	29
8.11.1 General	29
8.11.2 Procedure	29
8.11.3 Acceptance criterion	30
9 Group 2 tests	30
9.1 Equipment required for group 2 tests	30
9.2 Physical state and external aspects	31

9.2.1	Procedure.....	31
9.2.2	Acceptance criterion.....	31
9.3	Transmitter voltage, pulse rise time and duration	31
9.3.1	Procedure.....	31
9.3.2	Acceptance criteria	34
9.4	Receiver.....	34
9.4.1	General.....	34
9.4.2	Frequency response.....	34
9.4.3	Noise level.....	36
9.4.4	Gain linearity.....	37
9.4.5	Vertical display linearity	37
Annex A (normative) Special conditions for ultrasonic instruments with logarithmic amplifiers		39
Bibliography		40

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 135, *Non-destructive testing*, Subcommittee SC 3, *Ultrasonic testing*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 138, *Non-destructive testing*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 22232 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Non-destructive testing — Characterization and verification of ultrasonic test equipment —

Part 1: Instruments

1 Scope

This document specifies methods and acceptance criteria within the frequency range of 0,5 MHz to 15 MHz, for assessing the electrical performance of digital ultrasonic instruments for pulse operation using A-scan display, for manual ultrasonic non-destructive testing with single- or dual-transducer probes. This document is also applicable for multi-channel instruments. This document can partly be applicable to ultrasonic instruments in automated systems, but other tests can be needed to ensure satisfactory performance.

This document excludes ultrasonic instruments for continuous waves.

This document also excludes ultrasonic phased array instruments, see e.g. ISO 18563-1. If a phased array instrument has dedicated connectors for single- or dual-transducer probes this document is applicable for these channels.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5577, *Non-destructive testing — Ultrasonic testing — Vocabulary*

ISO/IEC 17050-1, *Conformity assessment — Supplier's declaration of conformity — Part 1: General requirements*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 5577 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

analogue output

output from the ultrasonic instrument which gives a d.c. voltage nominally proportional to the amplitude of the largest received signal within a monitor gate

3.2

cross talk during transmission

amount of signal transfer from the transmitter output to the receiver input during the transmission pulse, with the ultrasonic instrument set for separate transmitter-receiver operation (dual-transducer probe)