# **EESTI STANDARD**

Dentistry - Portable dental equipment for use in **non-permanent healthcare environment - Part 1:** General requirements (ISO 23402-1:2020)



### EESTI STANDARDI EESSÕNA

#### NATIONAL FOREWORD

3			
See Eesti standard EVS-EN ISO 23402-1:2020 sisaldab Euroopa standardi EN ISO 23402-1:2020 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 23402-1:2020 consists of the English text of the European standard EN ISO 23402-1:2020.		
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.		
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 07.10.2020.	Date of Availability of the European standard is 07.10.2020.		
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.		
B O CL.			

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile <u>standardiosakond@evs.ee</u>.

#### ICS 11.060.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega: Koduleht <u>www.evs.ee</u>; telefon 605 5050; e-post <u>info@evs.ee</u>

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

# EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

# EN ISO 23402-1

October 2020

ICS 11.060.20

**English Version** 

## Dentistry - Portable dental equipment for use in nonpermanent healthcare environment - Part 1: General requirements (ISO 23402-1:2020)

Médecine bucco-dentaire - Matériel dentaire portatif utilisable dans des environnements de soins de santé non permanents - Partie 1: Exigences générales (ISO 23402-1:2020) Zahnheilkunde - Bewegliche dentale Ausrüstung zur Anwendung in nicht-permanenten Gesundheitseinrichtungen - Teil 1: Allgemeine Anforderungen (ISO 23402-1:2020)

This European Standard was approved by CEN on 3 October 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels** 

### **European foreword**

This document (EN ISO 23402-1:2020) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2021, and conflicting national standards shall be withdrawn at the latest by April 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Endorsement notice**

The text of ISO 23402-1:2020 has been approved by CEN as EN ISO 23402-1:2020 without any modification.

## Contents

Page

Fore	word		iv	
Intro	oductio	0 <b>n</b>	v	
1	Scon	pe	1	
2	Norr	mativa rafaranças	1	
2	Тот			
3	Tern	ierms and definitions		
4	Class	sification	2	
	4.1	General	Z	
	4.2	For electrically operated devices	Z	
	4.5	4.3 According to intended use environment		
	4.4 4.5	4.4 According to supply sources		
_	4.5	According to transport conditions		
5	Requ	uirements		
	5.1	General		
	5.2	Iransport requirements		
		5.2.1 General.		
		5.2.2 Grips of other nationing devices		
		5.2.5 Maximum dimensions		
		5.2.4 Maximum unnensions		
		5.2.6 Impact		
		5.2.0 Impact	4	
		5.2.7 Drop	4	
		5.2.9 Particulate and liquid ingress during transport	4	
		5.2.10 Flammability	5	
	5.3	Assembly and disassembly requirements	5	
	5.4	Utility requirements.	5	
	5.5	Operational requirements	5	
		5.5.1 Ambient operating conditions		
		5.5.2 Usability	5	
		5.5.3 Applied parts not intended to supply heat to a patient.	5	
		5.5.4 Cleaning and disinfection of external surfaces		
		5.5.5 Noise	6	
6	Sam	unling	6	
-	Sumpring.			
7	Meas	isurement and test methods		
	7.1	Visual inspection of the device		
	1.2	Visual inspection of the documentation	6	
8	Man	nufacturer's instructions	6	
	8.1	General		
	8.2 Instructions for use			
	8.3	Technical description	7	
9	Marl	king	7	
-	9.1	Marking on the equipment		
	9.2	Marking of packaging		
10	Dach	kaging	0	
10	Fack	naging	δ	
Anne	ex A (in	nformative) Applicable testing methodology from IEC 60601-1:2005/Am	d 1:20129	

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see <a href="https://www.iso.org/directives">www.iso.org/directives</a>).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see <a href="https://www.iso.org/patents">www.iso.org/patents</a>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see <a href="https://www.iso.org/iso/foreword.html">www.iso.org/iso/foreword.html</a>.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 23402 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at <u>www.iso.org/members.html</u>.

## Introduction

Transportable dental equipment is used by dental professionals to provide care to patients in a variety of settings. Because the intended use applications and intended means for transporting such equipment vary considerably, a wide variety of transportable dental equipment is commercially available. For example, certain transportable equipment is designed and constructed to be carried or rolled on its own wheels between rooms within a healthcare facility, while other transportable dental equipment is made to be folded and packed to carry over terrain which can be rugged and used in transient dental care settings which can have only limited shelter and utility services.

Transportable equipment that can be moved from one location to another while being carried by one or more persons is referred to as portable equipment. The term, portable equipment, applies to equipment that can be carried from room to room in a given facility or to remote parts of the world. This document focuses on portable dental equipment which is specifically designed and constructed to be transported between non-clinical environments and used by dental professionals to provide dental care in such settings, including temporary field clinics.

Such portable dental equipment for use in non-permanent healthcare environments enables dental professionals to provide a high standard of care to patients who do not have access to, or are not able to, travel to traditional health care facilities. Settings in which this equipment is commonly used include military field environments, humanitarian aid field clinics, public health outreach clinics, patient residences, long-term care facilities, prisons, schools and workplaces.

A number of trends in health care have driven increased utilization of portable dental equipment in non-permanent healthcare environments. Military forces use portable dental equipment in support of mobilized forces or for humanitarian outreach. A variety of government and non-government organizations are increasingly providing humanitarian dental care to underserved populations and populations affected by disasters. Civilian health care workers are also increasingly providing dental services to a growing population who are simply unable to visit traditional dental clinics due to age, disability, or income. Academic and research bodies regularly conduct dental education programs, particularly at external/off-site locations (including dentistry, dental hygiene, dental assisting).

The transport and end-use conditions for portable dental equipment used in non-permanent healthcare environments drive certain unique requirements which generally do not apply to portable, mobile or stationary dental equipment used in traditional dental clinics or hospitals. Because portable equipment used in non-permanent healthcare environments is intended to be moved between venues, and in some cases carried over rugged terrain or in inclement conditions, it needs to be designed and constructed to be safely transported by humans without damage, be efficiently assembled and disassembled, and deliver reliable service at the point of use. Special consideration is given to the austerity of the environment in which the equipment can be used and the availability and quality of utility supplies (such as electrical power, water, compressed air). In order for the equipment to be sufficiently portable and capable of operating in extreme conditions, certain requirements for dental equipment intended for use in traditional clinical settings may not be practical and is to be reconsidered for portable dental equipment for use in non-permanent healthcare environments. There can also be unique safety and infection control concerns to consider.

This document is one in a series with the objective of standardizing requirements for portable dental equipment for use in non-permanent healthcare environments.

12

## Dentistry — Portable dental equipment for use in nonpermanent healthcare environment —

# Part 1: General requirements

#### 1 Scope

This document specifies general requirements and test methods for portable dental equipment for use in non-permanent healthcare environments.

Portable dental equipment within the scope of this document includes portable dental units, portable patient chairs, portable operator's stools, portable operating lights, portable suction source equipment, portable air compressors and other portable dental equipment in instances where these devices are designed and constructed to be transported for use in non-permanent healthcare environments.

NOTE Particular requirements for specific types of portable dental equipment for use in non-permanent healthcare environments are specified in subsequent parts of this document.

This document does not apply to stationary dental equipment, wearable equipment (such as headlamps and loupes), mobile dental equipment or portable dental equipment that is not intended to be used in non-permanent healthcare environments or not designed to be disassembled, folded or packed for human transport between non-permanent healthcare environments. Also, requirements for stationary dental equipment that can be installed in a dental mobile medical facility (e.g. vehicular or containerized mobile dental clinic) are not considered in this document.

#### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, Dentistry — Vocabulary

ISO 4180:2019, Packaging — Complete, filled transport packages — General rules for the compilation of performance test schedules

ISO 21530, Dentistry — Materials used for dental equipment surfaces — Determination of resistance to chemical disinfectants

IEC 60529:1989/Amd 1:1999, Degrees of protection provided by enclosures (IP Code)

IEC 60601-1:2005/Amd 1:2012, Medical electrical equipment — Part 1: General requirements for basic safety and essential performance

IEC 62366-1, Medical devices — Part 1: Application of usability engineering to medical devices

IEC 80601-2-60, Medical electrical equipment — Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment