
**Wood-based panels — Determination
of formaldehyde release —**

**Part 3:
Gas analysis method**

*Panneaux à base de bois — Détermination du dégagement de
formaldéhyde —*

Partie 3: Méthode d'analyse de gaz



This document is a preview generated by EKO



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	1
6 Apparatus	2
6.1 Main composites of test apparatus (see Figure 1).....	2
6.2 Laboratory equipment.....	3
7 Sampling and preparation of test pieces	4
7.1 Preparation of test pieces.....	4
7.2 Selection of test pieces for factory production control.....	5
7.3 Selection of test pieces for other purposes.....	5
7.4 Preparation of test pieces in case of dispute.....	5
8 Procedure	5
8.1 Number of determinations.....	5
8.2 Determination of moisture content.....	5
8.3 Determination of formaldehyde release.....	5
8.3.1 General.....	5
8.3.2 Determination of formaldehyde release — Option 1.....	7
8.3.3 Determination of formaldehyde release — Option 2.....	7
8.3.4 Determination of formaldehyde release — Option 3.....	7
8.3.5 Determination of formaldehyde release — Option 4.....	7
8.4 Determination of formaldehyde content of the absorber solutions.....	7
8.4.1 General.....	7
8.4.2 Principle.....	8
8.4.3 Analytical procedure.....	8
8.4.4 Calibration curve.....	8
9 Expression of results	10
9.1 Gas analysis value.....	10
9.2 Calculation of results.....	10
9.2.1 Calculation of results of test pieces.....	10
9.2.2 Calculation of emission report value.....	11
9.3 Moisture content.....	11
10 Test report	11
Bibliography	13

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 89, *Wood based panels*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 112, *Wood based panels*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 12460-3:2015), which has been technically revised in order to improve the detection limit and the reproducibility of the method with regard to boards with low formaldehyde content.

The main changes compared to the previous edition are as follows:

- test period can be reduced from 4 h to 3 h;
- in [8.3](#) four different options to determine the formaldehyde release are introduced;
- conditioning procedure for decor finish foils specified in [7.1](#).

A list of all parts in the ISO 12460 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Wood-based panels — Determination of formaldehyde release —

Part 3: Gas analysis method

1 Scope

This document specifies a procedure for determination of accelerated formaldehyde release from uncoated and coated wood-based panels using the gas analysis method. The procedure is also suitable for the testing of other materials (e.g. edge bands, floor coverings, foams, foils, laminated wood products, veneered wood products, coated wood products).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16979, *Wood-based panels — Determination of moisture content*

ISO 16999, *Wood-based panels — Sampling and cutting of test pieces*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

4 Principle

A test piece of known surface area is placed in a closed chamber in which the temperature, humidity, airflow, and pressure are controlled to defined values. Formaldehyde released from the test pieces mixes with the air in the chamber. This air is continually drawn from the chamber and passes through gas wash bottles, containing water, which absorbs the released formaldehyde. At the end of the test, the formaldehyde concentration is determined photometrically or fluorimetrically. The formaldehyde release is calculated from this concentration, the sampling time, and the exposed area of the test pieces and is expressed in milligrams per square meter and hour (mg/m²h).

5 Reagents

Reagents of recognized analytical purity and distilled or demineralised water (referred throughout the following text as distilled water) shall be used for the analysis.

5.1 4 ml acetylacetone solution are added to a 1 000 ml volumetric flask and made up to the mark with distilled water.