

ICS 13.040.20

English Version

**Ambient air - Application of EN 16909 for the  
determination of elemental carbon (EC) and organic  
carbon (OC) in PM10 and PMcoarse**

Air ambiant - Application de la norme EN 16909 pour  
le dosage du carbone élémentaire (EC) et du carbone  
organique (OC) dans les fractions PM10 et PMgrossière

Außenluft - Anwendung der EN 16909 zur  
Bestimmung von elementarem Kohlenstoff (EC) und  
organischem Kohlenstoff (OC) in PM10 und PMcoarse

This Technical Report was approved by CEN on 9 November 2020. It has been drawn up by the Technical Committee CEN/TC 264.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

Contents	Page
European foreword.....	3
Introduction .....	4
1 Scope .....	5
2 Normative references.....	5
3 Terms and definitions .....	5
4 Symbols and abbreviations .....	6
5 Principle .....	6
6 Previous studies on interferences from inorganic components.....	7
6.1 General.....	7
6.2 Carbonate carbon.....	7
6.3 Metal oxides .....	8
6.4 Inorganic salts .....	8
7 Information from the data obtained during the EN 16909 field validation campaigns .....	9
8 Procedures for evaluating the applicability of EN 16909 to PM <sub>10</sub> and PM <sub>coarse</sub> .....	14
8.1 General.....	14
8.2 Materials, instruments and analysis.....	15
8.3 Sampling.....	15
8.4 Procedures.....	15
8.4.1 General.....	15
8.4.2 Comparison of OC and EC concentrations in different PM size fractions.....	15
9 Assessment of the effect of coarse PM constituents on OC and EC determination .....	16
9.1 Carbonate carbon.....	16
9.2 Analytical artefacts in PM <sub>2,5</sub> filter samples spiked with PM <sub>coarse</sub> constituents that contain no EC or OC .....	16
9.2.1 Spiking material preparation .....	16
9.2.2 Test sample preparation and measurements.....	16
9.2.3 Test evaluation.....	17
9.3 Analytical artefacts in PM <sub>coarse</sub> filters spiked with known amounts of OC and/or EC.....	17
9.3.1 Spiking material preparation .....	17
9.3.2 Test sample preparation and measurements.....	17
9.3.3 Test evaluation.....	17
Annex A (informative) Details of PM <sub>2,5</sub> and PM <sub>10</sub> filters included in the laboratory comparison exercise .....	18
Annex B (informative) Estimation of the uncertainty of EC <sub>coarse</sub> and OC <sub>coarse</sub> .....	19
Bibliography .....	20

## European foreword

This document (CEN/TR 17554:2020) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document is a preview generated by EVS

## Introduction

The standard method EN 16909 provides guidance for the determination of organic carbon (OC) and elemental carbon (EC) in airborne particulate matter deposited on filters. It has been developed following the requirement for the EU member states to measure OC and EC in the PM<sub>2,5</sub> size fraction (less than 2,5 µm in aerodynamic diameter) at background sites [5]. EN 16909 standard states: “The same analysis method may also be used for smaller size fractions than PM<sub>2,5</sub>. Any possible additional artefacts for larger particles, e.g. pyrolysis or higher concentrations of carbonates, should be assessed.”

## 1 Scope

This document describes procedures to assess the applicability of the standard method EN 16909 (determination of OC and EC deposited on filters) to particle size fractions up to 10 µm in aerodynamic diameter (50 % cut off).

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 16909, *Ambient air - Measurement of elemental carbon (EC) and organic carbon (OC) collected on filters*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 16909 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 3.1

#### **PM<sub>x</sub>**

particulate matter suspended in air which is small enough to pass through a size-selective inlet with a 50 % efficiency cut-off at x µm aerodynamic diameter

[SOURCE: EN 12341:2014 [1], definition 3.1.14]

### 3.2

#### **PM<sub>coarse</sub> fraction**

the PM<sub>10</sub> fraction excluding the PM<sub>2,5</sub> fraction

### 3.3

#### **OC<sub>x</sub>**

organic carbon component of PM<sub>x</sub>

### 3.4

#### **EC<sub>x</sub>**

elemental carbon component of PM<sub>x</sub>

### 3.5

#### **PC<sub>x</sub>**

pyrolytic carbon component of PM<sub>x</sub>

### 3.6

#### **TC<sub>x</sub>**

Total carbon component of PM<sub>x</sub>