

---

---

**Water quality — Determination of  
cyclic volatile methylsiloxanes in  
water —**

**Part 2:  
Method using liquid-liquid extraction  
with gas chromatography-mass  
spectrometry (GC-MS)**

*Qualité de l'eau — Détermination de méthylsiloxanes cycliques  
volatiles dans l'eau —*

*Partie 2: Méthode par extraction liquide-liquide avec  
chromatographie en phase gazeuse-spectrométrie de masse (CG-SM)*



This document is a preview generated by EKO



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>v</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>2</b>
<b>4 Principle</b>	<b>2</b>
4.1 Principle of preservation and extraction	2
<b>5 Interferences</b>	<b>2</b>
5.1 Interferences with sampling and processing	2
5.2 Interferences with GC-MS	2
5.3 Interferences determination	3
<b>6 Reagents</b>	<b>3</b>
<b>7 Apparatus</b>	<b>5</b>
<b>8 Method detection limits</b>	<b>6</b>
<b>9 Quality control</b>	<b>6</b>
<b>10 Sampling and storage</b>	<b>6</b>
10.1 Sampling preparation	6
10.2 Sample collection	6
<b>11 Extraction and analysis</b>	<b>7</b>
11.1 Extraction	7
11.2 GC conditions and operation	7
<b>12 Calibration</b>	<b>7</b>
12.1 General requirements	7
12.2 Calibration calculations	8
12.3 Concentration calculations	8
12.4 Calculation of results	9
12.5 Treatment of results lying outside the calibration range	9
<b>13 Expression of results</b>	<b>10</b>
<b>14 Test report</b>	<b>10</b>
<b>Annex A (informative) GC-MS conditions</b>	<b>11</b>
<b>Annex B (informative) Method detection limit and limit of quantification</b>	<b>13</b>
<b>Annex C (informative) Example quality control samples</b>	<b>14</b>
<b>Annex D (informative) Performance data</b>	<b>15</b>
<b>Bibliography</b>	<b>16</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

A list of all parts in the ISO 20596 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The method described in this document uses low density polyethylene to prevent volatilization of samples during transit and storage. The samples are processed using a liquid-liquid extraction into a non-polar solvent with subsequent injection onto a gas chromatograph-mass spectrometer for separation and quantitation.



# Water quality — Determination of cyclic volatile methylsiloxanes in water —

## Part 2:

## Method using liquid-liquid extraction with gas chromatography-mass spectrometry (GC-MS)

**WARNING** — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure neutralization and proper disposal of waste solutions.

**IMPORTANT** — It is absolutely essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

### 1 Scope

This document specifies a method for the determination of certain cyclic volatile methylsiloxanes (cVMS) in environmental water samples with low density polyethylene (LDPE) as a preservative and subsequent liquid-liquid extraction with hexane containing  $^{13}\text{C}$ -labeled cVMS as internal standards. The extract is then analysed by gas chromatography-mass spectrometry (GC-MS).

**NOTE** Using the  $^{13}\text{C}$ -labeled, chemically identical substances as internal standards with the same properties as the corresponding analytes, minimizes possible substance-specific discrimination in calibrations. Since these substances are least soluble in water, they are introduced via the extraction solvent hexane into the system.

This document is applicable to the measurement of the following cVMS in rivers, streams, and waste water (influent and effluent):

**Table 1 — Analytes determined by this method**

Analyte	Formula	Abbreviation	CAS <sup>a</sup> -RN
Octamethylcyclotetrasiloxane	$\text{C}_8\text{H}_{24}\text{O}_4\text{Si}_4$	D4	556-67-2
Decamethylcyclopentasiloxane	$\text{C}_{10}\text{H}_{30}\text{O}_5\text{Si}_5$	D5	541-02-6
Dodecamethylcyclohexasiloxane	$\text{C}_{12}\text{H}_{36}\text{O}_6\text{Si}_6$	D6	540-97-6

<sup>a</sup> CAS-RN Chemical Abstracts Services Registration Number

This method can be used to determine cVMS from 0,1 µg/l to 250 µg/l. In well controlled laboratory environments, where contamination is minimized, the lower end of the application range can be diminished by a factor of up to 10.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-4, *Water quality — Sampling — Part 4: Guidance on sampling from lakes, natural and man-made*

ISO 5667-6, *Water quality — Sampling — Part 6: Guidance on sampling of rivers and streams*

ISO 5667-10, *Water quality — Sampling — Part 10: Guidance on sampling of waste waters*

ISO 5667-14, *Water quality — Sampling — Part 14: Guidance on quality assurance and quality control of environmental water sampling and handling*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

ISO/TS 13530, *Water quality — Guidance on analytical quality control for chemical and physicochemical water analysis*

### 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 4 Principle

#### 4.1 Principle of preservation and extraction

The siloxane compounds (D4), (D5), and (D6) are relatively volatile and have low solubility in water thus making accurate quantification in aqueous matrices challenging. Low density polyethylene (LDPE) is added to samples to prevent volatilization of the cVMS through a partial physical barrier between the water and headspace and a matrix to which the cVMS may adsorb. Hexane is then used to extract the dissolved and sorbed fractions of cVMS. The hexane extracts are then analysed by GC-MS ([Annex A](#)).

### 5 Interferences

#### 5.1 Interferences with sampling and processing

Silicones, including D4, D5, and D6 are widely used in industrial applications as well as personal care products such as conditioner, hand lotion, sunscreens, and cosmetics (not all inclusive). Persons involved with the collection and analysis of samples should refrain from using siloxane containing products to limit potential contamination of the sample.

Additionally, the users should refrain from using collection devices, sampling containers, laboratory equipment or consumables which may contain silicones/siloxanes. Sample contact surfaces should be suitably rinsed with acetone or hexane and subsequently dried in a clean area of the laboratory to remove any contamination.

#### 5.2 Interferences with GC-MS

Silicones are also commonly found in parts and consumables associated with gas chromatography including septa for the vials and inlet. Commonly used types of GC columns are polydimethylsiloxane based which when exposed to moisture or when heated may generate cVMS and in such a way can contribute to background. Thus, the use of non-polydimethylsiloxane-based GC columns is highly recommended, in particular when analysing sub-ppb concentrations. Autosampler vial septa should be silicone free or at a minimum coated with polytetrafluoroethylene on the side exposed to the sample.