

Instruments for use in association with non-active surgical implants - General requirements (ISO 16061:2021)

This document is a preview generated by EVS

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 16061:2021 sisaldab Euroopa standardi EN ISO 16061:2021 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 16061:2021 consists of the English text of the European standard EN ISO 16061:2021.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 31.03.2021.	Date of Availability of the European standard is 31.03.2021.
Standard on kättesaadav Eesti Standardimis-ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 11.040.40, 11.040.99

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation:

Homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

EUROPEAN STANDARD

**EN ISO 16061**

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2021

ICS 11.040.40; 11.040.99

Supersedes EN ISO 16061:2015

English Version

## Instruments for use in association with non-active surgical implants - General requirements (ISO 16061:2021)

Instruments à utiliser en association avec les implants chirurgicaux non actifs - Exigences générales (ISO 16061:2021)

Instrumente, die in Verbindung mit nichtaktiven chirurgischen Implantaten verwendet werden - Allgemeine Anforderungen (ISO 16061:2021)

This European Standard was approved by CEN on 5 January 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## European foreword

This document (EN ISO 16061:2021) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" in collaboration with Technical Committee CEN/TC 285 "Non-active surgical implants" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2021, and conflicting national standards shall be withdrawn at the latest by September 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 16061:2015.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Endorsement notice

The text of ISO 16061:2021 has been approved by CEN as EN ISO 16061:2021 without any modification.

# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Intended performance</b> .....	<b>2</b>
<b>5 Design attributes</b> .....	<b>3</b>
<b>6 Selection of materials</b> .....	<b>4</b>
<b>7 Design evaluation</b> .....	<b>4</b>
7.1 General.....	4
7.2 Pre-clinical evaluation.....	4
7.3 Clinical evaluation.....	5
7.4 Post-market surveillance.....	5
<b>8 Manufacture</b> .....	<b>6</b>
<b>9 Sterilization</b> .....	<b>6</b>
9.1 Instruments supplied sterile.....	6
9.2 Instruments supplied non-sterile.....	6
9.3 Instruments that are resterilizable.....	6
<b>10 Packaging</b> .....	<b>6</b>
10.1 Protection from damage in transport, storage and handling.....	6
10.2 Maintenance of sterility in transport, storage and handling.....	7
<b>11 Information supplied by the manufacturer</b> .....	<b>7</b>
11.1 General.....	7
11.2 Marking on instruments.....	8
11.3 Label.....	8
11.4 Instructions for use.....	9
11.5 Additional information for instruments with a measuring function.....	10
<b>Annex A (informative) Examples of typical applications of instruments to be used in association with non-active surgical implants and materials found acceptable for instrument manufacture</b> .....	<b>11</b>
<b>Bibliography</b> .....	<b>13</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 285, *Non-active surgical implants*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fourth edition cancels and replaces the third edition (ISO 16061:2015), which has been technically revised. The main changes compared to the previous edition are as follows:

- A requirement to include intended purpose has been added in the list of items to be included when establishing the intended performance of the instrument.
- The list of design attributes in [Clause 5](#) has been reorganized and several new attributes have added to the list.
- The selection of materials to be used in the instrument has been based on a risk analysis and the clause now includes a list of the minimum factors to be considered in the risk analysis.
- The requirement for pre-clinical evaluation has been expanded and includes the requirement for testing and biological evaluation of the final instrument.
- A clinical evaluation of the instrument has been added as a requirement in all cases. However, if the pre-clinical evaluation demonstrates the safety and intended performance of the instrument in the conditions of intended use, the results of the pre-clinical evaluation will satisfy the requirement for the clinical evaluation.
- A new requirement for post-market surveillance has been added to [Clause 7](#).
- The requirements in [Clause 11](#) have been reorganized and clarified to reflect current practice and to reference ISO 17664:2017, Clause 6 for instructions for applicable processing step (i.e. cleaning, disinfection, drying, packaging, and sterilization) that need to be carried out by someone other than the manufacturer.

- [Annex A](#) has been simplified to provide more consistent guidance on selection of material using a risk-based approach. The stainless-steel grade material characteristic tables have been removed.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

This document provides a method of addressing the fundamental principles outlined in ISO/TR 14283 as they apply to instruments to be used in association with non-active surgical implants. It also provides a method that can be used to demonstrate compliance with applicable regulatory requirements relevant to the general safety and performance of medical devices as they apply to instruments used in association with non-active surgical implants.

There are three levels of standards dealing with instruments to be used in association with non-active surgical implants. They are as follows, with level 1 being the highest.

- Level 1: general requirements for instruments to be used in association with non-active surgical implants.
- Level 2: particular requirements for families of instruments to be used in association with non-active surgical implants.
- Level 3: specific requirements for types of instruments to be used in association with non-active surgical implants.

Level 1 standards include this document which contains requirements that apply to all instruments to be used in association with non-active surgical implants, ISO 14630, which contains the requirement for non-active surgical implants and ISO 14708-1, which contains requirements for active implants. They also anticipate that there are additional requirements in the level 2 and level 3 standards.

Level 2 standards apply to a more restricted set or family of instruments, such as those designed for use with non-active surgical implants used in neurosurgery, cardiovascular surgery, or joint replacement.

Level 3 standards apply to specific types of instruments within a family of instruments used in association with non-active surgical implants, such as hip joints or arterial stents.

To address all requirements for a specific instrument, it is advisable that the standard of the lowest available level be consulted first.

Compliance with a level 3 standard is intended to imply compliance with the applicable level 2 standards, if available, and with the applicable level 1 standard.

**NOTE** The requirements in this document correspond to international consensus. Individual or national standards or regulatory bodies can prescribe other requirements.



# Instruments for use in association with non-active surgical implants — General requirements

## 1 Scope

This document specifies the general requirements for instruments to be used in association with non-active surgical implants. These requirements apply to instruments when they are manufactured and when they are supplied after refurbishment.

NOTE In this document, unless otherwise specified, the term “instrument” refers to an instrument for use in association with non-active surgical implants.

This document also applies to instruments which can be connected to power-driven systems, but it does not apply to the power-driven systems themselves.

With regard to safety, this document gives the requirements for intended performance, design attributes, materials, design evaluation, manufacture, sterilization, packaging, and information supplied by the instrument manufacturer, hereafter referred to as the manufacturer.

This document is not applicable to instruments associated with dental implants, transendodontic and transradicular implants and ophthalmic implants.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8601-1, *Date and time — Representations for information interchange — Part 1: Basic rules*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 11135, *Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-2, *Sterilization of health care products — Radiation — Part 2: Establishing the sterilization dose*

ISO 11137-3, *Sterilization of health care products — Radiation — Part 3: Guidance on dosimetric aspects of development, validation and routine control*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 17664, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices*

ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO 25424, *Sterilization of health care products — Low temperature steam and formaldehyde — Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 80000-1, *Quantities and units — Part 1: General*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 14971 and the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

#### 3.1 associated implant

specific *non-active surgical implant* (3.4) associated with a specific *instrument* (3.3) intended to be used during a surgical procedure

#### 3.2 final instrument

*instrument* (3.3) that has been subjected to all manufacturing processes for the “to be marketed” *instrument* including packaging and if applicable, sterilization

Note 1 to entry: Derived from the definition of “final product” in ISO 10993-1:2018, 3.8 to reinforce that the instrument to be evaluated is the final product referred to in ISO 10993-1.

#### 3.3 instrument

non-active medical device intended for use during surgical procedures related to specific *non-active surgical implants* (3.4)

Note 1 to entry: Examples of typical applications of instruments to be used in association with non-active surgical implants are presented in A.1.

#### 3.4 non-active surgical implant

surgical implant, the operation of which does not depend on a source of electrical energy or any source of power other than that directly generated by the human body or gravity

[SOURCE: ISO 14630:2012, 3.6, modified — the second term “implant” has been deleted.]

### 4 Intended performance

The intended performance of an instrument shall be described and documented by addressing the following, with particular regard to safety:

- a) intended purpose;
- b) functional characteristics;
- c) intended conditions of use.