
**Soil, treated biowaste and sludge –
Determination of pH**

Sols, biodéchets traités et boues — Détermination du pH

This document is a preview generated by EVS



This document is a preview generated by ELS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	1
6 Apparatus	2
7 Laboratory sample	3
8 Procedure	3
8.1 Preparation of the suspension.....	3
8.2 Calibration of the pH-meter.....	3
8.3 Measurement of the pH.....	4
9 Precision	4
10 Test report	4
Annex A (informative) Repeatability and reproducibility data	5
Bibliography	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical and physical characterization*.

This third edition cancels and replaces the second edition (ISO 10390:2005), which has been technically revised. The main changes compared to the previous edition are as follows:

- The content of ISO 10390:2005 and EN 15933:2012 were merged;
- The scope was widened to include sludge and treated biowaste;
- Additional validation data for soil, sludge and treated biowaste were included;
- The text was editorially revised.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Soil, treated biowaste and sludge – Determination of pH

1 Scope

This document specifies an instrumental method for the routine determination of pH within the range pH 2 to pH 12 using a glass electrode in a 1:5 (volume fraction) suspension of soil, sludge and treated biowaste in either water (pH in H₂O), in 1 mol/l potassium chloride solution (pH in KCl) or in 0,01 mol/l calcium chloride solution (pH in CaCl₂).

This document is applicable to all types of air-dried soil and treated biowaste samples.

NOTE For example, pretreated in accordance with ISO 11464 or EN 16179 or EN 15002.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

4 Principle

A suspension of a test portion is made up in five times its volume with one of the following solutions:

- water;
- a solution of potassium chloride (KCl) in water, $c = 1 \text{ mol/l}$;
- a solution of calcium chloride (CaCl₂) in water, $c = 0,01 \text{ mol/l}$.

The pH of the suspension is measured using a pH-meter.

NOTE 1 To make the procedure generally applicable to all types of soil, treated biowaste samples and sludges except liquid sludge, a volume-to-volume shaking ratio is chosen because all test samples can be treated in the same way. If a mass-to-volume ratio were chosen, the weighed amount of test sample would have to be adapted for e.g. soils with a low density, to enable the preparation of the suspension. For the purpose of this document, it is sufficiently accurate to measure the required volume of test portion with a measuring spoon.

NOTE 2 In samples with a high content of charged particles (e.g. organic matter, clay) the suspension effect can modify the potential difference between the electrodes, and thereby have an influence on the recorded pH value. This problem is minimized by gentle stirring of the suspension. For calcareous material, carbon dioxide can be absorbed by the suspension, which makes it difficult to reach an equilibrium value. Other sources of error are associated with materials containing sulfidic minerals or volatile acids.

5 Reagents

Use only reagents of recognized analytical grade.