

MEDITSIINISEADMED. TOOTJA ESITATAV TEAVE

Medical devices - Information to be supplied by the manufacturer (ISO 20417:2021, Corrected version 2021-12)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 20417:2021 sisaldab Euroopa standardi EN ISO 20417:2021 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 20417:2021 consists of the English text of the European standard EN ISO 20417:2021.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 05.05.2021.	Date of Availability of the European standard is 05.05.2021.
Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.01

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about standards copyright protection, please contact the Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 20417

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2021

ICS 11.040.01

Supersedes EN 1041:2008+A1:2013

English version

**Medical devices - Information to be supplied by the
manufacturer (ISO 20417:2021, Corrected version
2021-12)**

Dispositifs médicaux - Informations à fournir par le
fabricant (ISO 20417:2021, Version corrigée 2021-12)

Medizinprodukte - Anforderungen an vom Hersteller
bereitgestellte Informationen (ISO 20417:2021,
korrigierte Fassung 2021-12)

This European Standard was approved by CEN on 30 June 2020.

This European Standard was corrected and reissued by the CEN-CENELEC Management Centre on 22 December 2021.

CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN and CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN and CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



**CEN-CENELEC Management Centre:
Rue de la Science 23, B-1040 Brussels**

European foreword

This document (EN ISO 20417:2021) has been prepared by Technical Committee ISO/TC 210 "Quality management and corresponding general aspects for medical devices" in collaboration with Technical Committee CEN/CLC/JTC 3 "Quality management and corresponding general aspects for medical devices" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2021, and conflicting national standards shall be withdrawn at the latest by November 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 1041:2008+A1:2013.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20417:2021, Corrected version 2021-12 has been approved by CEN as EN ISO 20417:2021 without any modification.

Contents

Page

Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 General considerations.....	9
5 Information elements to be established.....	10
5.1 Units of measurement.....	10
5.2 Graphical information.....	10
5.3 Language and country identifiers.....	11
5.3.1 Language identifiers.....	11
5.3.2 Country identifiers.....	11
5.4 Dates.....	11
5.5 Full address.....	12
5.6 <i>Commercial product name</i>	12
5.7 <i>Model number</i>	12
5.8 <i>Catalogue number</i>	12
5.9 Production controls.....	12
5.10 Unique device identifier.....	13
5.11 Types of use/reuse.....	13
5.12 <i>Sterile</i>	13
6 Requirements for accompanying information.....	13
6.1 Requirements for information to be supplied on the <i>label</i>	13
6.1.1 Minimum requirements for the <i>label</i>	13
6.1.2 Identification of the <i>manufacturer</i>	14
6.1.3 Identification of the <i>medical device</i> or <i>accessory</i>	15
6.1.4 Other <i>label</i> requirements.....	17
6.1.5 Consult <i>instructions for use</i>	18
6.1.6 <i>Safety signs</i>	19
6.2 Identification requirements for detachable components of a <i>medical device</i> or <i>accessory</i>	20
6.3 Legibility of the <i>label</i>	20
6.4 Durability of <i>markings</i>	20
6.5 Information to be provided on the packaging.....	21
6.5.1 General information.....	21
6.5.2 Packaging for the <i>lay user</i>	22
6.5.3 Special conditions indicated on the packaging.....	23
6.6 Requirements for information in the <i>instructions for use</i> and <i>technical description</i>	24
6.6.1 General.....	24
6.6.2 Requirements for <i>instructions for use</i>	25
6.6.3 Additional requirements for the <i>instructions for use</i> for a <i>lay user</i>	30
6.6.4 Requirements for <i>technical description</i>	30
6.6.5 Requirements for <i>e-documentation</i>	33
7 Other information that is required to be supplied with the <i>medical device</i> or <i>accessory</i>.....	33
7.1 <i>Importer</i>	33
7.2 <i>Distributor</i>	33
7.3 Repackaging.....	34
7.4 Translation.....	34
7.5 Regulatory identification.....	35
Annex A (informative) Particular guidance and rationale.....	36

Annex B (informative) Example test method for assessing <i>clearly legible</i> requirements	39
Annex C (informative) Example test method for assessing durability	40
Annex D (informative) Cross reference between the document and the requirements considered	41
Annex E (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances	53
Annex F (informative) Reference to the <i>essential principles</i>	57
Annex G (informative) Reference to the general safety and performance requirements for <i>medical devices</i>	61
Annex H (informative) Reference to the general safety and performance requirements for <i>IVD medical devices</i>	65
Annex I (informative) Terminology — Alphabetized index of defined terms	69
Bibliography	71

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices* in collaboration with the European Committee for Standardization (CEN/CLC) Technical Committee CEN/CLC JTC 3, *Quality management and corresponding general aspects for medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO 20417:2021 incorporates the following corrections:

In 6.1.3. f):

If the *label* includes *symbols* or safety-related colours, they shall be explained in the *label*.

has been corrected to:

If the *label* includes *symbols* or safety-related colours, they shall be explained in the *instructions for use*.

Introduction

This document provides the requirements for the identification and *labels* on a *medical device* or *accessory*, the packaging, *marking* of a *medical device* or *accessory*, and *accompanying information*. The aim of this document is to serve as a central source of these common, generally applicable requirements, allowing each specific *product standard* or *group standard* to focus more concisely on the unique requirements for a *specific medical device* or group of *medical devices*.

The requirements of a *medical device product standard* or a *group standard* can make use of these general requirements. Where there is a conflict and a *product standard* or a *group standard* exists, this document should not be used separately. Specific requirements of *medical device product standards* or *group standards* take precedence over requirements of this document. Unless specified otherwise within a *product standard* or a *group standard*, the general requirements of this document apply.

Some *authorities having jurisdiction* have requirements that can differ from the requirements of this document.

This document has been prepared in consideration of:

- the application of *Essential Principles of Safety and Performance of Medical Devices and IVD Medical Devices*, IMDRF/GRRP WG/N47:2018^[3] on the *information supplied by the manufacturer* of a *medical device* (see [Annex D](#));
- the application of *Labelling Principles for Medical Devices and IVD Medical Devices*, IMDRF/GRRP WG/N52:2019^[4] on the *information supplied by the manufacturer* of a *medical device* (see [Annex E](#));
- the application of the *essential principles of safety and performance* on the *information supplied by the manufacturer* of a *medical device* according to ISO 16142-1:2016 (see [Annex F](#));
- the application of the *essential principles of safety and performance* on the *information supplied by the manufacturer* of an *IVD medical device* according to ISO 16142-2:2017 (see [Annex F](#));
- the general safety and performance requirements for the *information supplied by the manufacturer* of a *medical device* according to regulation (EU) 2017/745^[5] (see [Annex G](#)); and
- the general safety and performance requirements for the *information supplied by the manufacturer* of a *medical device* according to regulation (EU) 2017/746^[6] (see [Annex H](#)).

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

Requirements in this document have been decomposed so that each requirement is uniquely delineated. This is done to support automated requirements tracking.

Medical devices — Information to be supplied by the manufacturer

1 Scope

NOTE 1 There is guidance or rationale for this Clause contained in Clause A.2.

This document specifies the requirements for *information supplied by the manufacturer* for a *medical device* or by the *manufacturer* for an *accessory*, as defined in 3.1. This document includes the generally applicable requirements for identification and *labels* on a *medical device* or *accessory*, the packaging, *marking* of a *medical device* or *accessory*, and *accompanying information*. This document does not specify the means by which the information is to be supplied.

NOTE 2 Some *authorities having jurisdiction* impose different requirements for the identification, *marking* and documentation of a *medical device* or *accessory*.

Specific requirements of *medical device product standards* or *group standards* take precedence over requirements of this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3166-1, *Codes for the representation of names of countries and their subdivisions — Part 1: Country codes*

ISO 3864-1:2011, *Graphical symbols — Safety colours and safety signs — Part 1: Design principles for safety signs and safety markings*

ISO 7000, *Graphical symbols for use on equipment — Registered symbols*

ISO 7010:2019, *Graphical symbols — Safety colours and safety signs — Registered safety signs*

ISO 8601-1, *Date and time — Representations for information interchange — Part 1: Basic rules*

ISO 13485:2016, *Medical devices — Quality management systems — Requirements for regulatory purposes*

ISO 14971:2019, *Medical devices — Application of risk management to medical devices*

ISO 15223-1:—¹⁾, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 16142-1:2016, *Medical devices — Recognized essential principles of safety and performance of medical devices — Part 1: General essential principles and additional specific essential principles for all non-IVD medical devices and guidance on the selection of standards*

ISO 16142-2:2017, *Medical devices — Recognized essential principles of safety and performance of medical devices — Part 2: General essential principles and additional specific essential principles for all IVD medical devices and guidance on the selection of standards*

IEC 60417, (*database*), *Graphical symbols for use on equipment*

1) Under preparation. Stage at the time of publication: ISO/FDIS 15223-1:2021.

IEC 62366-1:2015+AMD1:2020, *Medical devices — Part 1: Application of the usability engineering process to medical devices*

ISO 80000-1, *Quantities and units — Part 1: General*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 13485:2016, ISO 14971:2019, ISO 16142-1:2016, ISO 16142-2:2017 and IEC 62366-1:2015+AMD1:2020 as specified in [Annex I](#) and the following definitions apply. ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

NOTE An alphabetized index of defined terms used in this document is found in [Annex I](#).

3.1 **accessory**

item, intended specifically by its *manufacturer*, to be used together with one or more *medical devices* to specifically enable or assist those *medical devices* to be used in accordance with their *intended use*

Note 1 to entry: An *accessory* is typically a consumable or separate item for use with one or more *medical devices*.

Note 2 to entry: Note 2 to entry: Some *authorities having jurisdiction* consider an *accessory* to be a *medical device*.

Note 3 to entry: Some *authorities having jurisdiction* have a different definition of *accessory*.

3.2 **accompanying information**

information accompanying or *marked* on a *medical device* or *accessory* (3.1) for the *user* or those accountable for the installation, use, *processing*, maintenance, decommissioning and disposal of the *medical device* or *accessory*, particularly regarding safe use

Note 1 to entry: The *accompanying information* shall be regarded as part of the *medical device* or *accessory*.

Note 2 to entry: The *accompanying information* can consist of the *label*, *marking*, *instructions for use*, *technical description*, installation manual, quick reference guide, etc.

Note 3 to entry: *Accompanying information* is not necessarily a written or printed document but could involve auditory, visual, or tactile materials and multiple media types (e.g., CD/DVD-ROM, USB stick, website).

Note 4 to entry: See [Figure 1](#).