
**Molecular in vitro diagnostic
examinations — Specifications for
pre-examination processes for frozen
tissue —**

**Part 3:
Isolated DNA**

*Analyses de diagnostic moléculaire in vitro — Spécifications relatives
aux processus préanalytiques pour les tissus congelés —*

Partie 3: ADN extrait



This document is a preview generated by EKO



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

| | Page |
|---|-----------|
| Foreword | iv |
| Introduction | v |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 General considerations | 5 |
| 5 Outside the laboratory | 6 |
| 5.1 Specimen collection..... | 6 |
| 5.1.1 General..... | 6 |
| 5.1.2 Information about the patient/specimen donor..... | 6 |
| 5.1.3 Information about the specimen..... | 6 |
| 5.1.4 Specimen processing..... | 6 |
| 5.2 Fresh tissue transport requirements..... | 7 |
| 5.2.1 General..... | 7 |
| 5.2.2 Preparations for the transport..... | 7 |
| 5.2.3 During transport..... | 8 |
| 6 Inside the laboratory | 8 |
| 6.1 Information about the reception of the specimen..... | 8 |
| 6.2 Evaluation of the pathology of the specimen and selection of the sample(s)..... | 8 |
| 6.3 Freezing of the specimen or sample(s)..... | 9 |
| 6.4 Storage requirements..... | 11 |
| 6.5 DNA isolation..... | 11 |
| 6.5.1 General..... | 11 |
| 6.5.2 Using commercial kits..... | 12 |
| 6.5.3 Using laboratory developed protocols..... | 12 |
| 6.6 Quantity and quality assessment of isolated DNA..... | 12 |
| 6.7 Storage of isolated DNA..... | 13 |
| Bibliography | 14 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 20184 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Molecular in vitro diagnostics, including molecular pathology, has enabled significant progress in medicine. Further progress is expected with new technologies analysing nucleic acids, proteins, and metabolites in human tissues and body fluids. However, integrity and profile of these molecules can change during specimen collection, transport, storage, and processing. As a consequence the outcome from diagnostics or research can become unreliable or even impossible because the subsequent examination assay might not determine the real situation in the patient but instead, an artificial profile which is generated during the pre-examination process.

DNA integrity in tissues can change during processing and storage. Modifications of the DNA molecules can impact the validity and reliability of the examination test results. It is essential to take special measures to minimize the described DNA changes and modifications for subsequent examination.

Therefore, a standardization of the entire process from specimen collection to DNA examination is needed. Studies have been undertaken to determine the important influencing factors. This document draws upon such work to codify and standardize the steps for frozen tissue with regard to DNA examination in what is referred to as the pre-examination phase.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for frozen tissue —

Part 3: Isolated DNA

1 Scope

This document specifies requirements and gives recommendations for the handling, storage, processing, and documentation of frozen tissue specimens intended for DNA examination during the pre-examination phase before a molecular examination is performed.

This document is applicable to molecular in vitro diagnostic examinations including laboratory developed tests performed by medical laboratories and molecular pathology laboratories that evaluate DNA isolated from frozen tissue. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions and commercial organizations performing biomedical research, and regulatory authorities.

Tissues that have undergone chemical stabilization pre-treatment before freezing are not covered in this document.

NOTE International, national, or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189, *Medical laboratories — Requirements for quality and competence*

ISO 15190, *Medical laboratories — Requirements for safety*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15189 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

aliquot

portion of a larger amount of homogenous material, assumed to be taken with negligible sampling error

Note 1 to entry: The term is usually applied to fluids. Solid tissues are heterogeneous and therefore cannot be aliquoted.