

Acoustics - Hearing aid fitting management (HAFM)
(ISO 21388:2020)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 21388:2021 sisaldab Euroopa standardi EN ISO 21388:2021 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 21388:2021 consists of the English text of the European standard EN ISO 21388:2021.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 21.07.2021.	Date of Availability of the European standard is 21.07.2021.
Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.020.10, 11.180.15, 17.140.01

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation.

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Acoustics - Hearing aid fitting management (HAFM) (ISO
21388:2020)

Acoustique - Gestion des appareils de correction
auditive (ISO 21388:2020)

Akustik - Hörgeräteanpassungsmanagement (HAFM)
(ISO 21388:2020)

This European Standard was approved by CEN on 11 July 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of ISO 21388:2020 has been prepared by Technical Committee ISO/TC 43 "Acoustics" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 21388:2021 by Technical Committee CEN/TC 211 "Acoustics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2022, and conflicting national standards shall be withdrawn at the latest by January 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 15927:2010.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 21388:2020 has been approved by CEN as EN ISO 21388:2021 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Service preconditions	4
4.1 General	4
4.2 Educational requirements	4
4.2.1 General	4
4.2.2 Requirements for hearing aid professionals	4
4.2.3 Maintenance of competencies and skills of hearing aid professionals	5
4.3 Facility requirements	5
4.3.1 General	5
4.3.2 Room requirements	5
4.4 Equipment requirements	5
4.4.1 General	5
4.4.2 Audiometric equipment	6
4.4.3 Equipment for otoscopy and earmould impressions	6
4.4.4 Hearing aid programming equipment	6
4.4.5 Electroacoustic measurement equipment	6
4.4.6 Maintenance tools	6
4.4.7 Demonstration samples	7
4.5 Ethical requirements	7
4.5.1 General	7
4.5.2 Professional competence	7
4.5.3 Relationship with clients	7
4.5.4 Conflict of interest	7
4.5.5 Relationship with medical and other health practitioners	7
4.5.6 Relationship with colleagues	8
4.5.7 Advertising	8
5 General stages of HAFM	8
5.1 General	8
5.2 Client profile	9
5.2.1 General	9
5.2.2 General assessment	9
5.2.3 Audiological assessment	10
5.2.4 Medical referral	11
5.3 Counselling	11
5.3.1 General	11
5.3.2 Selection of hearing aid system	11
5.4 Hearing aid fitting	11
5.4.1 Ear coupling elements	11
5.4.2 Pre-setting of hearing aids	12
5.4.3 Setting and fine-tuning of hearing aids	12
5.5 Verification and validation	12
5.6 Post-fitting counselling	13
5.7 Follow-up	13
6 Quality of service	14
6.1 General	14
6.2 Documentation	14
6.3 Client evaluation of services	15
6.4 Customer complaint handling	15

6.5	Corrective actions.....	15
Annex A	(informative) Minimum competencies of the hearing aid professional (HAP).....	16
Annex B	(informative) Recommendation for organisation of education and training for hearing aid professionals (ISCED level 5)	19
Annex C	(informative) Fitting room example.....	21
Annex D	(informative) Recommendation on the referral of clients for medical or other specialist examination and treatment.....	23
Annex E	(informative) Informational counselling to support hearing aid fitting management.....	24
Annex F	(informative) Terminology	27
Bibliography	41

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 43, *Acoustics*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The World Health Organisation (WHO) estimates that there are 360 million people with hearing impairment, approximately 5,3 % of the world population^[22]. Hearing aids (HAs) are one of the most widely-used treatment options for people with a hearing loss^{[39][40]}. For the proper use of HAs, hearing aid fitting management (HAFM) is a crucial issue for manufacturers, practitioners, hearing aid professionals and especially for HA users^{[39][42][43]}. Individually optimized outcome of HA use is supported by comprehensive HA fitting protocols^[42] and the impact of “poor fit and comfort” can lead to non-compliance, HA return^[43] and additional hearing loss with over-amplification. Accordingly, the whole process of HA fitting should be optimized to achieve functional benefits, user satisfaction and cost-effectiveness.

Two observations are important to take into account when developing an HAFM standard. Firstly, the term “hearing aid fitting” is widely used^{[16][44]–[46]} among service providers and industry sectors. Secondly, it has potentially conflicting interpretations: while guidelines for HA fitting have been written to tackle these issues by various national and professional bodies^{[17][18][23]–[32][34]–[37][47][48]}, many jurisdictions are still not covered worldwide and there is a need to promote a more common understanding of the HA fitting process. It is likely that different understanding of fitting has led to non-uniform care, outcome variability and, in many cases, dissatisfaction with the use of HAs.

The main purpose of this document is thus to provide a general framework for HAFM including the pre- and post-fitting stages to make it more explicit and transparent so that all related tasks, including professional services, administration and financial aspects can be systematized. The overall objective is to achieve the best possible hearing rehabilitation, which can only be accomplished through adequate knowledge, training and skills of the professional and a systematic approach to HA fitting in close collaboration with the client. The general framework of HAFM in this document is divided into six stages (client profile, counselling, hearing aid fitting, verification and validation, post-fitting counseling, and follow-up) based on the common practices of hearing aid professionals, and as recommended by various pre-existing guidelines.

By dividing the hearing aid fitting process into stages, HAFM service providers can systematically identify and administer the service components needed for high service quality, user satisfaction, client-centered services, client self-efficacy and compliance rates with HAs (e.g. consistently using HAs and attending follow-up appointments). The stages focus on the components of the framework to achieve high rehabilitation outcomes such as communication skills, speech intelligibility, perception of the acoustic environment, comfort for the HA users and sound quality. In addition, this document can be a basis for making cost assessments for each stage or component, which can help improve public health funding systems. Another possible application is to use this document as a minimum basis for the development of professional training programs in HAFM.

Acoustics — Hearing aid fitting management (HAFM)

1 Scope

This document applies to hearing aid fitting management (HAFM) services offered by hearing aid professionals (HAP) when providing benefit for their clients. The provision of hearing aids relies on the knowledge and practices of a hearing aid professional, to ensure the proper fitting and adequate service in the interest of the client with hearing loss.

This document specifies general processes of HAFM from the client profile to the follow-up through administering, organising and controlling hearing aid fitting through all stages. It also specifies important preconditions such as education, facilities and systems that are required to ensure proper services.

The focus of this document is the services offered to the majority of adult clients with hearing impairment. It is recognized that certain populations with hearing loss such as children, persons with other disabilities or persons with implantable devices can require services outside the scope of this document. This document generally applies to air conduction hearing aids and for the most part also to bone conduction devices.

Hearing loss can be a consequence of serious medical conditions. Hearing aid professionals are not in a position to diagnose or treat such conditions. When assisting clients seeking hearing rehabilitation without prior medical examination, hearing aid professionals are expected to be observant of symptoms of such conditions and refer to proper medical care.

Further to the main body of the document, which specifies the HAFM requirements and processes, several informative annexes are provided. Appropriate education of hearing aid professionals is vital for exercising HAFM. [Annex A](#) defines the competencies required for the HAFM processes. [Annex B](#) offers a recommended curriculum for the education of hearing aid professionals. [Annex C](#) is an example of an appropriate fitting room. [Annex D](#) gives guidance on the referral of clients for medical or other specialist examination and treatment. [Annex E](#) is a recommendation for important information to be exchanged with the client during the process of HAFM. [Annex F](#) is a comprehensive terminology list offering definitions of the most current terms related to HAFM.

It is the intention that these annexes be helpful to those who wish to deliver HAFM of the highest quality.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8253-1, *Acoustics — Audiometric test methods — Part 1: Pure-tone air and bone conduction audiometry*

ISO 8253-2, *Acoustics — Audiometric test methods — Part 2: Sound field audiometry with pure-tone and narrow-band test signals*

ISO 8253-3, *Acoustics — Audiometric test methods — Part 3: Speech audiometry*

IEC 60118-7, *Electroacoustics — Hearing aids — Part 7: Measurement of the performance characteristics of hearing aids for production, supply and delivery quality assurance purposes*

IEC 60645-1:2017, *Electroacoustics — Audiometric equipment — Part 1: Equipment for pure-tone and speech audiometry*

IEC 60645-5, *Electroacoustics — Audiometric equipment — Part 5: Instruments for the measurement of aural acoustic impedance/admittance*

IEC 61669, *Electroacoustics — Measurement of real-ear acoustic performance characteristics of hearing aids*

International Standard Classification of Education, ISCED. United Nations Educational, Scientific and Cultural Organization, 2011, ISBN 978-92-9189-123-8, <http://uis.unesco.org/sites/default/files/documents/international-standard-classification-of-education-isced-2011-en.pdf>

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 auditory dynamic range

difference between the *hearing* (3.7) threshold and the uncomfortable loudness level (UCL)

3.2 client

person with *hearing loss* (3.8) being serviced by a *HAP* (3.13)

3.3 client profile

comprehensive record of a *client's* (3.2) auditory functionality, social situation, activity opportunities, needs and expectations as well as a client's audiological and medical history

3.4 ear impression

representation of the three-dimensional geometry of the relevant part of the concha and ear canal

3.5 fine-tuning

adjustment of the *hearing aid system* (3.10) to best match the needs and preferences of the *client* (3.2)

3.6 fitting system

set of devices typically comprising a computer, fitting software and a programming interface used to adjust *hearing aids* (3.9)

3.7 hearing

manner in which a person detects, discriminates, identifies and cognitively processes sounds

3.8 hearing loss

reduction of the *hearing* (3.7) ability

3.9 hearing aid

wearable electroacoustic instrument intended to process sounds in order to compensate for *hearing loss* (3.8)

Note 1 to entry: Hearing aids are medical devices and comply with the requirements of IEC 60601-2-66.