

ELEKTRILISED MEDITSIINISEADMED. OSA 2-41:
ERINÕUDED KIRURGIAS JA DIAGNOOSIMISEL
KASUTATAVATE VALGUSTITE ESMASELE OHUTUSELE
JA OLULISTELE TOIMIMISNÄITAJATELE

Medical electrical equipment - Part 2-41: Particular requirements for the basic safety and essential performance of surgical luminaires and luminaires for diagnosis

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN IEC 60601-2-41:2021 sisaldab Euroopa standardi EN IEC 60101-2-41:2021 ingliskeelset teksti.	This Estonian standard EVS-EN IEC 60601-2-41:2021 consists of the English text of the European standard EN IEC 60101-2-41:2021.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 17.12.2021.	Date of Availability of the European standard is 17.12.2021.
Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.20, 11.040.55, 11.040.99

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about standards copyright protection, please contact the Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Medical electrical equipment - Part 2-41: Particular requirements
for the basic safety and essential performance of surgical
luminaires and luminaires for diagnosis
(IEC 60601-2-41:2021)

Appareils électromédicaux - Partie 2-41: Exigences
particulières pour la sécurité de base et les performances
essentielles des éclairages chirurgicaux et des éclairages
de diagnostic
(IEC 60601-2-41:2021)

Medizinische elektrische Geräte - Teil 2-41: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Operationsleuchten
und Untersuchungsleuchten
(IEC 60601-2-41:2021)

This European Standard was approved by CENELEC on 2021-10-08. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62D/1859/FDIS, future edition 3 of IEC 60601-2-41, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60101-2-41:2021.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2022-07-08 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2024-10-08 document have to be withdrawn

This document supersedes EN 60601-2-41:2009 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

Endorsement notice

The text of the International Standard IEC 60601-2-41:2021 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60364-7-710:2002 NOTE Harmonized as HD 60364-7-710:2012 (modified)

IEC 60598-1	NOTE Harmonized as EN IEC 60598-1
IEC 60598-2-1	NOTE Harmonized as EN IEC 60598-2-1
IEC 60598-2-4	NOTE Harmonized as EN 60598-2-4
IEC 60598-2-22	NOTE Harmonized as EN 60598-2-22
IEC 60598-2-25	NOTE Harmonized as EN 60598-2-25
IEC 60601-1-10	NOTE Harmonized as EN 60601-1-10
IEC 60601-2-18	NOTE Harmonized as EN 60601-2-18
IEC 60601-2-50	NOTE Harmonized as EN IEC 60601-2-50
ISO 9680	NOTE Harmonized as EN ISO 9680

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

The Annex ZA of EN 60601-1:2006 and EN 60601-1:2006/A1:2013 apply with the following changes:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Add the following references:</i>				
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-		+ corrigendum Mar.	2010
+ A1	2012		+ A1	2013
-	-		+ A12	2014
IEC 62471 (mod)	2006	Photobiological safety of lamps and lamp systems	EN 62471	2008
ISO/CIE 11664-1	-	Colorimetry - Part 1: CIE standard colorimetric observers	EN ISO/CIE 11664-1	-
ISO/CIE 19476	2014	Characterization of the performance of illuminance meters and luminance meters		-
ANSI C78.377	2017	Specifications for the Chromaticity of Solid-State Lighting (SSL) Products		-
CIE 13.3	1995	Method of measurement and specifying colour rendering properties of light sources		-
CIE 15	2018	Colorimetry	-	-

Replace the following references:

IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	EN 60529	1991
-	-		+ corrigendum May	1993
+ A1	1999		+ A1	2000
+ A2	2013		+ A2	2013

IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2:EN 60601-1-2 General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests	2015
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6:EN 60601-1-6 General requirements for basic safety and essential performance - Collateral standard: Usability	2010

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-41: Particular requirements for the basic safety and essential
performance of surgical luminaires and luminaires for diagnosis**

**Appareils électromédicaux –
Partie 2-41: Exigences particulières pour la sécurité de base et
les performances essentielles des éclairages chirurgicaux et des éclairages
de diagnostic**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2021 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC online collection - oc.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 18 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC online collection - oc.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-41: Particular requirements for the basic safety and essential
performance of surgical luminaires and luminaires for diagnosis**

**Appareils électromédicaux –
Partie 2-41: Exigences particulières pour la sécurité de base et
les performances essentielles des éclairages chirurgicaux et des éclairages
de diagnostic**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.20; 11.040.55; 11.040.99

ISBN 978-2-8322-1001-3

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references	10
201.3 Terms and definitions.....	10
201.4 General requirements.....	15
201.5 General requirements for testing ME EQUIPMENT.....	15
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	16
201.7 ME EQUIPMENT identification, marking and documents.....	17
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	19
201.9 Protection against mechanical hazards of ME EQUIPMENT and ME SYSTEMS	19
201.10 * Protection against unwanted and excessive radiation HAZARDS	23
201.11 Protection against excessive temperatures and other HAZARDS.....	38
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	40
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	41
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	41
201.15 Construction of ME EQUIPMENT	41
201.16 ME SYSTEMS.....	41
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	42
Annexes	43
Annex AA (informative) Particular guidance and rationale.....	44
Annex BB (informative) Requirements for SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS	50
Annex CC (informative) Technical information for conducting tests	52
Bibliography.....	58
Index of defined terms used in this particular standard.....	59
Figure 201.101 – Examples of providing power to a SURGICAL LUMINAIRE	11
Figure 201.102 – Examples of different distance and illuminance terms	14
Figure 201.103 – DETACHABLE HANDLE attachment and detachment tests	20
Figure 201.104 – Test for ease of motion.....	22
Figure 201.105 – Light distribution	25
Figure 201.106 – CENTRAL ILLUMINANCE measurement.....	26
Figure 201.107 – Measurements for determining LIGHT FIELD DIAMETERS d_{10} and d_{50}	26
Figure 201.108 – Illuminance measurement with one mask.....	27
Figure 201.109 – Illuminance measurement with two masks	28
Figure 201.110 – Illuminance measurement with four different positions of two masks.....	29
Figure 201.111 – Simulated cavity for illuminance measurements.....	30
Figure 201.112 – Illuminance measurement at the bottom of a simulated cavity, with one mask.....	30
Figure 201.113 – Illuminance measurement at the bottom of a simulated cavity, with two masks.....	31
Figure 201.114 – Illuminance measurement at the bottom of a simulated cavity, with four different positions of two masks	32

Figure 201.115 – Measurement of DEPTH OF ILLUMINATION	33
Figure 201.116 – Caution symbol for hot surface	39
Figure CC.1 – Location of the illuminance meter and mask on the luminaire	54
Figure CC.2 – Movement of the illuminance meter and mask on the luminaire	55
Figure CC.3 – Location of the spectrometer and mask on the luminaire	55
Figure CC.4 – Test conditions for measuring illuminance at 500 mm.....	56
Figure CC.5 – Test method for maintaining the measurement distance	56
Table 23 – Allowable maximum temperatures for ME EQUIPMENT parts that are likely to be touched.....	39
Table AA.1 – Summary of photobiological HAZARDS and exposure limits.....	45
Table BB.1 – Summary of requirements for SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS	50

This document is a preview generated by EVS

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-41: Particular requirements for the basic safety and essential performance of surgical luminaires and luminaires for diagnosis**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-41 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice. It is an International Standard.

An annex in this publication contains an embedded Microsoft Excel file intended to help in organizing data and calculating exposures associated with photobiological HAZARDS. This file is intended to be used as a complement and does not form an integral part of the publication.

This third edition cancels and replaces the second edition published in 2009 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) revised the statement of essential performance;
- b) added exposure limits, test conditions, calculation methods and safety warnings related to photobiological hazards;
- c) removed the terms "MINOR SURGICAL LUMINAIRES" and "MAJOR SURGICAL LUMINAIRES";

- d) added definitions of MAXIMUM ILLUMINANCE DISTANCE and REFERENCE DISTANCE and allowed MANUFACTURERS to measure some performance characteristics at the REFERENCE DISTANCE that they specify;
- e) replaced the region of acceptable chromaticity in (x,y) colour space with a requirement for $D_{u,v}$;
- f) added a requirement for acceptable drift of the lighthouse when attached to the suspension system;
- g) added a requirement for fluid ingress protection;
- h) revised Table 201.101 of IEC 60601-2-41:2009 and IEC 60601-2-41:2009/AMD1:2013 and moved it to Annex BB;
- i) specified a new device for measuring SHADOW DILUTION in a simulated cavity;
- j) specified test conditions for luminaires equipped with distance sensors.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
62D/1859/FDIS	62D/1879/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type*;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific publication. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The "colour inside" logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION

This particular standard concerns the basic safety and essential performance of SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS.

It amends and supplements IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, hereinafter referred to as the "general standard".

The requirements of this particular standard take priority over those of the general standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-41: Particular requirements for the basic safety and essential performance of surgical luminaires and luminaires for diagnosis

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS, hereafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

This particular standard does not apply to

- headlights;
- endoscopes, laparoscopes and their light sources, which are covered by IEC 60601-2-18;
- luminaires used in dentistry, which are covered by ISO 9680;
- luminaires for general purposes, which are covered by IEC 60598-2-1 and IEC 60598-2-4;
- luminaires dedicated to therapeutic purposes;
- special purpose lights with different conditions of use such as light sources intended solely for decontamination of air and surfaces, UV lights for dermatological diagnosis, slit lamps for ophthalmology, lights for surgical microscopes and lights for surgical navigation systems;
- lights connected to surgical instruments, such as luminous retractors;
- luminaires for emergency lighting, which are covered by IEC 60598-2-22.

NOTE See also 4.2 of the general standard.

SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS are medical devices and not general lighting equipment.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS as defined in 201.3.

¹ The general standard is IEC 60601-1 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.