

LABORATOORIUMI KLAAS- JA PLASTNÕUD.
MAHUMÕÕDUNÕUD. MAHU KATSETAMISE JA
KASUTAMISE MEETODID

Laboratory glass and plastic ware - Volumetric
instruments - Methods for testing of capacity and for
use (ISO 4787:2021)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 4787:2022 sisaldab Euroopa standardi EN ISO 4787:2021 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 4787:2022 consists of the English text of the European standard EN ISO 4787:2021.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 01.12.2021.	Date of Availability of the European standard is 01.12.2021.
Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 17.060

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about standards copyright protection, please contact the Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 4787

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2021

ICS 17.060

Supersedes EN ISO 4787:2011

English Version

Laboratory glass and plastic ware - Volumetric
instruments - Methods for testing of capacity and for use
(ISO 4787:2021)

Verrerie et matériel en plastique de laboratoire -
Instruments volumétriques - Méthodes d'essai de la
capacité et d'utilisation (ISO 4787:2021)

Laborgeräte aus Glas und Kunststoff -
Volumenmessgeräte - Prüfverfahren und Anwendung
(ISO 4787:2021)

This European Standard was approved by CEN on 20 November 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 4787:2021) has been prepared by Technical Committee ISO/TC 48 "Laboratory equipment" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2022, and conflicting national standards shall be withdrawn at the latest by June 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 4787:2011.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 4787:2021 has been approved by CEN as EN ISO 4787:2021 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	2
5 Volume and reference temperature	2
5.1 Unit of volume.....	2
5.2 Reference temperature.....	2
6 Apparatus and calibration liquid	2
6.1 Balance.....	2
6.2 Measurement devices.....	2
6.3 Calibration liquid.....	3
6.4 Receiving vessel.....	3
7 Factors affecting the accuracy of volumetric instruments	3
7.1 General.....	3
7.2 Temperature.....	3
7.2.1 Temperature of the volumetric instrument.....	3
7.2.2 Temperature of calibration liquid.....	3
7.3 Cleanliness of surface.....	4
7.4 Conditions of used volumetric instruments.....	4
7.5 Delivery time and waiting time.....	4
8 Setting the meniscus	5
8.1 General.....	5
8.2 Setting the meniscus.....	5
8.2.1 Meniscus of transparent liquids.....	5
8.2.2 Meniscus of opaque liquids.....	7
9 Calibration procedure	7
9.1 General.....	7
9.2 Test room.....	7
9.3 Filling and delivery.....	7
9.3.1 Volumetric flasks and measuring cylinders.....	7
9.3.2 Pipettes adjusted to deliver.....	7
9.3.3 Pipettes adjusted to contain.....	8
9.3.4 Burettes adjusted to deliver.....	8
9.3.5 Pycnometers.....	9
9.4 Weighing.....	9
9.5 Volume and uncertainty calculation.....	9
10 Procedure for use	10
10.1 General.....	10
10.2 Volumetric flasks (in accordance with ISO 1042 or ISO 5215).....	11
10.3 Measuring cylinders (in accordance with ISO 4788 or ISO 6706).....	11
10.4 Burettes (in accordance with ISO 385).....	11
10.5 Pipettes.....	12
10.5.1 Pipettes adjusted to deliver (see ISO 648 and ISO 835, or other pipettes, e.g. plastic ones).....	12
10.5.2 Pipettes adjusted to contain.....	12
10.6 Pycnometers.....	12
Annex A (informative) Cleaning of volumetric glassware	13

Annex B (informative) Cleaning of volumetric plasticware	14
Annex C (normative) Calculation formulae and tables	15
Annex D (informative) Coefficient of cubic thermal expansion	19
Annex E (informative) Uncertainty estimation and repeatability calculation	20
Bibliography	21

This document is a preview generated by EVS

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 332, *Laboratory equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 4787:2010), which has been technically revised.

The main changes compared to the previous edition are as follows:

- a) volumetric plastic ware has been included;
- b) new information on meniscus adjustment of convex meniscus has been added; namely, altered procedure "Upper edge of the graduation line is horizontally tangential to the highest point of meniscus" as compared to older procedure "Upper edge of the graduation line is horizontally tangential to the lowest point of the meniscus";
- c) improved figures for meniscus adjustment have been provided;
- d) [Table 1](#) has been improved;
- e) new [Table 2](#) for minimum requirements for the measurement devices has been added;
- f) new test room ambient conditions have been added;
- g) new information regarding repeatability and uncertainty has been added in [Annex E](#);
- h) [Formula \(C.1\)](#) has been changed to [Formula \(1\)](#).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The International Standards for the individual volumetric instruments include clauses on the specification of capacity (volume); these clauses describe the method of manipulation in sufficient detail to determine the capacity without ambiguity. This document contains supplementary information.

This document is a preview generated by EVS

Laboratory glass and plastic ware — Volumetric instruments — Methods for testing of capacity and for use

1 Scope

This document provides methods for the testing, calibration and use of volumetric instruments made from glass and plastic in order to obtain the best accuracy in use.

NOTE Testing is the process by which the conformity of the individual volumetric instrument with the appropriate standard is determined, resulting in the determination of its error of measurement at one or more points.

This document is applicable to volumetric instruments with nominal capacities in the range of 100 μ l to 10 000 ml. These include single-volume pipettes (see ISO 648), graduated pipettes (see ISO 835), burettes (see ISO 385), volumetric flasks (see ISO 1042 and ISO 5215), and graduated measuring cylinders (see ISO 4788 and ISO 6706).

The methods are not intended for testing of volumetric instruments with capacities below 100 μ l such as micro-glassware.

This document does not deal specifically with pycnometers as specified in ISO 3507. However, the procedures specified for the determination of volume of glassware can, for the most part, also be followed for the determination of a pycnometer volume. For some types of pycnometers, special handling can be necessary.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 385, *Laboratory glassware — Burettes*

ISO 648, *Laboratory glassware — Single-volume pipettes*

ISO 835, *Laboratory glassware — Graduated pipettes*

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

ISO 1773, *Laboratory glassware — Narrow-necked boiling flasks*

ISO 3507, *Laboratory glassware — Pycnometers*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 4788, *Laboratory glassware — Graduated measuring cylinders*

ISO 4797, *Laboratory glassware — Boiling flasks with conical ground joints*

ISO 5215¹⁾, *Laboratory plastic ware — Volumetric flasks*

ISO 6706, *Plastics laboratory ware — Graduated measuring cylinders*

ISO 24450, *Laboratory glassware — Wide-necked boiling flasks*

1) Under preparation. Stage at the time of publication: ISO/DIS 5215:2021.

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 99 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 Principle

The general procedure for testing the capacity (volume) and for use is based upon a determination of volume of water, either contained in or delivered by the volumetric instrument. This volume of water is based upon knowledge of its mass under consideration of buoyancy and its density (gravimetric method).

5 Volume and reference temperature

5.1 Unit of volume

The unit of volume shall be the millilitre (ml), which is equivalent to one cubic centimetre (cm³).

5.2 Reference temperature

The standard reference temperature, i.e. the temperature at which the volumetric instrument is intended to contain or deliver its volume (capacity), shall be 20 °C.

When the volumetric instrument is required for use in a country which has adopted a standard reference temperature of 27 °C (according to ISO 384), this figure shall replace 20 °C in [Formula \(1\)](#).

6 Apparatus and calibration liquid

6.1 Balance

The balance used for testing shall be chosen in accordance with the minimum requirements specified in [Table 1](#), depending on the nominal volume of the volumetric instrument under test.

Table 1 — Minimum requirements for the balance

Nominal capacity (volume) V	Resolution	Repeatability	Expanded uncertainty in use $U (k = 2)^a$
$100 \mu\text{l} \leq V \leq 10 \text{ ml}$	mg	mg	mg
$100 \mu\text{l} \leq V \leq 10 \text{ ml}$	0,1	0,2	0,4
$10 \text{ ml} < V \leq 1\ 000 \text{ ml}$	1	2	4
$V > 1\ 000 \text{ ml}$	10	10	40

^a Expanded uncertainty in use estimated according to Reference [1] (which includes applicable definitions) at the value of the nominal volume. If uncertainty in use is not available, then the uncertainty at calibration should be taken.

6.2 Measurement devices

The minimum requirements for each relevant measurement device are specified in [Table 2](#).