

Clinical laboratory testing and in vitro diagnostic test systems - Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices - Part 2: Evaluation of performance of antimicrobial susceptibility test devices against reference broth micro-dilution (ISO 20776-2:2021)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 20776-2:2022 sisaldab Euroopa standardi EN ISO 20776-2:2022 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 20776-2:2022 consists of the English text of the European standard EN ISO 20776-2:2022.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 19.01.2022.	Date of Availability of the European standard is 19.01.2022.
Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.100.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about standards copyright protection, please contact the Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Clinical laboratory testing and in vitro diagnostic test systems - Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices - Part 2: Evaluation of performance of antimicrobial susceptibility test devices against reference broth micro-dilution (ISO 20776-2:2021)

Systèmes d'essais en laboratoire et de diagnostic in vitro - Sensibilité in vitro des agents infectieux et évaluation des performances des dispositifs pour antibiogrammes - Partie 2: Évaluation des performances des dispositifs pour antibiogrammes par rapport à une méthode de référence de microdilution en bouillon (ISO 20776-2:2021)

Labormedizinische Untersuchungen und In-vitro-Diagnostika-Systeme - Empfindlichkeitsprüfung von Infektionserregern und Evaluation von Geräten zur antimikrobiellen Empfindlichkeitsprüfung - Teil 2: Evaluation der Leistung einer Vorrichtung zur antimikrobiellen Empfindlichkeitsprüfung (ISO 20776-2:2021)

This European Standard was approved by CEN on 16 December 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 20776-2:2022) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2022, and conflicting national standards shall be withdrawn at the latest by January 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 20776-2:2007.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard : Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20776-2:2021 has been approved by CEN as EN ISO 20776-2:2022 without any modification.

Contents

Page

Foreword	iv
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Test methods	5
4.1 General.....	5
4.2 Methods.....	5
4.2.1 Reference method.....	5
4.2.2 Strain selection.....	5
4.2.3 Quality control (QC).....	6
4.2.4 Quality control (QC) of the reference method.....	6
4.2.5 Reproducibility testing of test device.....	6
4.2.6 Isolate testing protocol.....	7
4.2.7 Inoculum preparation.....	7
4.2.8 Discrepancy resolution testing.....	7
4.2.9 System under evaluation.....	7
5 Data analysis and acceptance criteria	8
5.1 Accuracy of test device.....	8
5.1.1 General.....	8
5.1.2 MIC devices.....	8
5.1.3 Qualitative AST devices.....	8
5.2 Quality control (QC) of test device.....	8
5.3 Reproducibility of test device.....	8
5.4 Documents related to study.....	8
Annex A (informative) Evaluating the performance of MIC tests	9
Annex B (informative) Rationale for bias analysis	12
Annex C (informative) Sensitivity and specificity analyses for qualitative tests	14
Bibliography	16

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 20776-2:2007), which has been technically revised.

The main changes are as follows:

- Revision in the title of this document to better align with the intended information.
- Addition of an Introduction (not present in the first edition).
- Revised [Clause 3](#) as follows:
 - Removed definitions for category agreement, susceptible, intermediate, resistant, non-susceptible, major discrepancy, minor discrepancy, very major discrepancy, breakpoint test and zone diameter;
 - Added definition for contemporary isolate ([3.11.1](#)), and removed definitions for fresh isolate, recent isolate;
 - Added definitions for reproducibility ([3.9](#)), bias of the test method ([3.10.3](#)), sensitivity analysis ([3.10.4.1](#)), specificity analysis ([3.10.4.2](#)), bacterial organism group ([3.16](#));
 - Added definition for qualitative test ([3.7](#)) and removed definition for breakpoint test;
 - Revised definitions for minimum inhibitory concentration test ([3.4](#)), breakpoint ([3.6](#)), quality control ([3.8](#)), discrepancy ([3.10.1](#)).
- Reordered [Clause 4](#) (Test methods);

- Moved general requirements for a performance evaluation as a separate section, to the overview (now renamed general section, [subclause 4.1](#)) under test methods);
- Revised quality control section, [subclause 4.2](#), and referenced EUCAST and CLSI documents for quality control ranges;
- Revised [subclause 4.2.1](#) (Reference method) to add variability;
- Revised [subclause 4.2.2](#) (Strain selection) and incorporated new definition of contemporary isolates ([3.11.1](#));
- Revised [subclause 4.2.5](#) (Reproducibility testing);
- Updated [subclause 4.2.8](#) (Discrepancy resolution testing);
- Combined data analysis and acceptance criteria subclauses ([Clause 5](#));
- Revised [subclause 5.1](#) (Accuracy of test device) to remove category agreement;
- Revised data analysis for MIC devices to remove category agreement. Added bias requirement;
- Removed acceptance for breakpoint AST devices;
- Added provisions on acceptance criteria for qualitative AST devices ([5.1.3](#)) and included sensitivity and specificity requirements;
- Revised subclauses on quality control of test device and reproducibility of test device ([5.2](#) and [5.3](#));
- Revised Bibliography;
- Added [Annex A](#) — Evaluation the Performance of MIC Tests, [Annex B](#) — Rationale for Bias Analysis, and [Annex C](#) — Sensitivity and Specificity Analyses for Qualitative Tests.

A list of all parts in the ISO 20776 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

In vitro antimicrobial susceptibility tests are performed on bacteria suspected of causing disease, particularly if the isolate is thought to belong to a species that can exhibit resistance to frequently used antimicrobial agents. The tests are also important in resistance surveillance, epidemiological studies of susceptibility and in comparisons of new and existing agents.

Dilution procedures are used to determine the minimum inhibitory concentrations (MICs) of antimicrobial agents for antimicrobial susceptibility testing. MIC methods are used in resistance surveillance, defining and identifying wild type phenotypes, comparative testing of new agents, to establish the susceptibility of organisms that give equivocal results in routine tests, for tests on organisms where routine tests can be unreliable and when a quantitative result is required for clinical management. In dilution tests, bacterial strains are tested for their ability to produce visible growth in broth (broth dilution) containing serial dilutions of the antimicrobial agent or on a series of agar plates (agar dilution).

The lowest concentration of an antimicrobial agent (in mg/l) that, under defined in vitro conditions, prevents the appearance of visible growth of an isolated bacterial strain within a defined period of time, is known as the MIC. Careful control and standardization are required for intra- and interlaboratory reproducibility of broth MIC tests. The MICs of quality control (QC) strains generally span three doubling dilutions with a dominant central value, but can have a four-dilution range.

Broth micro-dilution denotes the performance of the broth dilution test in micro-dilution trays. Broth micro-dilution is now one of the most common methods used globally to perform antimicrobial susceptibility tests.

This document is the second edition of ISO 20776-2. It is designed for the evaluation of antimicrobial test devices against the standard broth micro-dilution reference method (ISO 20776-1) using pure cultures of aerobic bacteria that are easily grown by overnight incubation on agar and grow well in standardized micro-dilution trays containing standardized Mueller-Hinton broth (volume of $\leq 200 \mu\text{l}$), which can need to be modified depending on the antimicrobial agent being tested.

Quantitative MIC and qualitative evaluations detailed in this revised document measure the accuracy, reproducibility and QC of tests performed with antimicrobial test devices that generate MIC values against the standard broth micro-dilution reference method. Antimicrobial agar disc diffusion tests are not included in this document.

This document has been revised using the premise that the MIC test is an in vitro assay, subject to intra- and interlaboratory assay variation. When making the comparison between any derivative test and that of the reference method, it is appropriate to apply measures of assay performance only and not result interpretation. For this reason, and because interpretive categories were removed from the second edition of ISO 20776-1, categorical agreement (CA) and its associated terminology, as described by the U.S. Food and Drug Administration (FDA), the Clinical and Laboratory Standards Institute (CLSI) M23 document, and other international documents, has not been applied. Avoiding an assessment of CA also assists in reducing the requirement to reassess assay performance automatically when the only change has been a breakpoint change (which is external to the assay itself).

This document applies to new performance evaluations initiated after the publication date; studies conducted prior to the acceptance date of this document should not need to be re-designed and/or re-analysed using these criteria. Studies conducted prior to these standards or acceptance of this document follow standard practice or guidance at the time of the study.

For derivative tests with more than three two-fold dilutions, assay performance is assessed with tools designed to measure accuracy using essential agreement (EA) and bias, and precision using EA only. For derivative tests with 1 to 3 concentrations, assay performance is assessed using standard sensitivity and specificity measures.

Clinical laboratory testing and in vitro diagnostic test systems — Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices —

Part 2: Evaluation of performance of antimicrobial susceptibility test devices against reference broth micro-dilution

1 Scope

This document establishes acceptable performance criteria for antimicrobial susceptibility test (AST) devices that are used to determine minimum inhibitory concentrations (MIC) of bacteria to antimicrobial agents in medical laboratories.

This document specifies requirements for AST devices and procedures for assessing performance of such devices. It defines how a performance evaluation of an AST device is to be conducted.

This document has been developed to guide manufacturers in the conduct of performance evaluation studies.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20776-1, *Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices — Part 1: Broth micro-dilution reference method for testing the in vitro activity of antimicrobial agents against rapidly growing aerobic bacteria involved in infectious diseases*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

antimicrobial susceptibility test device

AST device

device, including all specified components used to obtain test results that allow MIC determination of bacteria with specific antimicrobial agents

Note 1 to entry: Specific components of the device include inoculators, disposables and reagents, media used to perform the test, and readers or analysers. Non-specific components, such as swabs, pipettes and tubes, are not part of the device.