
Ophthalmic instruments — Direct ophthalmoscopes

Instruments ophtalmiques — Ophtalmoscopes directs



This document is a preview generated by EUS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Classification.....	2
5 Requirements.....	2
5.1 General.....	2
5.2 Optical requirements.....	2
5.3 Construction and function of the viewing system.....	3
5.4 Construction and function of the illumination system.....	3
5.5 Optical radiation hazard with direct ophthalmoscopes.....	4
6 Test methods.....	4
6.1 General.....	4
6.2 Checking the optical and functional requirements.....	4
7 Accompanying documents.....	5
8 Marking.....	5
Bibliography.....	6

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 10942:2006), which has been technically revised.

The main changes compared to the previous edition are as follows:

- revision of the dated references;
- editorial update of the whole document.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Ophthalmic instruments — Direct ophthalmoscopes

1 Scope

This document, together with ISO 15004-1 and ISO 15004-2, specifies minimum requirements and test methods for hand-held direct ophthalmoscopes designed for directly observing the eye fundus.

This document takes precedence over ISO 15004-1 and ISO 15004-2, if differences exist.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15004-1, *Ophthalmic instruments — Fundamental requirements and test methods — Part 1: General requirements applicable to all ophthalmic instruments*

ISO 15004-2, *Ophthalmic instruments — Fundamental requirements and test methods — Part 2: Light hazard protection*

IEC 60601-1:2005+A1:2012, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

ophthalmoscope

optical instrument used to examine the external and internal parts of the eye, particularly the media and the fundus

3.2

direct ophthalmoscope

ophthalmoscope (3.1) which provides an illuminating system, an observation system and viewing lenses which allow the observer to visualize the patient's eye directly, that is without the formation of an intermediate image

3.3

viewing lens

lens which is positioned between the observer's eye(s) and the eye to be examined in order to achieve optimum focus, i.e. to correct for patient's and/or observer's refractive error and/or accommodation

Note 1 to entry: In direct ophthalmoscopes when a selection of such lenses is required, these are integrated with or mounted in a disc or other mechanical means by which the user can easily position the lens of choice centrally in the visual path.