TECHNICAL SPECIFICATION

ISO/TS 24498

First edition 2022-01

Paper, board and pulps — Estimation of uncertainty for test methods by



Reference number ISO/TS 24498:2022(E)



© ISO 2022

mentation, no part of all including photod from either All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

| Coı | ntents | Page |
|------|--|-----------|
| Fore | word | iv |
| Intr | oduction | v |
| 1 | Scope | 1 |
| 2 | Normative references | |
| 3 | Terms and definitions | |
| 4 | Procedure | |
| - | | |
| 5 | Preparation of an interlaboratory study 5.1 Laboratories | 3 |
| | 5.1.1 Qualification of laboratories | 3 |
| | 5.1.2 Number of participating laboratories | |
| | 5.2 Sample preparation and distribution | 3 |
| | 5.2.1 Number and type of material | |
| | 5.2.2 Selection of the material | |
| | 5.2.3 Identification and packaging | |
| | 5.2.4 Additional and specific care | |
| | | |
| 6 | Testing | |
| 7 | Analysis of the results | |
| | 7.1 Calculations | 6 |
| | 7.2 Case of interlaboratory study organised by the working group | |
| | 7.2.2 Consistency evaluation | |
| | 7.2.3 Handling of outlying data | 7 |
| | 7.3 Case of data provided from proficiency testing services | |
| 8 | Re-evaluation of the uncertainty | 8 |
| 9 | Report | 8 |
| 10 | Archiving raw data | 8 |
| | ex A (informative) Tappi and ISO codifications | |
| | | |
| Ann | ex B (informative) Example of a recommended form to be distributed to participate laboratories for the collation of results for analysis | ang 10 |
| Δnn | ex C (informative) Example of precision statements for physical properties | |
| | | |
| DIUI | ography | 12 |
| | | |
| | 9 | |
| | | |
| | | 5 |
| | | |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 6, Paper, board and pulps.

This first edition cancels and replaces the second edition of ISO/TR 24498:2019, which has been technically revised.

The main changes are as follows:

- ISO/TR 24498 has been changed into ISO/TS 24498 adding normative language
- Lignins and kraft liquors have been introduced in the scope of the document, and a subclause on the sampling of these materials in <u>5.2</u> has been added
- Subclause <u>7.3</u> has been updated.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

One step in the development of any new standard test method is to estimate the uncertainty of the method. After such a procedure, a "Precision statement" is usually included in ISO test methods for pulp, paper and board and is recommended by ISO/TC 6 for all new and revised ISO/TC 6 standards. This is normally performed in a precision experiment, in which samples are sent to a number of laboratories and the results are compared. Such a precision experiment is often referred to as "interlaboratory comparative testing".

The procedures for conducting a precision statement are outlined in the ISO 5725 series^[1], which is general and does not cover the special conditions that apply in the testing of pulp, paper, board and cellulosic nanomaterials (this is the reason why some countries have published national standards or test methods dedicated to pulp, paper and board^{[2][3]}).

For example, paper and board materials as well as cellulosic nanomaterials are very sensitive to changes in relative humidity and temperature. Changes in the environmental conditions may induce significant moisture content variations in paper and board, which may induce changes in physical and mechanical properties.

Due to product heterogeneity, randomisation of the samples and /or test pieces is essential to minimize the impact of such variability. For the same reason, the variation in the properties can increase drastically when the test piece size decreases, for example when measuring grammage or Cobb water absorptiveness.

These reasons make it necessary to have special instructions for precision experiments for pulp, paper, board and cellulosic nanomaterials.

One effect of the heterogeneity of the product is that a large number of measurements is required in order to achieve sufficient precision. Most standardized test methods are therefore based on 10 or more measurements. The result is generally the average of these measurements.

Uncertainty has multiple components including a random component and a systematic component. This document focuses on the random component, defined by a repeatability and reproducibility of the measurements.

There are four main purposes for testing:

- Research, where the main question is whether there is an expected maximum difference between two samples, for instance, papers produced using different pulp mixtures.
- Verification of conformance with a specification. This can be at the production central testing laboratory site or in an independent laboratory.
- Evaluation of a new test method, where the aim is to verify that the precision of the test method is acceptable.
- Determination of a precision statement for an existing test method either where one does not exist or where it requires revision.

When the uncertainty of a test method is to be expressed, the following aspects should be considered.

- The conditions for the tests. Are the conditions as similar as possible, or as different as possible?
- The uncertainty can be expressed in different statistical measures, as a standard deviation or as a confidence interval.
- The uncertainty can be expressed either as a variation in the test results themselves, or as the difference between two test results.

This document is a preview general ded by tills

Paper, board and pulps — Estimation of uncertainty for test methods by interlaboratory comparisons

1 Scope

This document presents guidelines for a methodology for the estimation of the uncertainty of methods for testing lignins and kraft liquors, pulps, paper, board, cellulosic nanomaterials, as well as products thereof containing any portion of recycled material or material intended for recycling.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 187, Paper, board and pulps — Standard atmosphere for conditioning and testing and procedure for monitoring the atmosphere and conditioning of samples

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at https://www.electropedia.org/

3.1

interlaboratory comparison

organization, performance and evaluation of measurements or tests on the same or similar items by two or more laboratories in accordance with predetermined conditions

[SOURCE: ISO 13528:2015, 3.1]

3.2

interlaboratory test

series of measurements of one or more quantities performed independently by a number of laboratories on samples of a given material

[SOURCE: ISO 11459:1997, 3.16]

3.3

repeatability condition of measurement

repeatability condition

condition of measurement, out of a set of conditions that includes the same measurement procedure, same operators, same measuring system, same operating conditions and same location, and replicate measurements on the same or similar objects over a short period of time

Note 1 to entry: A condition of measurement is a repeatability condition only with respect to a specified set of repeatability conditions.

Note 2 to entry: In chemistry, the term "intra-serial precision condition of measurement" is sometimes used to designate this concept.