

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Ultrasonics – Hydrophones –

Part 1: Measurement and characterization of medical ultrasonic fields

Ultrasons – Hydrophones –

Partie 1: Mesurage et caractérisation des champs ultrasoniques médicaux



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Ultrasonics – Hydrophones –
Part 1: Measurement and characterization of medical ultrasonic fields**

**Ultrasons – Hydrophones –
Partie 1: Mesurage et caractérisation des champs ultrasoniques médicaux**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-1080-1

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope.....	9
2 Normative references	9
3 Terms and definitions	10
4 Symbols	32
5 Measurement requirements	34
5.1 Requirements for hydrophones and amplifiers.....	34
5.1.1 Preface.....	34
5.1.2 General	34
5.1.3 Sensitivity of a hydrophone.....	35
5.1.4 Directional response of a hydrophone.....	35
5.1.5 Effective hydrophone size.....	35
5.1.6 Choice of the size of a hydrophone active element	35
5.1.7 Bandwidth.....	37
5.1.8 Linearity	40
5.1.9 Hydrophone signal amplifier	40
5.1.10 Hydrophone cable length and amplifiers	40
5.2 Requirements for positioning and water baths.....	41
5.2.1 General	41
5.2.2 Positioning systems	41
5.2.3 Water bath.....	42
5.3 Requirements for data acquisition and analysis systems.....	43
5.4 Recommendations for ultrasonic equipment being characterized.....	43
6 Measurement procedure	43
6.1 General.....	43
6.2 Preparation and alignment	44
6.2.1 Preparation.....	44
6.2.2 Aligning an ultrasonic transducer and a hydrophone	44
6.3 Measurement.....	44
6.4 Analysis	44
6.4.1 Corrections for restricted bandwidth and spatial resolution	44
6.4.2 Uncertainties	44
7 Beam characterization	45
7.1 General.....	45
7.2 Primary pressure parameters	46
7.2.1 General	46
7.2.2 Peak-compressional acoustic pressure and peak-rarefactional acoustic pressure	47
7.2.3 Spatial-peak RMS acoustic pressure	47
7.2.4 Local distortion parameter	48
7.3 Intensity parameters derived from acoustic pressure.....	48
7.3.1 General	48
7.3.2 Intensity parameters using pulse-pressure-squared integral	49
8 Requirements for specific ultrasonic fields.....	52

8.1	General.....	52
8.2	Diagnostic fields	52
8.2.1	Simplified procedures and guidelines.....	52
8.2.2	Pulsed wave diagnostic equipment	52
8.2.3	Continuous wave diagnostic equipment	53
8.2.4	Diagnostic equipment with low acoustic output	54
8.3	Therapy fields	54
8.3.1	Physiotherapy equipment.....	54
8.3.2	High intensity therapeutic ultrasonic fields	55
8.3.3	Non-focused and weakly focused pressure pulses	55
8.4	Surgical fields	55
8.4.1	Lithotripters and pressure pulse sources for other therapeutic purposes	55
8.4.2	Low frequency surgical applications.....	56
8.5	Fields from other medical applications	56
9	Conformity statement.....	56
9.1	General.....	56
9.2	Maximum probable values.....	56
9.3	Sampling.....	57
Annex A	(informative) General rationale.....	58
Annex B	(informative) Hydrophones and positioning	60
B.1	General.....	60
B.2	Electrical loading considerations	60
B.3	Hydrophone signal amplifier.....	60
B.4	Hydrophone cable length and amplifiers.....	60
B.5	Transducer positioning.....	61
B.6	Alignment of hydrophones.....	62
B.7	Water bath lining material	62
B.8	Recommendations for ultrasonic equipment being characterized.....	62
B.9	Types of hydrophones.....	63
B.9.1	Ceramic needle hydrophones	63
B.9.2	PVDF needle hydrophones	63
B.9.3	PVDF membrane hydrophones	63
B.9.4	Fibre-optic and optic hydrophones	64
B.9.5	Relative performance of different types.....	65
B.10	Typical specification data for hydrophones.....	65
Annex C	(informative) Acoustic pressure and intensity	66
Annex D	(informative) Voltage to pressure conversion	68
D.1	General.....	68
D.2	Hydrophone deconvolution procedure	69
D.3	Converting the data between double-sided and single-sided spectra	70
D.4	Use of hydrophone calibration data	72
D.4.1	Calibration data interpolation	72
D.4.2	Calibration data extrapolation	72
D.4.3	Regularization filtering.....	73
D.5	Implication of the hydrophone deconvolution process on measurement duration	74
D.6	Validation of deconvolution implementation.....	75
Annex E	(informative) Correction for spatial averaging.....	76

E.1	Linear and quasilinear fields	76
E.2	Linear fields, quasilinear fields, and broadband nonlinearly distorted waveforms	78
Annex F (informative) Acoustic output parameters for multi-mode medical ultrasonic fields in the absence of scan-frame synchronization		81
F.1	General.....	81
F.2	Current philosophy.....	81
F.3	Need for an alternative approach	82
F.4	Proposed approach.....	82
F.4.1	Alternative philosophy	82
F.4.2	Alternative parameters.....	83
F.5	Measurement methods.....	84
F.5.1	General	84
F.5.2	Peak pressures.....	84
F.5.3	Temporal-average intensity	84
F.5.4	Frequency	85
F.5.5	Power	85
F.6	Discussion	85
F.6.1	Relationship to existing standards	85
F.6.2	Advantages	86
F.6.3	Disadvantages.....	86
Annex G (informative) Propagation medium and degassing.....		87
Annex H (informative) Specific ultrasonic fields.....		88
H.1	Diagnostic fields	88
H.1.1	Useful relationships between acoustical parameters	88
H.1.2	Pulsed wave diagnostic equipment	89
H.1.3	Continuous wave diagnostic equipment	89
H.2	Therapy fields	90
H.2.1	Physiotherapy equipment.....	90
H.2.2	High intensity therapeutic ultrasonic equipment	90
H.2.3	Non-focused and weakly focused pressure pulses	90
H.3	Surgical fields	90
H.3.1	Lithotripters	90
H.3.2	Low frequency surgical applications.....	90
Annex I (informative) Assessment of uncertainty in the acoustic quantities obtained by hydrophone measurements.....		91
I.1	General.....	91
I.2	Overall (expanded) uncertainty	91
I.3	Common sources of uncertainty	91
Annex J (informative) Transducer and hydrophone positioning systems.....		93
Annex K (informative) Beamwidth midpoint method.....		94
Bibliography.....		95
Figure 1 – Schematic diagram of the different planes and lines in an ultrasonic field		12
Figure 2 – Several apertures and planes for a transducer of unknown geometry		26
Figure 3 – Parameters for describing a focusing transducer of known geometry.....		29
Figure 4 – Schematic diagram of the method of determining pulse duration		46
Figure D.1 – A flow diagram of the hydrophone deconvolution process		70

Figure D.2 – Example of waveform deconvolution	74
Figure J.1 – Schematic diagram of the ultrasonic transducer and hydrophone degrees of freedom	93
Table 1 – Acoustic parameters appropriate to various types of medical ultrasonic equipment.....	45
Table B.1 – Typical specification data for hydrophones, in this case given at 1 MHz [69]	65
Table C.1 – Properties of distilled or de-ionized water as a function of temperature [71]	67
Table D.1 – Method of conversion from a double- to a single-sided spectrum	71
Table D.2 – Method of conversion from a single- to a double-sided spectrum	71
Table F.1 – Main basic parameters defined in this document or in IEC 61161	82
Table F.2 – List of parameters that are to be used or are to be deleted.....	83
Table K.1 – Decibel beamwidth levels for determining midpoints.....	94

This document is a preview generated by EVS

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – HYDROPHONES –

Part 1: Measurement and characterization of medical ultrasonic fields

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62127-1 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2007 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) The upper frequency limit of 40 MHz has been removed.
- b) Hydrophone sensitivity definitions have been changed to recognize sensitivities as complex-valued quantities.
- c) Procedures and requirements for narrow-band approximation and broadband measurements have been modified; details on waveform deconvolution have been added.
- d) Procedures for spatial averaging correction have been amended.
- e) Annex D, Annex E and bibliography have been updated to support the changes of the normative parts.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/783/FDIS	87/788/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

A list of all parts of IEC 62127 series, published under the general title *Ultrasonics – Hydrophones*, can be found on the IEC website.

NOTE Words in **bold** in the text are terms defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

The main purpose of this document is to define various acoustic parameters that can be used to specify and characterize ultrasonic fields propagating in liquids, and, in particular, water, using hydrophones. Measurement procedures are outlined that may be used to determine these parameters. Specific device related measurement standards, for example IEC 61689, IEC 61157, IEC 61847 or IEC 62359, can refer to this document for appropriate acoustic parameters. In IEC 62359, some additional measurement methods for attenuated parameters and indices are described addressing the specific needs of acoustic output characterization of ultrasonic diagnostic equipment in accordance with IEC 60601-2-37.

The philosophy behind this document is the specification of the acoustic field in terms of acoustic pressure parameters, acoustic pressure being the primary measurement quantity when hydrophones are used to characterize the field.

Intensity parameters are specified in this document, but these are regarded as derived quantities that are meaningful only under certain assumptions related to the ultrasonic field being measured.

ULTRASONICS – HYDROPHONES –

Part 1: Measurement and characterization of medical ultrasonic fields

1 Scope

This part of IEC 62127 specifies methods of use of calibrated **hydrophones** for the measurement in liquids of acoustic fields generated by ultrasonic medical equipment including **bandwidth** criteria and calibration frequency range requirements in dependence on the spectral content of the fields to be characterized.

This document:

- defines a group of acoustic parameters that can be measured on a physically sound basis;
- defines a second group of parameters that can be derived under certain assumptions from these measurements, and called derived intensity parameters;
- defines a measurement procedure that can be used for the determination of acoustic pressure parameters;
- defines the conditions under which the measurements of acoustic parameters can be made using calibrated **hydrophones**;
- defines procedures for correcting for limitations caused by the use of **hydrophones** with finite **bandwidth** and finite active element size, and for estimating the corresponding **uncertainties**.

NOTE 1 Throughout this document, SI units are used. In the specification of certain parameters, such as **beam areas** and intensities, it can be convenient to use decimal multiples or submultiples. For example, **beam area** is likely to be specified in cm^2 and intensities in W/cm^2 or mW/cm^2 .

NOTE 2 The **hydrophone** as defined can be of a piezoelectric or an optic type.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60565-1, *Underwater acoustics – Hydrophones – Calibration of hydrophones – Part 1: Procedures for free-field calibration of hydrophones*

IEC 61689, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 0,5 MHz to 5 MHz*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields up to 40 MHz*

IEC 62127-3, *Ultrasonics – Hydrophones – Part 3: Properties of hydrophones for ultrasonic fields up to 40 MHz*

IEC 63009, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 20 kHz to 500 kHz*

ISO 16269-6, *Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 62127-2, IEC 62127-3 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

acoustic pulse waveform

temporal waveform of the **instantaneous acoustic pressure** at a specified position in an acoustic field and displayed over a period sufficiently long to include all significant acoustic information in a single pulse or tone-burst, or one or more cycles in a continuous wave

Note 1 to entry: Temporal waveform is a representation (e.g. oscilloscope presentation or equation) of the **instantaneous acoustic pressure**.

3.2

acoustic repetition period

arp

pulse repetition period for non-automatic scanning systems and the **scan repetition period** for automatic scanning systems, equal to the time interval between corresponding points of consecutive cycles for continuous wave systems

Note 1 to entry: The **acoustic repetition period** is expressed in seconds (s).

3.3

acoustic-working frequency

acoustic frequency

frequency of an acoustic signal based on the observation of the output of a **hydrophone** placed in an acoustic field at the position corresponding to the **spatial-peak temporal-peak acoustic pressure**

Note 1 to entry: The signal is analysed using either the **zero-crossing acoustic-working frequency** technique or a spectrum analysis method. Acoustic-working frequencies are defined in 3.3.1, 3.3.2, 3.3.3, 3.3.4 and 3.3.5.

Note 2 to entry: In a number of cases the present definition is not very helpful or convenient, especially for **broadband transducers**. In that case, a full description of the frequency spectrum is expected to be given in order to enable any frequency-dependent correction to the signal.

Note 3 to entry: **Acoustic frequency** is expressed in hertz (Hz).

3.3.1

zero-crossing acoustic-working frequency

f_{awf}

number, n , of consecutive half-cycles (irrespective of polarity) divided by twice the time between the commencement of the first half-cycle and the end of the n -th half-cycle

Note 1 to entry: None of the n consecutive half-cycles are expected to show evidence of phase change.

Note 2 to entry: The measurement are performed at terminals in the receiver that are as close as possible to the receiving transducer (**hydrophone**) and, in all cases, before rectification.

Note 3 to entry: This frequency is determined according to the procedure specified in IEC TR 60854.

Note 4 to entry: This frequency is intended for continuous-wave systems only.