
Wheelchair seating —
Part 11:
Determination of dissipation
characteristics of sensible
perspiration into seat cushions



This document is a preview generated by EUS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Abbreviated terms	2
5 Principle	2
6 Test apparatus	2
6.1 Perspiration rigid cushion loading indenter	2
6.1.1 Shape of PRCLI	2
6.1.2 Temperature control on the PRCLI	2
6.1.3 Loading mechanism	3
6.1.4 Discharge mechanism	3
6.1.5 Sensing system	4
6.2 Electrical impedance measurement system	4
6.3 Test environment	4
7 Preparation for testing	5
8 Test method	5
9 Test report	5
Bibliography	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 173, *Assistive products*, Subcommittee SC 1, *Wheelchairs*.

This first edition of ISO 16840-11 cancels and replaces ISO/TS 16840-11:2014, which has been technically revised.

The main changes are as follows:

- addition of details around the test apparatus, preparation and method to make the test easier to process.

A list of all parts in the ISO 16840 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Dissipating perspiration and moisture on seat cushions is important to prevent pressure injuries and tissue maceration, and for comfortable sitting. Perspiration can occur as water vapour, and/or as moisture – sensible moisture has not undergone a phase change to vapour, whereas insensible moisture has undergone the phase change and is in the vapour phase.

The results of the testing methods in this document have not been clinically validated to support any medical claims of prevention of pressure injuries or prevention of tissue maceration.

Wheelchair seating —

Part 11:

Determination of dissipation characteristics of sensible perspiration into seat cushions

1 Scope

This document specifies a method for determining the dissipation characteristics of simulated sensible perspiration exposure on seat cushions of a wheelchair under test conditions that simulate body loading on support surfaces with flat and contoured profiles. The test method is applicable to a wide variety of seat cushion materials and constructions that are used across the world.

The test method is intended to determine how the cushion handles a liquid load. It is not intended to determine how the cushion responds to a continuous release of liquid or vapour.

This document also specifies apparatus to measure dissipation characteristics (of perspiration) of seat cushions and the approach to be employed in measuring these characteristics.

This document is applicable to seat cushions for wheelchairs that include a cushion cover.

NOTE The significance of the dissipation characteristics of the seat cushion can be greatly affected by the clothing worn, for example a moisture wicking cover might not help the skin condition of the patient sitting in a wet adult diaper.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16840-2:2018, *Wheelchair seating — Part 2: Determination of physical and mechanical characteristics of seat cushions intended to manage tissue integrity*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

indenter ischial tuberosity

one of two locations on the inferior surface of an indenter that corresponds in position to the location of a human ischial tuberosity

3.2

electrical impedance

complex number ratio of the voltage to the current in an alternating current circuit

Note 1 to entry: Impedance extends the concept of electrical resistance to the alternating current circuit.