
**Dentistry — Elastomeric auxiliaries
for use in orthodontics**

*Médecine bucco-dentaire — Auxiliaires élastomériques utilisés en
orthodontie*



This document is a preview generated by ELS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	3
4.1 General	3
4.2 Dimensions	3
4.3 Mechanical properties	3
4.3.1 Initial extension force	3
4.3.2 24 h residual force	3
4.3.3 Ultimate extension	3
5 Test methods	4
5.1 Sampling	4
5.2 Ambient conditions	4
5.3 Dimensions	4
5.3.1 Apparatus	4
5.3.2 Procedure	4
5.4 Initial extension force, F_0	4
5.4.1 Apparatus	4
5.4.2 Procedure	5
5.5 24 h residual force, F_{24}	5
5.5.1 Apparatus	5
5.5.2 Procedure	6
5.6 Ultimate extension, A	6
5.6.1 Apparatus	6
5.6.2 Procedure	7
5.7 Treatment of results	7
6 Marking, labelling and packaging	7
6.1 General requirements	7
6.2 Packaging and labelling	7
Bibliography	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 1, *Filling and restorative materials*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 21606:2007), which has been technically revised.

The main changes compared to the previous edition are as follows:

- subclause [5.1](#) has been clarified to type test with specification of specimen test quantity;
- subclause [5.5.2](#) has been clarified with the addition of the word “quickly” to the transfer of specimens from support plate to tensile tester adapters.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document has been developed as a result of the difficulty often encountered by clinicians in making meaningful comparisons between elastomeric auxiliaries using the information currently available from manufacturers and suppliers.

NOTE Auxiliaries include orthodontic elastics, orthodontic elastomeric chains, orthodontic thread, orthodontic elastomeric ligatures and orthodontic elastomeric separators.

Specific qualitative and quantitative requirements for freedom from biological hazards are not included in this document, but it is recommended that in assessing possible biological hazards reference should be made to ISO 10993-1 and ISO 7405.

Dentistry — Elastomeric auxiliaries for use in orthodontics

1 Scope

This document specifies the requirements and their test methods applicable to all elastomeric auxiliaries used for orthodontics both inside and outside the mouth, in conjunction with fixed and removable appliances.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 1942, *Dentistry — Vocabulary*

ISO 8601-1, *Date and time — Representations for information interchange — Part 1: Basic rules*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

elastomeric auxiliary

device with elastomeric properties used for transmission of forces including orthodontic *elastics* (3.2), orthodontic *elastomeric chains* (3.4), orthodontic threads, orthodontic *elastomeric ligatures* (3.5) and orthodontic *elastomeric separators* (3.6)

3.2

elastic

intra-oral and extra-oral elastomeric ring used to apply forces to orthodontic devices

3.3

elastomeric thread

elastomeric strand (that can be hollow) of constant cross-section used to apply forces to orthodontic devices

3.4

elastomeric chain

interconnected elastomeric rings or a multi-perforated elastomeric band used to apply forces to orthodontic devices

3.5

elastomeric ligature

elastomeric ring used to retain wires to orthodontic attachments