
Dentistry — Mobile dental units and dental patient chairs —

Part 1: General requirements

*Médecine bucco-dentaire — Units dentaires et fauteuils dentaires
patient mobiles —*

Partie 1: Exigences générales



This document is a preview generated by ELS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

| | |
|--|-----------|
| Foreword | iv |
| Introduction | v |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 Classification | 2 |
| 5 Requirements | 2 |
| 5.1 General requirements | 2 |
| 5.1.1 Basic safety and essential performance | 2 |
| 5.1.2 Controls and indicators | 2 |
| 5.1.3 Function stop system | 2 |
| 5.1.4 Usability | 2 |
| 5.1.5 Cleaning and disinfection | 3 |
| 5.1.6 Excessive temperatures | 3 |
| 5.1.7 Biocompatibility | 3 |
| 5.1.8 Solids filter | 3 |
| 5.1.9 Amalgam separator device | 3 |
| 5.1.10 Upholstery and padding | 3 |
| 5.1.11 Air, water, suction, and waste water systems | 3 |
| 5.2 Mechanical requirements | 4 |
| 5.2.1 General mechanical requirements | 4 |
| 5.2.2 Mechanical requirements for mobile dental units | 5 |
| 5.2.3 Mechanical requirements for mobile dental patient chairs | 5 |
| 5.3 Electrical requirements | 5 |
| 5.3.1 General electrical requirements | 5 |
| 5.3.2 Test points for periodic safety checks | 6 |
| 5.3.3 Position limiting of mobile dental patient chair | 6 |
| 5.4 Test report | 6 |
| 6 Sampling | 6 |
| 7 Testing | 6 |
| 7.1 Visual inspection | 6 |
| 7.1.1 Visual inspection of device | 6 |
| 7.1.2 Visual inspection of documentation or test reports | 6 |
| 7.2 Tipping and stability of mobile dental patient chair | 6 |
| 8 Instructions for use | 7 |
| 9 Technical description | 8 |
| 10 Marking | 8 |
| 11 Packaging | 8 |
| Annex A (informative) Example of a test report | 9 |
| Bibliography | 13 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 5467 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document was developed with consideration of ISO 7494-1:2018 to be applicable to mobile dental units and dental patient chairs by

- limiting the scope to mobile dental units and dental patient chairs,
- providing instability and rough handling requirements in [5.2.1.4](#) and [5.2.1.5](#),
- providing a requirement for locking in position in [5.2.1.6](#),
- providing a requirement for dimension in [5.2.1.7](#), and
- providing a requirement for impact to [5.2.1.8](#).

Dentistry — Mobile dental units and dental patient chairs —

Part 1: General requirements

1 Scope

This document specifies the requirements and test methods for mobile dental units and dental patient chairs that is intended to be used within a permanent healthcare facility regardless of whether they are or not electrically powered.

This document also specifies the requirements for the instructions for use, for the technical description, for marking and for packaging.

Operator's stools, stationary dental equipment, other types of mobile dental equipment, portable dental equipment and operating lights are not in the scope of this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 4073, *Dentistry — Information system on the location of dental equipment in the working area of the oral health care provider*

ISO 5467-2, *Dentistry — Mobile dental units and dental patient chairs — Part 2: Air, water, suction and wastewater systems*

ISO 7494-1:2018, *Dentistry — Stationary dental units and dental patient chairs — Part 1: General requirements*

ISO 23402-1:2020, *Dentistry — Portable dental equipment for use in non-permanent healthcare environment — Part 1: General requirements*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 80601-2-60:2019, *Medical electrical equipment — Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942, ISO 4073, ISO 7494-1, ISO 23402-1, IEC 60601-1, IEC 80601-2-60 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>